



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040124\
Data File : BM044856.D
Acq On : 01 Apr 2024 09:30
Operator : MA/JU
Sample : DFTPP015
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP015

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030524.MA.M
Title : SVOA CALIBRATION
Last Update : Sun Mar 31 13:54:26 2024



AutoFind: Scans 2451, 2452, 2453; Background Corrected with Scan 2445

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.0	66589	PASS
68	69	0.00	2	1.7	916	PASS
69	198	0.00	100	47.1	54083	PASS
70	69	0.00	2	0.5	294	PASS
127	198	10	80	53.6	61563	PASS
197	198	0.00	2	0.4	412	PASS
198	198	100	100	100.0	114869	PASS
199	198	5	9	6.7	7706	PASS
275	198	10	60	25.3	29048	PASS
365	198	1	100	3.9	4455	PASS
441	198	0.01	100	13.9	15982	PASS
442	198	50	100	83.9	96427	PASS
443	442	15	24	19.4	18700	PASS

m/z	Abundance
39.00	185.0
40.00	1637.0
44.00	2091.0
49.00	329.0
73.00	616.0
83.80	205.0
95.90	218.0
146.80	181.0
191.00	227.0
192.80	152.0
206.90	1583.0
207.90	236.0
209.00	244.0
280.90	382.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

m/z	Abundance
36.90	333.0
38.10	947.0
39.00	5595.0
40.00	2206.0
41.00	304.0
44.00	2629.0
44.90	236.0
48.90	549.0
50.00	21456.0
51.00	77632.0
52.00	3936.0
52.90	187.0
54.90	497.0
56.00	1741.0
57.00	4381.0
57.90	204.0
61.00	851.0
62.00	820.0
62.90	2130.0
64.00	247.0
65.10	1408.0
67.90	855.0
68.90	60464.0
69.90	314.0
73.00	1122.0
74.00	6636.0
75.00	10134.0
76.00	3724.0
77.00	76128.0
78.00	4764.0
79.00	4348.0
80.00	3482.0
81.00	4578.0
82.00	1096.0
82.90	1270.0
84.00	280.0
84.90	720.0
86.00	1196.0
86.90	500.0
87.90	166.0
90.80	837.0
92.00	1200.0
92.90	7819.0
93.90	520.0
96.00	495.0
98.00	5790.0
99.00	5621.0
100.00	594.0
101.00	2563.0
102.00	153.0
102.90	971.0
103.90	1591.0
104.90	1342.0
105.90	534.0
107.00	20144.0
108.00	3148.0
108.80	422.0
109.90	31992.0
110.90	5545.0
112.00	637.0
112.90	334.0
115.90	1062.0
116.90	19008.0
118.00	1296.0
118.90	240.0
119.90	336.0
121.90	1597.0
122.90	1941.0
123.90	1092.0
124.90	824.0
126.10	309.0
127.00	66392.0
127.90	5796.0
128.90	26264.0
129.90	2332.0
130.90	536.0
131.90	272.0
132.90	230.0
134.00	549.0
134.90	2131.0
136.00	855.0
136.90	1192.0
138.00	317.0
139.90	387.0
141.00	3235.0
142.00	959.0
142.90	725.0
143.80	198.0
144.80	170.0
145.90	654.0
146.90	2023.0
147.90	4738.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

148.90	881.0
149.90	270.0
151.10	387.0
151.70	166.0
153.00	1062.0
153.80	525.0
154.90	2150.0
156.00	2943.0
157.00	477.0
157.80	731.0
158.90	378.0
159.90	1222.0
160.90	1542.0
161.90	473.0
162.80	174.0
163.90	260.0
164.90	1029.0
165.90	732.0
166.90	7534.0
167.90	5008.0
168.90	773.0
170.00	178.0
170.90	294.0
171.80	659.0
172.90	744.0
173.90	1344.0
174.90	2787.0
175.90	727.0
176.80	1437.0
178.00	299.0
178.90	4498.0
179.90	3495.0
180.90	1450.0
181.80	369.0
182.80	155.0
183.80	293.0
184.90	2284.0
185.90	15882.0
187.00	4839.0
187.80	523.0
188.90	1113.0
189.80	194.0
190.90	762.0
191.90	1450.0
193.00	1863.0
193.90	439.0
196.00	3047.0
197.90	110064.0
198.90	7241.0
199.90	792.0
201.20	491.0
202.90	1049.0
203.90	5282.0
205.00	7643.0
206.00	30936.0
207.00	5773.0
207.90	1138.0
208.90	526.0
209.90	452.0
211.00	1258.0
214.80	442.0
215.90	587.0
216.90	8351.0
217.90	1156.0
221.00	4842.0
221.90	1183.0
223.00	2114.0
224.00	17776.0
225.00	4899.0
226.00	438.0
226.90	7235.0
228.00	873.0
228.90	1617.0
229.90	309.0
230.80	536.0
232.80	274.0
233.90	320.0
234.90	489.0
235.90	444.0
236.90	883.0
238.80	245.0
239.90	175.0
240.90	511.0
241.90	898.0
243.00	1122.0
244.00	13309.0
244.90	1879.0
246.00	2314.0
246.80	492.0
247.90	216.0
248.90	503.0
252.00	187.0
252.90	230.0
253.80	295.0
254.90	64672.0
255.90	10542.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

256.90	618.0
258.00	3941.0
259.00	568.0
259.90	208.0
264.90	1723.0
265.90	343.0
270.90	222.0
272.90	1901.0
273.90	4658.0
275.00	27688.0
276.00	4072.0
276.90	2421.0
277.90	558.0
280.90	376.0
283.00	416.0
283.90	232.0
285.00	563.0
292.90	553.0
294.80	217.0
295.90	8937.0
296.90	1246.0
303.00	978.0
303.90	304.0
313.90	306.0
314.80	794.0
316.00	474.0
321.00	189.0
323.00	2895.0
324.00	440.0
326.80	498.0
328.00	311.0
332.00	181.0
332.90	168.0
333.90	1459.0
335.00	372.0
341.00	450.0
345.80	722.0
352.00	887.0
353.00	379.0
354.00	947.0
355.10	199.0
364.90	3532.0
365.90	613.0
370.90	239.0
371.90	1249.0
373.00	266.0
383.00	306.0
389.80	151.0
401.90	512.0
403.00	532.0
404.00	177.0
420.90	616.0
421.90	440.0
423.00	3943.0
424.00	1313.0
425.10	186.0
441.00	12591.0
442.00	73912.0
443.00	14875.0
443.90	1327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

m/z	Abundance
35.90	177.0
37.00	376.0
37.90	1214.0
39.00	5550.0
39.90	1916.0
40.90	220.0
41.90	158.0
42.90	185.0
44.00	2374.0
44.90	337.0
48.90	759.0
50.00	21984.0
51.00	77824.0
52.00	3702.0
53.00	205.0
54.90	400.0
55.90	2233.0
56.90	4479.0
57.90	204.0
61.00	837.0
61.90	788.0
63.00	2719.0
64.00	468.0
64.90	1363.0
65.90	189.0
68.00	1245.0
68.90	65432.0
70.00	362.0
73.00	1337.0
74.00	6916.0
75.00	10045.0
76.00	4214.0
77.00	79000.0
78.00	5670.0
79.00	4886.0
79.90	3629.0
81.00	5418.0
81.90	1364.0
83.00	1557.0
84.00	378.0
84.90	799.0
85.90	1590.0
86.90	519.0
88.00	254.0
88.80	154.0
90.90	1216.0
92.00	1395.0
92.90	8858.0
93.90	586.0
95.90	572.0
97.90	7086.0
98.90	5777.0
99.80	536.0
101.00	2844.0
101.80	205.0
102.90	772.0
103.90	1967.0
105.00	1469.0
105.90	589.0
106.90	22944.0
108.00	3396.0
108.90	663.0
109.90	37160.0
111.00	6295.0
111.80	573.0
113.00	202.0
116.00	1094.0
117.00	21736.0
117.90	1874.0
119.00	223.0
119.80	273.0
121.90	1092.0
122.90	2042.0
123.90	1208.0
125.00	1232.0
126.00	307.0
127.00	73888.0
128.00	5602.0
128.90	28792.0
129.90	2538.0
131.00	486.0
132.10	282.0
132.90	172.0
134.00	832.0
135.00	2517.0
135.90	969.0
136.90	1492.0
137.80	301.0
139.90	454.0
140.90	3321.0
141.90	1128.0
142.80	970.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

143.80	157.0
144.90	284.0
145.90	769.0
146.90	2154.0
147.90	5550.0
148.90	1255.0
149.90	267.0
150.90	468.0
151.80	287.0
153.00	985.0
154.00	963.0
154.90	2180.0
155.90	3154.0
157.00	641.0
157.70	602.0
158.90	488.0
159.90	1332.0
161.00	1823.0
162.00	387.0
162.90	201.0
163.90	214.0
164.90	1393.0
165.80	1137.0
166.90	9101.0
167.90	5603.0
168.90	795.0
169.80	358.0
170.90	302.0
171.80	664.0
172.90	972.0
173.90	1785.0
175.00	3267.0
176.00	1030.0
176.90	1488.0
177.90	556.0
178.90	5923.0
179.90	4306.0
180.90	1686.0
181.90	381.0
182.90	207.0
183.90	537.0
185.00	2737.0
186.00	21240.0
187.00	5531.0
188.00	746.0
188.90	1328.0
189.90	262.0
190.90	869.0
191.90	2275.0
193.00	2142.0
194.00	378.0
195.00	290.0
196.00	4005.0
196.90	795.0
197.90	140160.0
198.90	9806.0
199.80	903.0
201.20	672.0
202.00	252.0
202.90	1239.0
204.00	5789.0
205.00	9332.0
206.00	39184.0
207.00	6615.0
207.90	1542.0
209.00	751.0
210.10	525.0
210.90	1429.0
211.90	272.0
213.00	204.0
214.90	418.0
215.90	1095.0
216.90	10946.0
217.90	1510.0
220.20	254.0
220.90	5442.0
221.80	1478.0
222.90	2473.0
224.00	23584.0
225.00	5980.0
225.90	644.0
226.90	10153.0
228.00	1276.0
228.90	2027.0
229.90	262.0
230.90	684.0
231.90	248.0
233.00	216.0
234.00	609.0
234.90	783.0
235.80	707.0
237.00	821.0
238.90	456.0
240.00	539.0
240.80	661.0
242.00	1460.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

242.90	1396.0
244.00	17656.0
244.90	1970.0
246.00	3236.0
246.90	805.0
247.80	158.0
248.80	556.0
250.80	197.0
251.80	309.0
253.00	485.0
253.90	615.0
255.00	85864.0
255.90	13176.0
256.90	1076.0
257.90	5491.0
258.90	949.0
264.90	2385.0
265.90	398.0
270.80	252.0
271.80	243.0
273.00	2677.0
274.00	6850.0
275.00	34296.0
276.00	4859.0
276.90	2889.0
277.90	815.0
281.00	407.0
282.10	159.0
283.00	494.0
283.90	203.0
284.90	620.0
285.90	169.0
288.90	194.0
289.90	194.0
291.90	201.0
292.90	884.0
293.90	174.0
294.80	166.0
295.90	11735.0
296.90	1639.0
300.80	190.0
301.80	150.0
303.00	1417.0
304.00	335.0
309.90	187.0
314.00	632.0
315.00	1418.0
316.00	711.0
321.00	267.0
323.00	3707.0
323.90	433.0
326.90	787.0
327.80	296.0
328.00	300.0
332.00	257.0
332.90	362.0
334.00	2321.0
335.00	536.0
340.90	466.0
341.90	213.0
345.90	905.0
346.80	255.0
351.90	1171.0
353.00	700.0
353.90	1226.0
354.90	336.0
364.90	5628.0
365.90	823.0
371.00	178.0
372.00	1731.0
372.90	345.0
382.90	462.0
390.00	239.0
391.90	202.0
400.90	200.0
401.80	874.0
402.90	1186.0
403.90	429.0
421.00	818.0
422.00	880.0
423.00	6103.0
424.00	1723.0
424.70	170.0
441.00	19560.0
442.00	116152.0
443.00	22208.0
444.00	1977.0
444.90	171.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

m/z	Abundance
36.80	248.0
38.00	606.0
39.00	3199.0
40.00	2080.0
41.00	204.0
43.00	187.0
44.00	2147.0
44.90	188.0
49.00	376.0
50.00	13027.0
51.00	44312.0
52.10	1916.0
54.90	256.0
55.90	1166.0
57.00	2450.0
61.00	468.0
62.00	761.0
63.00	1798.0
64.00	208.0
65.10	701.0
67.90	648.0
69.00	36352.0
69.90	207.0
73.00	952.0
74.00	4431.0
75.00	6409.0
76.00	2096.0
77.00	47008.0
78.00	3287.0
79.00	2976.0
79.90	1876.0
80.90	3176.0
81.90	897.0
82.90	636.0
83.90	258.0
85.00	417.0
85.90	1008.0
86.90	358.0
90.90	966.0
91.90	824.0
93.00	4787.0
93.80	268.0
95.10	156.0
95.90	337.0
98.00	3858.0
98.90	3021.0
99.80	302.0
100.90	1756.0
103.10	518.0
103.90	992.0
105.00	836.0
106.00	335.0
107.00	13581.0
107.90	1828.0
108.80	397.0
110.00	19512.0
110.90	3650.0
112.00	468.0
113.00	234.0
115.90	675.0
117.00	12576.0
117.80	692.0
119.90	177.0
122.00	671.0
122.90	1384.0
123.80	397.0
124.90	753.0
126.00	151.0
127.00	44408.0
128.00	3313.0
128.90	17400.0
129.90	1353.0
130.90	422.0
133.00	205.0
133.90	518.0
135.00	1605.0
136.00	698.0
136.90	953.0
139.80	269.0
140.90	2080.0
141.90	564.0
142.80	412.0
144.90	154.0
146.00	379.0
146.90	1638.0
147.90	2958.0
149.00	745.0
150.00	231.0
150.90	303.0
152.90	775.0
153.90	479.0
154.90	1764.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

156.00	1820.0
157.10	324.0
157.90	476.0
158.90	318.0
159.90	686.0
161.00	1182.0
161.80	371.0
164.80	781.0
165.90	942.0
167.00	5840.0
167.90	3405.0
169.00	387.0
169.80	211.0
170.90	258.0
171.90	573.0
172.90	661.0
173.90	1114.0
175.00	1910.0
176.10	373.0
177.00	938.0
177.90	364.0
178.90	3668.0
179.90	2718.0
181.00	1260.0
183.80	229.0
184.90	2024.0
186.00	12969.0
187.00	3540.0
187.90	286.0
188.90	1013.0
189.80	163.0
190.90	571.0
191.90	1627.0
193.00	1601.0
193.90	310.0
195.90	2865.0
197.00	442.0
197.90	94384.0
198.90	6072.0
199.90	494.0
201.30	475.0
201.50	489.0
202.90	843.0
204.00	3543.0
204.90	5857.0
206.00	26144.0
207.00	4880.0
207.90	1003.0
208.80	518.0
210.10	631.0
210.90	894.0
214.80	215.0
215.80	525.0
216.90	7507.0
218.00	847.0
221.00	3738.0
221.90	1106.0
222.90	1587.0
224.00	15072.0
224.90	3800.0
226.00	419.0
226.90	6781.0
227.90	1004.0
228.90	1343.0
230.90	610.0
233.90	571.0
234.90	804.0
236.00	242.0
236.90	696.0
240.90	435.0
241.90	1041.0
242.90	778.0
244.00	12150.0
244.90	1496.0
245.90	2311.0
246.90	540.0
248.90	178.0
251.80	205.0
252.90	242.0
253.90	439.0
255.00	61160.0
255.90	8947.0
256.90	933.0
257.90	3354.0
258.90	665.0
265.00	1704.0
266.00	285.0
270.90	193.0
272.90	2271.0
273.90	4790.0
274.90	25160.0
275.90	3476.0
276.90	2523.0
278.00	358.0
281.00	311.0
282.90	407.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015

283.90	254.0
284.90	398.0
292.00	157.0
292.90	435.0
293.90	257.0
294.80	182.0
295.90	9371.0
296.90	1354.0
302.90	825.0
304.00	225.0
313.90	537.0
314.90	1026.0
315.90	643.0
321.00	205.0
323.00	2483.0
324.00	593.0
326.80	657.0
327.90	335.0
331.80	182.0
333.00	198.0
334.00	1808.0
334.80	428.0
340.90	296.0
345.90	650.0
351.90	885.0
352.90	478.0
354.00	944.0
354.90	172.0
364.90	4206.0
365.90	709.0
370.80	188.0
371.90	1312.0
372.90	312.0
383.00	524.0
389.60	169.0
402.00	698.0
403.00	1024.0
403.90	268.0
420.90	822.0
421.90	536.0
423.00	4874.0
424.00	1142.0
441.00	15795.0
442.00	99216.0
443.00	19016.0
443.90	1665.0

Instrument :
BNA_M
ClientSampleId :
DFTPP015