

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
 Data File : BM049777.D
 Acq On : 01 Apr 2025 16:48
 Operator : RC/JU
 Sample : Q1679-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP-9

Quant Time: Apr 01 17:48:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 13 14:55:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	305297	20.000	ng	-0.01
21) Naphthalene-d8	10.586	136	1026961	20.000	ng	-0.02
39) Acenaphthene-d10	14.433	164	626906	20.000	ng	-0.01
64) Phenanthrene-d10	17.174	188	1099906	20.000	ng	-0.01
76) Chrysene-d12	21.403	240	767632	20.000	ng	-0.02
86) Perylene-d12	24.409	264	720807	20.000	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	2581065	142.651	ng	-0.01
7) Phenol-d6	6.963	99	2961279	130.416	ng	-0.02
23) Nitrobenzene-d5	8.951	82	2071628	103.569	ng	-0.01
42) 2,4,6-Tribromophenol	15.921	330	1207051	164.889	ng	-0.01
45) 2-Fluorobiphenyl	13.057	172	5347845	106.835	ng	-0.01
79) Terphenyl-d14	19.798	244	6480882	142.270	ng	-0.01


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
Data File : BM049777.D
Acq On : 01 Apr 2025 16:48
Operator : RC/JU
Sample : Q1679-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP-9

Quant Time: Apr 01 17:48:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 14:55:34 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
Data File : BM049777.D
Acq On : 01 Apr 2025 16:48
Operator : RC/JU
Sample : Q1679-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP-9

Quant Time: Apr 01 17:48:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 14:55:34 2025
Response via : Initial Calibration

