

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
 Data File : BM049767.D
 Acq On : 01 Apr 2025 10:18
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 04/02/2025
 Supervised By :Jagrut Upadhyay 04/02/2025

Quant Time: Apr 01 11:46:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 13 14:55:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	337112	20.000	ng	-0.01
21) Naphthalene-d8	10.586	136	1252823	20.000	ng	-0.02
39) Acenaphthene-d10	14.433	164	834176	20.000	ng	-0.01
64) Phenanthrene-d10	17.174	188	1532241	20.000	ng	-0.01
76) Chrysene-d12	21.409	240	1182765	20.000	ng	-0.02
86) Perylene-d12	24.415	264	962538	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	1572130	78.689	ng	-0.01
7) Phenol-d6	6.969	99	2144767	85.542	ng	-0.01
23) Nitrobenzene-d5	8.951	82	1982948	81.263	ng	-0.01
42) 2,4,6-Tribromophenol	15.921	330	841905	86.432	ng	-0.01
45) 2-Fluorobiphenyl	13.057	172	5529414	83.015	ng	-0.01
79) Terphenyl-d14	19.803	244	7109403	101.290	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.275	88	307910	35.601	ng	97
3) Pyridine	3.675	79	856347m	38.770	ng	
4) n-Nitrosodimethylamine	3.581	42	387653	39.665	ng	96
6) Aniline	7.122	93	861876	41.031	ng	99
8) 2-Chlorophenol	7.363	128	794901	40.177	ng	99
9) Benzaldehyde	6.934	77	435903	35.168	ng	99
10) Phenol	6.992	94	1011917	41.623	ng	98
11) bis(2-Chloroethyl)ether	7.216	93	813640	41.329	ng	96
12) 1,3-Dichlorobenzene	7.681	146	928678	38.624	ng	98
13) 1,4-Dichlorobenzene	7.828	146	941528	38.816	ng	99
14) 1,2-Dichlorobenzene	8.145	146	919994	39.908	ng	98
15) Benzyl Alcohol	8.034	79	664440	42.229	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.322	45	1004495	43.412	ng	97
17) 2-Methylphenol	8.239	107	616084	42.684	ng	100
18) Hexachloroethane	8.875	117	330882	38.628	ng	93
19) n-Nitroso-di-n-propyla...	8.604	70	696178	47.885	ng	96
20) 3+4-Methylphenols	8.569	107	846734	43.775	ng	98
22) Acetophenone	8.616	105	1304418	40.380	ng	# 98
24) Nitrobenzene	8.992	77	911220	39.603	ng	99
25) Isophorone	9.522	82	1550110	41.374	ng	100
26) 2-Nitrophenol	9.704	139	393929	40.225	ng	99
27) 2,4-Dimethylphenol	9.769	122	497943	39.946	ng	97
28) bis(2-Chloroethoxy)met...	10.004	93	1092065	41.386	ng	98
29) 2,4-Dichlorophenol	10.239	162	854259	42.057	ng	97
30) 1,2,4-Trichlorobenzene	10.451	180	1000462	39.367	ng	99
31) Naphthalene	10.639	128	2565330	39.821	ng	100
32) Benzoic acid	9.916	122	439432	41.450	ng	97
33) 4-Chloroaniline	10.745	127	897486	40.822	ng	99
34) Hexachlorobutadiene	10.933	225	601986	38.371	ng	99
35) Caprolactam	11.539	113	198870	42.171	ng	96
36) 4-Chloro-3-methylphenol	11.880	107	746125	42.804	ng	98
37) 2-Methylnaphthalene	12.251	142	1902642	42.238	ng	100
38) 1-Methylnaphthalene	12.474	142	1861529	42.833	ng	100
40) 1,2,4,5-Tetrachloroben...	12.621	216	1179803	38.504	ng	99
41) Hexachlorocyclopentadiene	12.610	237	438408	38.621	ng	98
43) 2,4,6-Trichlorophenol	12.863	196	695798	40.545	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
 Data File : BM049767.D
 Acq On : 01 Apr 2025 10:18
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025

Quant Time: Apr 01 11:46:32 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Mar 13 14:55:34 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.939	196	753129	40.216	ng	98
46) 1,1'-Biphenyl	13.269	154	2556224	39.760	ng	99
47) 2-Chloronaphthalene	13.310	162	1963725	39.242	ng	99
48) 2-Nitroaniline	13.510	65	455968	40.589	ng	100
49) Acenaphthylene	14.157	152	2810898	40.186	ng	100
50) Dimethylphthalate	13.898	163	2324364	40.471	ng	100
51) 2,6-Dinitrotoluene	14.010	165	472221	41.035	ng	94
52) Acenaphthene	14.498	154	1852901	40.385	ng	99
53) 3-Nitroaniline	14.333	138	430177	39.283	ng	99
54) 2,4-Dinitrophenol	14.539	184	218039	38.924	ng	95
55) Dibenzofuran	14.833	168	3088131	40.393	ng	99
56) 4-Nitrophenol	14.651	139	350690	36.790	ng	96
57) 2,4-Dinitrotoluene	14.792	165	615313	41.553	ng	99
58) Fluorene	15.480	166	2684693	43.130	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.063	232	625262	39.887	ng	98
60) Diethylphthalate	15.262	149	2220802	41.056	ng	99
61) 4-Chlorophenyl-phenyle...	15.480	204	1461965	43.011	ng	97
62) 4-Nitroaniline	15.498	138	455574	35.986	ng	96
63) Azobenzene	15.768	77	2184719	42.285	ng	97
65) 4,6-Dinitro-2-methylph...	15.557	198	313084	39.734	ng	93
66) n-Nitrosodiphenylamine	15.692	169	1940059	41.790	ng	100
67) 4-Bromophenyl-phenylether	16.374	248	727065	41.099	ng	96
68) Hexachlorobenzene	16.486	284	788814	39.652	ng	97
69) Atrazine	16.639	200	321907	31.464	ng	98
70) Pentachlorophenol	16.827	266	490034	39.197	ng	99
71) Phenanthrene	17.221	178	3446408	40.507	ng	100
72) Anthracene	17.309	178	3339045	40.524	ng	100
73) Carbazole	17.574	167	2901083	38.877	ng	99
74) Di-n-butylphthalate	18.151	149	3364666	41.139	ng	100
75) Fluoranthene	19.233	202	3742501	38.207	ng	100
77) Benzidine	19.415	184	590016	41.522	ng	99
78) Pyrene	19.592	202	3826162	46.302	ng	99
80) Butylbenzylphthalate	20.497	149	1084171	42.277	ng	99
81) Benzo(a)anthracene	21.392	228	3155239	39.939	ng	100
82) 3,3'-Dichlorobenzidine	21.315	252	933134	36.279	ng	99
83) Chrysene	21.456	228	2925890	39.070	ng	99
84) Bis(2-ethylhexyl)phtha...	21.327	149	1720829	43.057	ng	99
85) Di-n-octyl phthalate	22.462	149	2177889	35.060	ng	99
87) Indeno(1,2,3-cd)pyrene	27.803	276	2679054	38.670	ng	99
88) Benzo(b)fluoranthene	23.468	252	2636894	40.674	ng	99
89) Benzo(k)fluoranthene	23.533	252	2629220	41.215	ng	100
90) Benzo(a)pyrene	24.274	252	2177053	39.979	ng	99
91) Dibenzo(a,h)anthracene	27.862	278	2224114	38.533	ng	99
92) Benzo(g,h,i)perylene	28.856	276	2247299	38.653	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
Data File : BM049767.D
Acq On : 01 Apr 2025 10:18
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025


Quant Time: Apr 01 11:46:32 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Mar 13 14:55:34 2025

Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040125\
Data File : BM049767.D
Acq On : 01 Apr 2025 10:18
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025


Quant Time: Apr 01 11:46:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 14:55:34 2025
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.792 min Scan# 81
 Delta R.T. -0.013 min
 Lab File: BM049767.D
 Acq: 01 Apr 2025 10:18




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040



Tgt Ion:152 Resp: 337112
 Ion Ratio Lower Upper
 152 100
 150 151.7 123.8 185.8
 115 52.8 43.0 64.6




Manual Integrations
 APPROVED


Reviewed By :Anahy Claudio 04/02/2025
 Supervised By :Jagrut Upadhyay 04/02/2025



#2
 1,4-Dioxane
 Concen: 35.601 ng
 RT: 3.275 min Scan# 49
 Delta R.T. -0.012 min
 Lab File: BM049767.D
 Acq: 01 Apr 2025 10:18


Tgt Ion: 88 Resp: 307910
 Ion Ratio Lower Upper
 88 100
 58 73.1 56.7 85.1
 43 30.2 22.6 34.0





#3
Pyridine
Concen: 38.770 ng m
RT: 3.675 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 85634
Ion Ratio Lower Upper
79 100
52 66.8 50.1 75.1
51 34.3 25.2 37.8

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#4
n-Nitrosodimethylamine
Concen: 39.665 ng
RT: 3.581 min Scan# 101
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion: 42 Resp: 387653
Ion Ratio Lower Upper
42 100
74 136.2 112.6 168.8
44 6.5 5.0 7.6





#5
2-Fluorophenol
Concen: 78.689 ng
RT: 5.375 min Scan# 408
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 112 Resp: 1572130
Ion Ratio Lower Upper
112 100
64 58.3 45.3 67.9
63 31.2 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#6
Aniline
Concen: 41.031 ng
RT: 7.122 min Scan# 703
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion: 93 Resp: 861876
Ion Ratio Lower Upper
93 100
66 38.5 30.6 46.0
65 20.5 15.4 23.2





#7
Phenol-d6
Concen: 85.542 ng
RT: 6.969 min Scan# 61
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 99 Resp: 214476
Ion Ratio Lower Upper
99 100
42 17.3 13.6 20.4
71 33.3 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#8
2-Chlorophenol
Concen: 40.177 ng
RT: 7.363 min Scan# 744
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18

Tgt Ion:128 Resp: 794901
Ion Ratio Lower Upper
128 100
130 32.7 12.5 52.5
64 46.3 24.9 64.9









#9
Benzaldehyde
Concen: 35.168 ng
RT: 6.934 min Scan# 61
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 43590
Ion Ratio Lower Upper
77 100
105 97.3 75.0 115.0
106 92.6 72.3 112.3




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#10
Phenol
Concen: 41.623 ng
RT: 6.992 min Scan# 681
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion: 94 Resp: 1011917
Ion Ratio Lower Upper
94 100
65 29.2 10.4 50.4
66 35.5 17.1 57.1





#11
bis(2-Chloroethyl)ether
Concen: 41.329 ng
RT: 7.216 min Scan# 719
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 93 Resp: 813640
Ion Ratio Lower Upper
93 100
63 79.2 54.8 94.8
95 32.8 12.8 52.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#12
1,3-Dichlorobenzene
Concen: 38.624 ng
RT: 7.681 min Scan# 798
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:146 Resp: 928678
Ion Ratio Lower Upper
146 100
148 63.9 52.3 78.5
75 31.1 24.2 36.2





#13
1,4-Dichlorobenzene
Concen: 38.816 ng
RT: 7.828 min Scan# 81
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:146 Resp: 941528
Ion Ratio Lower Upper
146 100
148 66.0 52.6 79.0
111 36.8 30.8 46.2

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#14
1,2-Dichlorobenzene
Concen: 39.908 ng
RT: 8.145 min Scan# 877
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:146 Resp: 919994
Ion Ratio Lower Upper
146 100
148 65.6 50.9 76.3
111 39.6 31.0 46.6





#15
Benzyl Alcohol
Concen: 42.229 ng
RT: 8.034 min Scan# 80
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 664440
Ion Ratio Lower Upper
79 100
108 77.1 60.7 91.1
77 66.7 55.4 83.2

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 43.412 ng
RT: 8.322 min Scan# 907
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion: 45 Resp: 1004495
Ion Ratio Lower Upper
45 100
77 14.9 0.0 36.1
79 11.7 0.0 32.8





#17
2-Methylphenol
Concen: 42.684 ng
RT: 8.239 min Scan# 895
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:107 Resp: 616084
Ion Ratio Lower Upper
107 100
108 112.2 89.6 134.4
77 49.8 39.8 59.6
79 47.1 37.4 56.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#18
Hexachloroethane
Concen: 38.628 ng
RT: 8.875 min Scan# 1001
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:117 Resp: 330882
Ion Ratio Lower Upper
117 100
119 99.0 73.1 109.7
201 119.0 90.0 135.0





#19
n-Nitroso-di-n-propylamine
Concen: 47.885 ng
RT: 8.604 min Scan# 910
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 70 Resp: 696178
Ion Ratio Lower Upper
70 100
42 56.4 41.8 62.6
101 8.9 7.7 11.5
130 19.8 16.2 24.4




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#20
3+4-Methylphenols
Concen: 43.775 ng
RT: 8.569 min Scan# 949
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:107 Resp: 846734
Ion Ratio Lower Upper
107 100
108 89.8 71.0 111.0
77 33.8 15.1 55.1
79 30.2 11.2 51.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.586 min Scan# 11
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:136 Resp: 125282
Ion Ratio Lower Upper
136 100
137 10.9 8.6 12.8
54 7.9 5.8 8.8
68 5.6 4.5 6.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#22
Acetophenone
Concen: 40.380 ng
RT: 8.616 min Scan# 957
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18

Tgt Ion:105 Resp: 1304418
Ion Ratio Lower Upper
105 100
71 4.3 2.5 3.7#
51 26.7 19.9 29.9
120 22.6 18.3 27.5





#23

Nitrobenzene-d5

Concen: 81.263 ng

RT: 8.951 min Scan# 1014

Delta R.T. -0.012 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 198294

Ion Ratio Lower Upper

82 100

128 41.0 33.4 50.0





54 46.8 36.4 54.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#24

Nitrobenzene

Concen: 39.603 ng

RT: 8.992 min Scan# 1021

Delta R.T. -0.012 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18




Tgt Ion: 77 Resp: 911220


Ion Ratio Lower Upper

77 100

123 44.1 34.9 52.3

65 13.0 10.6 15.8





#25

Isophorone

Concen: 41.374 ng

RT: 9.522 min Scan# 1111

Delta R.T. -0.012 min

Lab File: BM049767.D


Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion: 82 Resp: 1550110

Ion Ratio Lower Upper

82 100

95 6.6 5.3 7.9




138 18.0 14.4 21.6

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#26

2-Nitrophenol

Concen: 40.225 ng

RT: 9.704 min Scan# 1142

Delta R.T. -0.012 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18




Tgt Ion:139 Resp: 393929


Ion Ratio Lower Upper

139 100

109 28.0 23.3 34.9


65 48.6 38.6 58.0





#27
2,4-Dimethylphenol
Concen: 39.946 ng
RT: 9.769 min Scan# 1153
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:122 Resp: 49794
Ion Ratio Lower Upper
122 100
107 114.7 94.3 141.5
121 58.8 45.4 68.2




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#28
bis(2-Chloroethoxy)methane
Concen: 41.386 ng
RT: 10.004 min Scan# 1193
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion: 93 Resp: 1092065
Ion Ratio Lower Upper
93 100
95 32.2 26.6 40.0
123 11.2 9.0 13.6





#29
2,4-Dichlorophenol
Concen: 42.057 ng
RT: 10.239 min Scan# 1233
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 85425
Ion Ratio Lower Upper
162 100
164 64.1 46.3 86.3
98 30.9 12.8 52.8

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#30
1,2,4-Trichlorobenzene
Concen: 39.367 ng
RT: 10.451 min Scan# 1269
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:180 Resp: 1000462
Ion Ratio Lower Upper
180 100
182 95.7 77.7 116.5
145 26.4 21.0 31.4





#31
Naphthalene
Concen: 39.821 ng
RT: 10.639 min Scan# 1177
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 2565330
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 12.9 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#32
Benzoic acid
Concen: 41.450 ng
RT: 9.916 min Scan# 1178
Delta R.T. 0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:122 Resp: 439432
Ion Ratio Lower Upper
122 100
105 121.5 105.6 145.6
77 99.5 76.9 116.9





#33
4-Chloroaniline
Concen: 40.822 ng
RT: 10.745 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:127 Resp: 897480
Ion Ratio Lower Upper
127 100
129 32.7 25.7 38.5
65 31.2 24.1 36.1
92 18.1 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#34
Hexachlorobutadiene
Concen: 38.371 ng
RT: 10.933 min Scan# 1351
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:225 Resp: 601986
Ion Ratio Lower Upper
225 100
223 62.3 49.8 74.8
227 64.0 52.1 78.1





#35
Caprolactam
Concen: 42.171 ng
RT: 11.539 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:113 Resp: 198870
Ion Ratio Lower Upper
113 100
55 166.0 141.8 181.8
56 124.3 98.0 138.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#36
4-Chloro-3-methylphenol
Concen: 42.804 ng
RT: 11.880 min Scan# 1512
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:107 Resp: 746125
Ion Ratio Lower Upper
107 100
144 30.3 23.0 34.6
142 89.5 69.8 104.8





#37
2-Methylnaphthalene
Concen: 42.238 ng
RT: 12.251 min Scan# 1575
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:142 Resp: 190264
Ion Ratio Lower Upper
142 100
141 88.0 70.5 105.7
115 33.3 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#38
1-Methylnaphthalene
Concen: 42.833 ng
RT: 12.474 min Scan# 1613
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:142 Resp: 1861529
Ion Ratio Lower Upper
142 100
141 92.1 73.4 110.2
116 3.5 2.8 4.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.433 min Scan# 1946
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:164 Resp: 834170
Ion Ratio Lower Upper
164 100
162 101.3 80.5 120.7
160 45.9 36.9 55.3

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.504 ng
RT: 12.621 min Scan# 1638
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:216 Resp: 1179803
Ion Ratio Lower Upper
216 100
214 78.4 63.5 95.3
179 19.9 15.7 23.5
108 15.3 12.2 18.2





#41
Hexachlorocyclopentadiene
Concen: 38.621 ng
RT: 12.610 min Scan# 1636
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:237 Resp: 438408
Ion Ratio Lower Upper
237 100
235 63.2 41.1 81.1
272 12.6 0.0 32.5

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#42
2,4,6-Tribromophenol
Concen: 86.432 ng
RT: 15.921 min Scan# 2199
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:330 Resp: 841905
Ion Ratio Lower Upper
330 100
332 95.3 77.0 115.4
141 35.3 27.8 41.6





#43
2,4,6-Trichlorophenol
Concen: 40.545 ng
RT: 12.863 min Scan# 1692
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:196 Resp: 695798
Ion Ratio Lower Upper
196 100
198 96.6 78.3 117.5
200 31.1 25.1 37.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#44
2,4,5-Trichlorophenol
Concen: 40.216 ng
RT: 12.939 min Scan# 1692
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:196 Resp: 753129
Ion Ratio Lower Upper
196 100
198 99.3 76.7 115.1
97 41.5 32.9 49.3
132 21.4 16.8 25.2





#45
2-Fluorobiphenyl
Concen: 83.015 ng
RT: 13.057 min Scan# 1712
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:172 Resp: 5529414
Ion Ratio Lower Upper
172 100
171 35.4 27.9 41.9
170 23.6 18.6 28.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#46
1,1'-Biphenyl
Concen: 39.760 ng
RT: 13.269 min Scan# 1748
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:154 Resp: 2556224
Ion Ratio Lower Upper
154 100
153 39.6 19.7 59.7
76 11.9 0.0 31.2





#47
2-Chloronaphthalene
Concen: 39.242 ng
RT: 13.310 min Scan# 1755
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 1963725
Ion Ratio Lower Upper
162 100
127 31.9 25.8 38.6
164 33.5 27.1 40.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#48
2-Nitroaniline
Concen: 40.589 ng
RT: 13.510 min Scan# 1789
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18



Tgt Ion: 65 Resp: 455968
Ion Ratio Lower Upper
65 100
92 67.4 53.4 80.2
138 109.4 87.6 131.4





#49

Acenaphthylene

Concen: 40.186 ng

RT: 14.157 min Scan# 1899

Delta R.T. -0.012 min

Lab File: BM049767.D


Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:152 Resp: 2810898

Ion Ratio Lower Upper

152 100

151 19.7 15.8 23.8




153 13.1 10.2 15.4

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#50

Dimethylphthalate

Concen: 40.471 ng

RT: 13.898 min Scan# 1855

Delta R.T. -0.006 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18




Tgt Ion:163 Resp: 2324364


Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.0

164 10.1 8.2 12.2






#51
2,6-Dinitrotoluene
Concen: 41.035 ng
RT: 14.010 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:165 Resp: 47222
Ion Ratio Lower Upper
165 100
63 49.8 42.3 63.5
89 50.1 43.9 65.9


Manual Integrations

APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#52
Acenaphthene
Concen: 40.385 ng
RT: 14.498 min Scan# 1957
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:154 Resp: 1852901
Ion Ratio Lower Upper
154 100
153 113.5 90.4 135.6
152 54.0 44.2 66.2





#53
3-Nitroaniline
Concen: 39.283 ng
RT: 14.333 min Scan# 1929
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:138 Resp: 43017
Ion Ratio Lower Upper
138 100
108 9.3 7.5 11.3
92 110.8 88.0 132.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#54
2,4-Dinitrophenol
Concen: 38.924 ng
RT: 14.539 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:184 Resp: 218039
Ion Ratio Lower Upper
184 100
63 56.9 47.2 70.8
154 57.5 49.9 74.9





#55
Dibenzenofuran
Concen: 40.393 ng
RT: 14.833 min Scan# 2016
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 308813
Ion Ratio Lower Upper
168 100
139 33.7 27.4 41.0
169 13.5 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#56
4-Nitrophenol
Concen: 36.790 ng
RT: 14.651 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:139 Resp: 350690
Ion Ratio Lower Upper
139 100
109 72.1 54.7 94.7
65 91.8 76.2 116.2





#57
2,4-Dinitrotoluene
Concen: 41.553 ng
RT: 14.792 min Scan# 2009
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 615311
Ion Ratio Lower Upper
165 100
63 40.0 30.8 46.2
89 66.4 53.1 79.7
182 2.7 2.1 3.1




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#58
Fluorene
Concen: 43.130 ng
RT: 15.480 min Scan# 2124
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:166 Resp: 2684693
Ion Ratio Lower Upper
166 100
165 95.6 77.5 116.3
167 13.7 10.9 16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 39.887 ng
RT: 15.063 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:232 Resp: 625261
Ion Ratio Lower Upper
232 100
131 38.4 30.0 45.0
130 2.0 1.6 2.4
166 29.8 22.5 33.7




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#60
Diethylphthalate
Concen: 41.056 ng
RT: 15.262 min Scan# 2087
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:149 Resp: 2220802
Ion Ratio Lower Upper
149 100
177 22.8 18.2 27.2
150 11.9 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 43.011 ng
RT: 15.480 min Scan# 2126
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:204 Resp: 146196
Ion Ratio Lower Upper
204 100
206 33.4 25.2 37.8
141 45.2 34.5 51.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#62
4-Nitroaniline
Concen: 35.986 ng
RT: 15.498 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:138 Resp: 455574
Ion Ratio Lower Upper
138 100
92 49.4 29.8 69.8
108 75.5 61.3 101.3





#63
Azobenzene
Concen: 42.285 ng
RT: 15.768 min Scan# 2173
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 2184719
Ion Ratio Lower Upper
77 100
182 25.2 8.0 48.0
105 18.3 0.0 40.0
51 28.1 7.3 47.3

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.174 min Scan# 2412
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:188 Resp: 1532241
Ion Ratio Lower Upper
188 100
94 7.4 5.8 8.6
80 8.3 6.3 9.5





#65
4,6-Dinitro-2-methylphenol
Concen: 39.734 ng
RT: 15.557 min Scan# 2137
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:198 Resp: 313084
Ion Ratio Lower Upper
198 100
51 37.3 13.8 53.8
105 38.0 14.0 54.0

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#66
n-Nitrosodiphenylamine
Concen: 41.790 ng
RT: 15.692 min Scan# 2160
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:169 Resp: 1940059
Ion Ratio Lower Upper
169 100
168 68.1 54.8 82.2
167 37.5 29.9 44.9





#67
4-Bromophenyl-phenylether
Concen: 41.099 ng
RT: 16.374 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:248 Resp: 727065
Ion Ratio Lower Upper
248 100
250 96.9 77.9 116.9
141 53.0 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#68
Hexachlorobenzene
Concen: 39.652 ng
RT: 16.486 min Scan# 2295
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18

Tgt Ion:284 Resp: 788814
Ion Ratio Lower Upper
284 100
142 33.9 25.5 38.3
249 31.3 24.2 36.2





#69

Atrazine

Concen: 31.464 ng

RT: 16.639 min Scan# 21

Delta R.T. -0.012 min

Lab File: BM049767.D


Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:	200	Resp:	32190
Ion Ratio	Lower	Upper	
200	100		
173	25.3	5.4	45.4
215	51.9	29.8	69.8

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#70

Pentachlorophenol

Concen: 39.197 ng




RT: 16.827 min Scan# 2353


Delta R.T. -0.012 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18


Tgt Ion:	266	Resp:	490034
Ion Ratio	Lower	Upper	
266	100		
268	63.5	50.2	75.2
264	63.3	51.3	76.9





#71
Phenanthrene
Concen: 40.507 ng
RT: 17.221 min Scan# 2421
Delta R.T. -0.006 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:178 Resp: 3446408
Ion Ratio Lower Upper
178 100
176 19.5 15.5 23.3
179 15.5 12.5 18.7

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#72
Anthracene
Concen: 40.524 ng
RT: 17.309 min Scan# 2435
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:178 Resp: 3339045
Ion Ratio Lower Upper
178 100
176 19.1 15.1 22.7
179 15.6 12.4 18.6





#73
 Carbazole
 Concen: 38.877 ng
 RT: 17.574 min Scan# 2482
 Delta R.T. -0.012 min
 Lab File: BM049767.D
 Acq: 01 Apr 2025 10:18




Instrument :
 BNA_M
 ClientSampleId :
 SSTCCCC040




Tgt Ion:167 Resp: 2901083
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.0 25.6
 139 11.7 9.4 14.2

Manual Integrations
 APPROVED



Reviewed By :Anahy Claudio 04/02/2025
 Supervised By :Jagrut Upadhyay 04/02/2025




#74
 Di-n-butylphthalate
 Concen: 41.139 ng
 RT: 18.151 min Scan# 2578
 Delta R.T. -0.012 min
 Lab File: BM049767.D
 Acq: 01 Apr 2025 10:18




Tgt Ion:149 Resp: 3364666
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.2 10.8
 104 4.8 3.9 5.9





#75
Fluoranthene
Concen: 38.207 ng
RT: 19.233 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 374250
Ion Ratio Lower Upper
202 100
101 8.1 0.0 27.8
203 17.5 0.0 37.6




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.409 min Scan# 3132
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18

Tgt Ion:240 Resp: 1182765
Ion Ratio Lower Upper
240 100
120 8.5 6.5 9.7
236 25.6 20.7 31.1





#77

Benzidine

Concen: 41.522 ng

RT: 19.415 min Scan# 21

Delta R.T. -0.012 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:184 Resp: 590010

Ion Ratio Lower Upper

184 100

185 14.0 11.7 17.5





183 11.5 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#78

Pyrene

Concen: 46.302 ng

RT: 19.592 min Scan# 2823

Delta R.T. -0.012 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18




Tgt Ion:202 Resp: 3826162


Ion Ratio Lower Upper

202 100

200 20.9 17.0 25.4


203 17.7 14.0 21.0





#79
 Terphenyl-d14
 Concen: 101.290 ng
 RT: 19.803 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BM049767.D
 Acq: 01 Apr 2025 10:18




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion:244 Resp: 710940
 Ion Ratio Lower Upper
 244 100
 212 6.7 5.8 8.6
 122 6.8 6.0 9.0

Manual Integrations
 APPROVED



Reviewed By :Anahy Claudio 04/02/2025
 Supervised By :Jagrut Upadhyay 04/02/2025




#80
 Butylbenzylphthalate
 Concen: 42.277 ng
 RT: 20.497 min Scan# 2977
 Delta R.T. -0.012 min
 Lab File: BM049767.D
 Acq: 01 Apr 2025 10:18




Tgt Ion:149 Resp: 1084171
 Ion Ratio Lower Upper
 149 100
 91 66.1 52.2 78.4
 206 21.5 17.6 26.4





#81
Benzo(a)anthracene
Concen: 39.939 ng
RT: 21.392 min Scan# 3116
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 3155239
Ion Ratio Lower Upper
228 100
226 27.8 22.0 33.0
229 19.8 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025



#82
3,3'-Dichlorobenzidine
Concen: 36.279 ng
RT: 21.315 min Scan# 3116
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18


Tgt Ion:252 Resp: 933134
Ion Ratio Lower Upper
252 100
254 63.3 51.0 76.4
126 9.2 6.3 9.5





#83
Chrysene
Concen: 39.070 ng
RT: 21.456 min Scan# 3142
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:228 Resp: 2925890
Ion Ratio Lower Upper
228 100
226 30.9 24.5 36.7
229 19.9 15.8 23.6

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 43.057 ng
RT: 21.327 min Scan# 3118
Delta R.T. -0.012 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Tgt Ion:149 Resp: 1720829
Ion Ratio Lower Upper
149 100
167 28.1 22.9 34.3
279 4.3 3.5 5.3





#85
Di-n-octyl phthalate
Concen: 35.060 ng
RT: 22.462 min Scan# 31
Delta R.T. -0.018 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:149 Resp: 2177889
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 9.0 6.8 10.2

Manual Integrations
APPROVED



Reviewed By :Anahy Claudio 04/02/2025
Supervised By :Jagrut Upadhyay 04/02/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.415 min Scan# 3643
Delta R.T. -0.024 min
Lab File: BM049767.D
Acq: 01 Apr 2025 10:18



Tgt Ion:264 Resp: 962538
Ion Ratio Lower Upper
264 100
260 23.2 19.6 29.4
265 21.8 18.2 27.2





#87

Indeno(1,2,3-cd)pyrene

Concen: 38.670 ng

RT: 27.803 min Scan# 41

Delta R.T. -0.042 min

Lab File: BM049767.D


Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:276 Resp: 2679054

Ion Ratio Lower Upper

276 100

138 21.3 16.3 24.5




277 24.8 19.9 29.9

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#88

Benzo(b)fluoranthene

Concen: 40.674 ng

RT: 23.468 min Scan# 3482

Delta R.T. -0.024 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18




Tgt Ion:252 Resp: 2636894


Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.0

125 8.3 5.9 8.9





#89

Benzo(k)fluoranthene

Concen: 41.215 ng

RT: 23.533 min Scan# 34

Delta R.T. -0.018 min

Lab File: BM049767.D


Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:252 Resp: 2629220

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9




125 7.8 6.2 9.2

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#90

Benzo(a)pyrene

Concen: 39.979 ng

RT: 24.274 min Scan# 3619

Delta R.T. -0.024 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18




Tgt Ion:252 Resp: 2177053


Ion Ratio Lower Upper

252 100

253 22.0 17.2 25.8

125 8.9 7.0 10.6





#91

Dibenzo(a,h)anthracene

Concen: 38.533 ng

RT: 27.862 min Scan# 41

Delta R.T. -0.042 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:278 Resp: 2224114

Ion Ratio Lower Upper

278 100

139 13.2 10.1 15.1





279 23.8 19.1 28.7

Manual Integrations

APPROVED

Reviewed By :Anahy Claudio 04/02/2025

Supervised By :Jagrut Upadhyay 04/02/2025



#92

Benzo(g,h,i)perylene

Concen: 38.653 ng

RT: 28.856 min Scan# 4398

Delta R.T. -0.047 min

Lab File: BM049767.D

Acq: 01 Apr 2025 10:18

Tgt Ion:276 Resp: 2247299

Ion Ratio Lower Upper

276 100

277 23.7 18.8 28.2

138 18.5 14.0 21.0

