

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040215\
 Data File : BM000806.D
 Acq On : 02 Apr 2015 10:22
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 03 02:07:17 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.483	-4.3	102	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
5 S	Nitrobenzene-d5	0.247	0.251	-1.6	106	0.00
6	Naphthalene	1.162	1.200	-3.3	104	0.00
7	2-Methylnaphthalene	0.734	0.760	-3.5	104	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	104	-0.02
9 S	2-Fluorobiphenyl	1.551	1.616	-4.2	105	0.00
10	Acenaphthylene	2.213	2.295	-3.7	108	0.00
11	Acenaphthene	1.533	1.534	-0.1	105	0.00
12	Fluorene	1.716	1.824	-6.3	109	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
14	Hexachlorobenzene	0.293	0.287	2.0	110	0.00
15	Pentachlorophenol	0.069	0.046	33.3#	113	0.00
16	Phenanthrene	1.370	1.328	3.1	109	0.00
17	Anthracene	1.148	1.090	5.1	108	-0.02
18	Fluoranthene	1.353	1.295	4.3	110	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
20	Pyrene	1.631	1.690	-3.6	112	0.00
21 S	Terphenyl-d14	0.602	0.627	-4.2	110	0.00
22	Benzo(a)anthracene	1.428	1.484	-3.9	117	-0.01
23	Chrysene	1.497	1.552	-3.7	110	0.00
24	Indeno(1,2,3-cd)pyrene	1.549	1.614	-4.2	114	-0.01
25 I	Perylene-d12	1.000	1.000	0.0	114	0.00
26	Benzo(b)fluoranthene	1.575	1.720	-9.2	123	-0.01
27	Benzo(k)fluoranthene	1.511	1.432	5.2	105	0.00
28 C	Benzo(a)pyrene	1.352	1.418	-4.9	120	-0.01
29	Dibenzo(a,h)anthracene	1.335	1.356	-1.6	114	-0.01
30	Benzo(g,h,i)perylene	1.441	1.453	-0.8	113	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0