

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040215\
 Data File : BM000809.D
 Acq On : 02 Apr 2015 14:12
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Quant Time: Apr 03 02:12:44 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.469	-1.3	101	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
5 S	Nitrobenzene-d5	0.247	0.246	0.4	104	0.00
6	Naphthalene	1.162	1.192	-2.6	104	0.00
7	2-Methylnaphthalene	0.734	0.755	-2.9	104	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
9 S	2-Fluorobiphenyl	1.551	1.605	-3.5	105	0.00
10	Acenaphthylene	2.213	2.239	-1.2	106	0.00
11	Acenaphthene	1.533	1.528	0.3	105	0.00
12	Fluorene	1.716	1.779	-3.7	107	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
14	Hexachlorobenzene	0.293	0.284	3.1	105	0.00
15	Pentachlorophenol	0.069	0.032	53.6#	78	0.00
16	Phenanthrene	1.370	1.369	0.1	109	0.00
17	Anthracene	1.148	1.101	4.1	105	-0.02
18	Fluoranthene	1.353	1.302	3.8	107	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
20	Pyrene	1.631	1.636	-0.3	107	0.00
21 S	Terphenyl-d14	0.602	0.621	-3.2	109	0.00
22	Benzo(a)anthracene	1.428	1.407	1.5	110	-0.01
23	Chrysene	1.497	1.526	-1.9	108	0.00
24	Indeno(1,2,3-cd)pyrene	1.549	1.577	-1.8	111	-0.01
25 I	Perylene-d12	1.000	1.000	0.0	110	0.00
26	Benzo(b)fluoranthene	1.575	1.631	-3.6	113	0.00
27	Benzo(k)fluoranthene	1.511	1.505	0.4	107	0.00
28 C	Benzo(a)pyrene	1.352	1.362	-0.7	111	-0.01
29	Dibenzo(a,h)anthracene	1.335	1.367	-2.4	111	0.00
30	Benzo(g,h,i)perylene	1.441	1.478	-2.6	111	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0