

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040315\
 Data File : BM000816.D
 Acq On : 02 Apr 2015 20:57
 Operator : TP/IZ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Quant Time: Apr 03 09:39:27 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 09:26:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	27053	5.00	ng	0.00
11) Naphthalene-d8	10.39	136	123709	5.00	ng	0.00
19) Acenaphthene-d10	14.28	164	60881	5.00	ng	0.00
26) Phenanthrene-d10	17.00	188	123846	5.00	ng	0.00
30) Chrysene-d12	21.21	240	116391	5.00	ng	0.00
34) Perylene-d12	23.40	264	104671	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.21	112	15502	2.08	ng	0.00
5) Phenol-d6	6.77	99	23587	2.27	ng	0.00
12) Nitrobenzene-d5	8.77	82	11485	2.20	ng	0.00
20) 2,4,6-Tribromophenol	15.77	330	4486	2.29	ng	0.00
32) Terphenyl-d14	19.65	244	29327	2.09	ng	0.00


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.12	88	6683	1.55	ng	91
3) n-Nitrosodimethylamine	3.43	42	7369	2.14	ng	# 38
6) 2-Chlorophenol	7.17	128	16141	2.18	ng	93
7) Phenol	6.81	94	23410	2.10	ng	93
8) bis(2-Chloroethyl)ether	7.02	93	16331	2.13	ng	89
9) 2-Methylphenol	8.06	107	16972	2.13	ng	95
10) 3+4-Methylphenols	8.38	107	19705	2.20	ng	95
13) Nitrobenzene	8.80	77	14106	2.09	ng	90
14) 2-Nitrophenol	9.52	139	6589	0.71	ng	# 90
15) 2,4-Dimethylphenol	9.58	122	14886	2.12	ng	97
16) 2,4-Dichlorophenol	10.04	162	13702	2.21	ng	95
17) Hexachlorobutadiene	10.75	225	9899	2.07	ng	100
18) 4-Chloro-3-methylphenol	11.69	107	13210	2.13	ng	98
21) 2,4,6-Trichlorophenol	12.69	196	8912	2.34	ng	99
22) 2,4,5-Trichlorophenol	12.75	196	10641	2.33	ng	95
23) 2-Fluorobiphenyl	12.87	172	39382	2.10	ng	99
24) 2,4-Dinitrophenol	14.40	184	904	1.51	ng	# 64
25) 4-Nitrophenol	14.49	139	3191	2.15	ng	87
27) 4,6-Dinitro-2-methylphenol	15.39	198	2628	2.03	ng	# 80
28) Hexachlorobenzene	16.33	284	15153	2.17	ng	97
29) Pentachlorophenol	16.67	266	3553	2.29	ng	95
31) Benzidine	19.28	184	10468	1.80	ng	# 50
33) 3,3'-Dichlorobenzidine	21.13	252	18760	2.20	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040315\
Data File : BM000816.D
Acq On : 02 Apr 2015 20:57
Operator : TP/IZ
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC002

Quant Time: Apr 03 09:39:27 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 09:26:49 2015
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.63 min Scan# 262
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Instrument :
BNA_M
ClientSampleId :
SSTDIC002




Tgt Ion: 152 Resp: 27053
Ion Ratio Lower Upper
152 100
150 123.9 99.8 149.6
115 15.9 13.5 20.3




#2

1,4-Dioxane
Concen: 1.55 ng
RT: 3.12 min Scan# 29
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Tgt Ion: 88 Resp: 6683
Ion Ratio Lower Upper
88 100
58 81.4 57.0 85.4
43 35.0 29.2 43.8









#3
 n-Nitrosodimethylamine
 Concen: 2.14 ng
 RT: 3.43 min Scan# 49
 Delta R.T. -0.02 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002


Tgt Ion: 42 Resp: 7369
 Ion Ratio Lower Upper
 42 100
 74 44.8 93.1 139.7#
 44 23.5 56.6 85.0#



#4
 2-Fluorophenol
 Concen: 2.08 ng
 RT: 5.21 min Scan# 164
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57

Tgt Ion: 112 Resp: 15502
 Ion Ratio Lower Upper
 112 100
 64 47.4 40.3 60.5
 63 21.7 19.0 28.4





#5

Phenol-d6

Concen: 2.27 ng

RT: 6.77 min Scan# 238

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002


Tgt Ion: 99 Resp: 23587

Ion Ratio Lower Upper

99 100

42 9.5 10.2 15.4#




71 32.0 28.6 43.0



Abundance Ion 99.00 (98.70 to 99.70): BM000816.D (-230) (-)

Ion 42.00 (41.70 to 42.70): BM000816.D (-230) (-)

Ion 71.00 (70.70 to 71.70): BM000816.D (-230) (-)



#6

2-Chlorophenol

Concen: 2.18 ng

RT: 7.17 min Scan# 249

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57


Tgt Ion: 128 Resp: 16141

Ion Ratio Lower Upper

128 100

130 27.3 7.6 47.6



64 35.6 23.1 63.1




Abundance Ion 128.00 (127.70 to 128.70): BM000816.D (-235) (-)

Ion 130.00 (129.70 to 130.70): BM000816.D (-235) (-)

Ion 64.00 (63.70 to 64.70): BM000816.D (-235) (-)





#7

Phenol

Concen: 2.10 ng

RT: 6.81 min Scan# 239

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002


Tgt Ion: 94 Resp: 23410

Ion Ratio Lower Upper

94 100

65 9.2 0.0 32.0




66 14.4 0.0 37.6



Abundance Ion 94.00 (93.70 to 94.70): BM000816.D (-231) (-)

Ion 65.00 (64.70 to 65.70): BM000816.D (-231) (-)

Ion 66.00 (65.70 to 66.70): BM000816.D (-231) (-)



#8

bis(2-Chloroethyl)ether

Concen: 2.13 ng

RT: 7.02 min Scan# 245

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57


Tgt Ion: 93 Resp: 16331

Ion Ratio Lower Upper

93 100

63 125.8 118.9 158.9



95 0.0 0.0 20.0




Abundance Ion 93.00 (92.70 to 93.70): BM000816.D (-237) (-)

Ion 63.00 (62.70 to 63.70): BM000816.D (-237) (-)

Ion 95.00 (94.70 to 95.70): BM000816.D (-237) (-)









#9
2-Methylphenol
Concen: 2.13 ng
RT: 8.06 min Scan# 274
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Instrument :
BNA_M
ClientSampleId :
SSTDICC002

Tgt Ion:	107	Resp:	16972
Ion	Ratio	Lower	Upper
107	100		
108	118.0	92.0	138.0
77	21.2	20.6	31.0
79	24.4	22.9	34.3




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BM0
Ion 79.00 (78.70 to 79.70): BM0






#10
3+4-Methylphenols
Concen: 2.20 ng
RT: 8.38 min Scan# 283
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Tgt Ion:	107	Resp:	19705
Ion	Ratio	Lower	Upper
107	100		
108	86.5	62.6	102.6
77	28.9	12.8	52.8
79	27.6	10.3	50.3



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BM0
Ion 79.00 (78.70 to 79.70): BM0





#11

Naphthalene-d8

Concen: 5.00 ng

RT: 10.39 min Scan# 350

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002

Tgt Ion:136 Resp: 123709


Ion Ratio Lower Upper

136 100

137 11.5 9.1 13.7

54 2.7 2.2 3.4

68 2.7 2.2 3.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BM

Ion 68.00 (67.70 to 68.70): BM



#12

Nitrobenzene-d5

Concen: 2.20 ng

RT: 8.77 min Scan# 294

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57


Tgt Ion: 82 Resp: 11485

Ion Ratio Lower Upper

82 100

128 101.3 72.5 108.7



54 10.3 11.8 17.8#




Abundance Ion 82.00 (81.70 to 82.70): BM

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BM









#13
Nitrobenzene
Concen: 2.09 ng
RT: 8.80 min Scan# 295
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57




Instrument :
BNA_M
ClientSampleId :
SSTDICC002


Tgt Ion: 77 Resp: 14106
Ion Ratio Lower Upper
77 100
123 65.3 45.8 68.6
65 8.5 8.2 12.4



#14
2-Nitrophenol
Concen: 0.71 ng
RT: 9.52 min Scan# 318
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Tgt Ion: 139 Resp: 6589
Ion Ratio Lower Upper
139 100
109 15.1 15.0 22.6
65 12.5 14.2 21.4#









#15
 2,4-Dimethylphenol
 Concen: 2.12 ng
 RT: 9.58 min Scan# 320
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002


Tgt Ion:122 Resp: 14886
 Ion Ratio Lower Upper
 122 100
 107 94.4 78.2 117.4
 121 55.4 45.4 68.0



#16
 2,4-Dichlorophenol
 Concen: 2.21 ng
 RT: 10.04 min Scan# 337
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57

Tgt Ion:162 Resp: 13702
 Ion Ratio Lower Upper
 162 100
 164 58.4 37.1 77.1
 98 33.9 19.5 59.5









#17
Hexachlorobutadiene
Concen: 2.07 ng
RT: 10.75 min Scan# 363
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57




Instrument :
BNA_M
ClientSampleId :
SSTDICC002


Tgt Ion: 225 Resp: 9899
Ion Ratio Lower Upper
225 100
223 50.8 40.8 61.2
227 79.0 62.8 94.2



#18
4-Chloro-3-methylphenol
Concen: 2.13 ng
RT: 11.69 min Scan# 408
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Tgt Ion: 107 Resp: 13210
Ion Ratio Lower Upper
107 100
144 33.0 25.8 38.8
142 91.1 71.3 106.9





#19

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.28 min Scan# 563

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002


Tgt Ion:164 Resp: 60881

Ion Ratio Lower Upper

164 100

162 92.5 74.2 111.2




160 36.7 30.4 45.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#20

2,4,6-Tribromophenol

Concen: 2.29 ng

RT: 15.77 min Scan# 638

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57


Tgt Ion:330 Resp: 4486

Ion Ratio Lower Upper

330 100

332 103.2 78.7 118.1



141 0.0 0.0 0.0




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E





#21

2,4,6-Trichlorophenol

Concen: 2.34 ng

RT: 12.69 min Scan# 466

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002


Tgt Ion:196 Resp: 8912

Ion Ratio Lower Upper

196 100

198 74.7 59.5 89.3




200 20.9 17.4 26.0



Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#22

2,4,5-Trichlorophenol

Concen: 2.33 ng

RT: 12.75 min Scan# 470

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Tgt Ion:196 Resp: 10641

Ion Ratio Lower Upper

196 100

198 76.4 57.9 86.9

97 0.0 0.0 0.0




132 0.0 0.0 0.0


Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BM

Ion 132.00 (131.70 to 132.70): E









#23
2-Fluorobiphenyl
Concen: 2.10 ng
RT: 12.87 min Scan# 477
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57




Instrument :
BNA_M
ClientSampleId :
SSTDICC002


Tgt Ion:172 Resp: 39382
Ion Ratio Lower Upper
172 100
171 37.7 30.6 46.0
170 27.1 22.2 33.2



#24
2,4-Dinitrophenol
Concen: 1.51 ng
RT: 14.40 min Scan# 571
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Tgt Ion:184 Resp: 904
Ion Ratio Lower Upper
184 100
63 42.4 62.7 94.1#
154 51.6 63.6 95.4#









#25
4-Nitrophenol
Concen: 2.15 ng
RT: 14.49 min Scan# 577
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Instrument :
BNA_M
ClientSampleId :
SSTDIC002

Tgt Ion:139 Resp: 3191
Ion Ratio Lower Upper
139 100
109 68.1 58.2 98.2
65 95.8 90.4 130.4




Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BM






#26
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.00 min Scan# 686
Delta R.T. -0.00 min
Lab File: BM000816.D
Acq: 02 Apr 2015 20:57

Tgt Ion:188 Resp: 123846
Ion Ratio Lower Upper
188 100
94 17.1 13.0 19.6
80 19.5 14.7 22.1



Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BM
Ion 80.00 (79.70 to 80.70): BM









#27
 4,6-Dinitro-2-methylphenol
 Concen: 2.03 ng
 RT: 15.39 min Scan# 623
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Tgt Ion:198 Resp: 2628
 Ion Ratio Lower Upper
 198 100
 51 124.2 139.3 179.3#
 105 97.0 88.8 128.8




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 51.00 (50.70 to 51.70): BM
 Ion 105.00 (104.70 to 105.70): E






#28
 Hexachlorobenzene
 Concen: 2.17 ng
 RT: 16.33 min Scan# 660
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57

Tgt Ion:284 Resp: 15153
 Ion Ratio Lower Upper
 284 100
 142 6.8 6.5 9.7
 249 14.5 12.6 18.8



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E





#29

Pentachlorophenol

Concen: 2.29 ng

RT: 16.67 min Scan# 673

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002


Tgt Ion:266 Resp: 3553

Ion Ratio Lower Upper

266 100

268 59.5 51.4 77.0




264 68.4 57.1 85.7



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#30

Chrysene-d12

Concen: 5.00 ng

RT: 21.21 min Scan# 970

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57


Tgt Ion:240 Resp: 116391

Ion Ratio Lower Upper

240 100

120 13.8 11.3 16.9



236 26.5 21.1 31.7




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#31

Benzidine

Concen: 1.80 ng

RT: 19.28 min Scan# 823

Delta R.T. -0.02 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57

Instrument :

BNA_M

ClientSampleId :

SSTDICC002


Tgt Ion:184 Resp: 10468

Ion Ratio Lower Upper

184 100

185 15.7 39.1 58.7#




183 11.6 36.1 54.1#



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#32

Terphenyl-d14

Concen: 2.09 ng

RT: 19.65 min Scan# 854

Delta R.T. -0.00 min

Lab File: BM000816.D

Acq: 02 Apr 2015 20:57


Tgt Ion:244 Resp: 29327

Ion Ratio Lower Upper

244 100

212 10.0 8.0 12.0



122 18.9 14.7 22.1




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#33
 3,3'-Dichlorobenzidine
 Concen: 2.20 ng
 RT: 21.13 min Scan# 965
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC002

Tgt Ion: 252 Resp: 18760
 Ion Ratio Lower Upper
 252 100
 254 55.9 44.5 66.7
 126 15.6 13.6 20.4



#34
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.40 min Scan# 1112
 Delta R.T. -0.00 min
 Lab File: BM000816.D
 Acq: 02 Apr 2015 20:57

Tgt Ion: 264 Resp: 104671
 Ion Ratio Lower Upper
 264 100
 260 23.9 19.1 28.7
 265 22.2 17.5 26.3

