

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040315\
 Data File : BM000817.D
 Acq On : 02 Apr 2015 21:33
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: Apr 03 09:39:41 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 09:26:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	26705	5.00	ng	0.00
11) Naphthalene-d8	10.39	136	127942	5.00	ng	0.00
19) Acenaphthene-d10	14.28	164	64893	5.00	ng	0.00
26) Phenanthrene-d10	17.00	188	128339	5.00	ng	0.00
30) Chrysene-d12	21.21	240	126271	5.00	ng	0.00
34) Perylene-d12	23.40	264	112992	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.21	112	38072	5.18	ng	0.00
5) Phenol-d6	6.77	99	61603	6.01	ng	0.00
12) Nitrobenzene-d5	8.77	82	28502	5.29	ng	0.00
20) 2,4,6-Tribromophenol	15.77	330	13691	6.56	ng	0.00
32) Terphenyl-d14	19.65	244	76181	5.01	ng	0.00


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.12	88	14961	3.51	ng	# 86
3) n-Nitrosodimethylamine	3.43	42	18198	5.36	ng	# 74
6) 2-Chlorophenol	7.17	128	41292	5.66	ng	89
7) Phenol	6.81	94	58851	5.34	ng	89
8) bis(2-Chloroethyl)ether	7.02	93	43173	5.71	ng	# 75
9) 2-Methylphenol	8.06	107	43572	5.54	ng	# 93
10) 3+4-Methylphenols	8.38	107	53334	6.05	ng	93
13) Nitrobenzene	8.80	77	36416	5.23	ng	# 77
14) 2-Nitrophenol	9.52	139	17071	1.79	ng	# 84
15) 2,4-Dimethylphenol	9.58	122	39762	5.48	ng	95
16) 2,4-Dichlorophenol	10.04	162	37518	5.84	ng	93
17) Hexachlorobutadiene	10.75	225	23820	4.82	ng	98
18) 4-Chloro-3-methylphenol	11.67	107	35717	5.56	ng	# 81
21) 2,4,6-Trichlorophenol	12.69	196	26323	6.48	ng	98
22) 2,4,5-Trichlorophenol	12.75	196	30734	6.32	ng	95
23) 2-Fluorobiphenyl	12.87	172	97055	4.85	ng	99
24) 2,4-Dinitrophenol	14.40	184	4375	6.87	ng	# 44
25) 4-Nitrophenol	14.49	139	12231	7.73	ng	# 76
27) 4,6-Dinitro-2-methylphenol	14.93	198	1236	0.92	ng	91
28) Hexachlorobenzene	16.33	284	37494	5.18	ng	# 95
29) Pentachlorophenol	16.67	266	13145	8.19	ng	94
31) Benzidine	19.28	184	48751	7.72	ng	# 48
33) 3,3'-Dichlorobenzidine	21.13	252	58375	6.32	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM040315\
Data File : BM000817.D
Acq On : 02 Apr 2015 21:33
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Quant Time: Apr 03 09:39:41 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 09:26:49 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.63 min Scan# 262

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005


Tgt Ion:152 Resp: 26705

Ion Ratio Lower Upper

152 100

150 127.0 99.8 149.6




115 16.2 13.5 20.3



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 3.51 ng

RT: 3.12 min Scan# 29

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion: 88 Resp: 14961

Ion Ratio Lower Upper

88 100

58 88.3 57.0 85.4#



43 36.1 29.2 43.8




Abundance Ion 88.00 (87.70 to 88.70): BM0

Ion 58.00 (57.70 to 58.70): BM0

Ion 43.00 (42.70 to 43.70): BM0





#3

n-Nitrosodimethylamine

Concen: 5.36 ng

RT: 3.43 min Scan# 49

Delta R.T. -0.02 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005


Tgt Ion: 42 Resp: 18198

Ion Ratio Lower Upper

42 100

74 116.3 93.1 139.7




44 14.4 56.6 85.0#



Abundance Ion 42.00 (41.70 to 42.70): BM000817.D (-31) (-)

Ion 74.00 (73.70 to 74.70): BM000817.D (-31) (-)

Ion 44.00 (43.70 to 44.70): BM000817.D (-31) (-)



#4

2-Fluorophenol

Concen: 5.18 ng

RT: 5.21 min Scan# 164

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion: 112 Resp: 38072

Ion Ratio Lower Upper

112 100

64 46.0 40.3 60.5



63 20.6 19.0 28.4




Abundance Ion 112.00 (111.70 to 112.70): BM000817.D (-145) (-)

Ion 64.00 (63.70 to 64.70): BM000817.D (-145) (-)

Ion 63.00 (62.70 to 63.70): BM000817.D (-145) (-)





#5

Phenol-d6

Concen: 6.01 ng

RT: 6.77 min Scan# 238

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005


Tgt Ion: 99 Resp: 61603

Ion Ratio Lower Upper

99 100

42 8.0 10.2 15.4#




71 30.7 28.6 43.0



Abundance Ion 99.00 (98.70 to 99.70): BM000817.D (-236) (-)

Ion 42.00 (41.70 to 42.70): BM000817.D (-236) (-)

Ion 71.00 (70.70 to 71.70): BM000817.D (-236) (-)



#6

2-Chlorophenol

Concen: 5.66 ng

RT: 7.17 min Scan# 249

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion: 128 Resp: 41292

Ion Ratio Lower Upper

128 100

130 27.1 7.6 47.6



64 31.9 23.1 63.1




Abundance Ion 128.00 (127.70 to 128.70): BM000817.D (-246) (-)

Ion 130.00 (129.70 to 130.70): BM000817.D (-246) (-)

Ion 64.00 (63.70 to 64.70): BM000817.D (-246) (-)





#7

Phenol

Concen: 5.34 ng

RT: 6.81 min Scan# 239

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005


Tgt Ion: 94 Resp: 58851

Ion Ratio Lower Upper

94 100

65 7.7 0.0 32.0




66 12.8 0.0 37.6



Abundance Ion 94.00 (93.70 to 94.70): BM000817.D (-231) (-)

Ion 65.00 (64.70 to 65.70): BM000817.D (-231) (-)

Ion 66.00 (65.70 to 66.70): BM000817.D (-231) (-)



#8

bis(2-Chloroethyl)ether

Concen: 5.71 ng

RT: 7.02 min Scan# 245

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion: 93 Resp: 43173

Ion Ratio Lower Upper

93 100

63 109.1 118.9 158.9#



95 0.0 0.0 20.0




Abundance Ion 93.00 (92.70 to 93.70): BM000817.D (-237) (-)

Ion 63.00 (62.70 to 63.70): BM000817.D (-237) (-)

Ion 95.00 (94.70 to 95.70): BM000817.D (-237) (-)





#9

2-Methylphenol

Concen: 5.54 ng

RT: 8.06 min Scan# 274

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005

Tgt Ion:107 Resp: 43572





Ion Ratio Lower Upper

107 100

108 119.4 92.0 138.0

77 19.2 20.6 31.0#

79 22.7 22.9 34.3#



#10

3+4-Methylphenols

Concen: 6.05 ng

RT: 8.38 min Scan# 283

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Tgt Ion:107 Resp: 53334




Ion Ratio Lower Upper


107 100

108 86.4 62.6 102.6

77 25.9 12.8 52.8

79 25.7 10.3 50.3





#11

Naphthalene-d8

Concen: 5.00 ng

RT: 10.39 min Scan# 350

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005

Tgt Ion: 136 Resp: 127942


Ion Ratio Lower Upper

136 100

137 11.4 9.1 13.7

54 2.8 2.2 3.4

68 2.8 2.2 3.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BM

Ion 68.00 (67.70 to 68.70): BM



#12

Nitrobenzene-d5

Concen: 5.29 ng

RT: 8.77 min Scan# 294

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion: 82 Resp: 28502

Ion Ratio Lower Upper

82 100

128 120.8 72.5 108.7#



54 8.0 11.8 17.8#




Abundance Ion 82.00 (81.70 to 82.70): BM

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BM









#13
 Nitrobenzene
 Concen: 5.23 ng
 RT: 8.80 min Scan# 295
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Tgt Ion: 77 Resp: 36416
 Ion Ratio Lower Upper
 77 100
 123 76.6 45.8 68.6#
 65 7.7 8.2 12.4#



#14
 2-Nitrophenol
 Concen: 1.79 ng
 RT: 9.52 min Scan# 318
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33

Tgt Ion: 139 Resp: 17071
 Ion Ratio Lower Upper
 139 100
 109 13.1 15.0 22.6#
 65 9.6 14.2 21.4#





#15

2,4-Dimethylphenol

Concen: 5.48 ng

RT: 9.58 min Scan# 320

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDICC005


Tgt Ion:122 Resp: 39762

Ion Ratio Lower Upper

122 100

107 91.7 78.2 117.4




121 54.9 45.4 68.0



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#16

2,4-Dichlorophenol

Concen: 5.84 ng

RT: 10.04 min Scan# 337

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion:162 Resp: 37518

Ion Ratio Lower Upper

162 100

164 57.9 37.1 77.1



98 30.0 19.5 59.5




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BM





#17

Hexachlorobutadiene

Concen: 4.82 ng

RT: 10.75 min Scan# 363

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005


Tgt Ion: 225 Resp: 23820

Ion Ratio Lower Upper

225 100

223 49.5 40.8 61.2




227 79.8 62.8 94.2



Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#18

4-Chloro-3-methylphenol

Concen: 5.56 ng

RT: 11.67 min Scan# 407

Delta R.T. -0.02 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33


Tgt Ion: 107 Resp: 35717

Ion Ratio Lower Upper

107 100

144 19.7 25.8 38.8#



142 72.1 71.3 106.9




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#19

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.28 min Scan# 563

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33

Instrument :

BNA_M

ClientSampleId :

SSTDIC005





Tgt Ion:164 Resp: 64893

Ion Ratio Lower Upper

164 100

162 93.0 74.2 111.2

160 37.1 30.4 45.6



#20

2,4,6-Tribromophenol

Concen: 6.56 ng

RT: 15.77 min Scan# 638

Delta R.T. -0.00 min

Lab File: BM000817.D

Acq: 02 Apr 2015 21:33




Tgt Ion:330 Resp: 13691


Ion Ratio Lower Upper

330 100

332 103.7 78.7 118.1

141 0.0 0.0 0.0









#21
 2,4,6-Trichlorophenol
 Concen: 6.48 ng
 RT: 12.69 min Scan# 466
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Tgt Ion:196 Resp: 26323
 Ion Ratio Lower Upper
 196 100
 198 73.5 59.5 89.3
 200 19.9 17.4 26.0



#22
 2,4,5-Trichlorophenol
 Concen: 6.32 ng
 RT: 12.75 min Scan# 470
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33

Tgt Ion:196 Resp: 30734
 Ion Ratio Lower Upper
 196 100
 198 76.8 57.9 86.9
 97 0.0 0.0 0.0
 132 0.0 0.0 0.0









#23
2-Fluorobiphenyl
Concen: 4.85 ng
RT: 12.87 min Scan# 477
Delta R.T. -0.00 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33




Instrument :
BNA_M
ClientSampleId :
SSTDIC005


Tgt Ion:172 Resp: 97055
Ion Ratio Lower Upper
172 100
171 37.7 30.6 46.0
170 27.0 22.2 33.2



#24
2,4-Dinitrophenol
Concen: 6.87 ng
RT: 14.40 min Scan# 571
Delta R.T. -0.00 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33

Tgt Ion:184 Resp: 4375
Ion Ratio Lower Upper
184 100
63 24.1 62.7 94.1#
154 35.8 63.6 95.4#









#25
4-Nitrophenol
Concen: 7.73 ng
RT: 14.49 min Scan# 577
Delta R.T. -0.00 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33




Instrument :
BNA_M
ClientSampleId :
SSTDIC005


Tgt Ion:139 Resp: 12231
Ion Ratio Lower Upper
139 100
109 62.7 58.2 98.2
65 80.1 90.4 130.4#



#26
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.00 min Scan# 686
Delta R.T. -0.00 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33

Tgt Ion:188 Resp: 128339
Ion Ratio Lower Upper
188 100
94 17.3 13.0 19.6
80 19.8 14.7 22.1









#27
4,6-Dinitro-2-methylphenol
Concen: 0.92 ng
RT: 14.93 min Scan# 605
Delta R.T. -0.46 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33

Instrument :
BNA_M
ClientSampleId :
SSTDIC005

Tgt Ion:198 Resp: 1236
Ion Ratio Lower Upper
198 100
51 168.0 139.3 179.3
105 95.0 88.8 128.8




Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BM0
Ion 105.00 (104.70 to 105.70): E






#28
Hexachlorobenzene
Concen: 5.18 ng
RT: 16.33 min Scan# 660
Delta R.T. -0.00 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33

Tgt Ion:284 Resp: 37494
Ion Ratio Lower Upper
284 100
142 6.4 6.5 9.7#
249 13.8 12.6 18.8



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#29
 Pentachlorophenol
 Concen: 8.19 ng
 RT: 16.67 min Scan# 673
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Tgt Ion: 266 Resp: 13145
 Ion Ratio Lower Upper
 266 100
 268 58.1 51.4 77.0
 264 67.6 57.1 85.7



#30
 Chrysene-d12
 Concen: 5.00 ng
 RT: 21.21 min Scan# 970
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33

Tgt Ion: 240 Resp: 126271
 Ion Ratio Lower Upper
 240 100
 120 14.4 11.3 16.9
 236 26.8 21.1 31.7









#31
Benzidine
Concen: 7.72 ng
RT: 19.28 min Scan# 823
Delta R.T. -0.02 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:184 Resp: 48751
Ion Ratio Lower Upper
184 100
185 14.8 39.1 58.7#
183 9.8 36.1 54.1#



#32
Terphenyl-d14
Concen: 5.01 ng
RT: 19.65 min Scan# 854
Delta R.T. -0.00 min
Lab File: BM000817.D
Acq: 02 Apr 2015 21:33

Tgt Ion:244 Resp: 76181
Ion Ratio Lower Upper
244 100
212 9.8 8.0 12.0
122 18.6 14.7 22.1









#33
 3,3'-Dichlorobenzidine
 Concen: 6.32 ng
 RT: 21.13 min Scan# 965
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Tgt Ion: 252 Resp: 58375
 Ion Ratio Lower Upper
 252 100
 254 55.2 44.5 66.7
 126 14.9 13.6 20.4



#34
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.40 min Scan# 1112
 Delta R.T. -0.00 min
 Lab File: BM000817.D
 Acq: 02 Apr 2015 21:33

Tgt Ion: 264 Resp: 112992
 Ion Ratio Lower Upper
 264 100
 260 23.7 19.1 28.7
 265 22.3 17.5 26.3

