

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040315\
 Data File : BM000817.D
 Acq On : 02 Apr 2015 21:33
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: Apr 03 18:17:37 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 17:58:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	26705	5.00	ng	0.00
11) Naphthalene-d8	10.39	136	127942	5.00	ng	0.00
19) Acenaphthene-d10	14.28	164	64893	5.00	ng	0.00
26) Phenanthrene-d10	17.00	188	128339	5.00	ng	0.00
30) Chrysene-d12	21.21	240	126271	5.00	ng	0.00
34) Perylene-d12	23.40	264	112992	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.21	112	38072	5.18	ng	0.00
5) Phenol-d6	6.77	99	61603	6.01	ng	0.00
12) Nitrobenzene-d5	8.77	82	28502	4.77	ng	0.00
20) 2,4,6-Tribromophenol	15.77	330	13691	7.08	ng	0.00
23) 2-Fluorobiphenyl	12.87	172	97055	4.85	ng	0.00
32) Terphenyl-d14	19.65	244	76181	5.01	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.43	42	18198	5.36	ng	# 74
6) 2-Chlorophenol	7.17	128	41292	5.66	ng	89
7) Phenol	6.81	94	58851	5.34	ng	89
8) bis(2-Chloroethyl)ether	7.02	93	43173	5.63	ng	# 75
9) 2-Methylphenol	8.06	107	43572	5.54	ng	# 93
10) 3+4-Methylphenols	8.38	107	53334	6.41	ng	93
13) Nitrobenzene	8.80	77	36416	5.23	ng	# 77
14) 2-Nitrophenol	9.52	139	17071	5.36	ng	# 84
15) 2,4-Dimethylphenol	9.58	122	39762	5.48	ng	95
16) 2,4-Dichlorophenol	10.04	162	37518	5.84	ng	93
17) Hexachlorobutadiene	10.75	225	23820	4.82	ng	98
18) 4-Chloro-3-methylphenol	11.67	107	35717	5.56	ng	# 81
24) 2,4-Dinitrophenol	14.40	184	4375	6.76	ng	# 44
25) 4-Nitrophenol	14.49	139	12231	9.59	ng	# 76
27) 4,6-Dinitro-2-methylphenol	15.39	198	9580m	8.36	ng	
28) Hexachlorobenzene	16.33	284	37494	5.18	ng	# 95
29) Pentachlorophenol	16.67	266	13145	10.49	ng	94
31) Benzdine	19.28	184	48751	9.57	ng	# 48
33) 3,3'-Dichlorobenzidine	21.13	252	58375	6.88	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040315\
Data File : BM000817.D
Acq On : 02 Apr 2015 21:33
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Quant Time: Apr 03 18:17:37 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 17:58:44 2015
Response via : Initial Calibration

