

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040315\
 Data File : BM000837.D
 Acq On : 03 Apr 2015 11:18
 Operator : TP/IZ
 Sample : MDL-06-W
 Misc : MDL-WATER-0.1/0.2
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-06-W

Quant Time: Apr 03 18:50:35 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 18:16:53 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	25286	5.00	ng	0.00
11) Naphthalene-d8	10.39	136	119113	5.00	ng	0.00
19) Acenaphthene-d10	14.28	164	60733	5.00	ng	0.00
26) Phenanthrene-d10	17.00	188	126049	5.00	ng	0.00
30) Chrysene-d12	21.21	240	113150	5.00	ng	0.00
34) Perylene-d12	23.40	264	102751	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.21	112	86563	12.43	ng	0.00
5) Phenol-d6	6.77	99	146990	11.00	ng	0.00
12) Nitrobenzene-d5	8.77	82	40338	7.26	ng	0.00
20) 2,4,6-Tribromophenol	15.77	330	22656	8.82	ng	0.00
23) 2-Fluorobiphenyl	12.87	172	172703	9.23	ng	0.00
32) Terphenyl-d14	19.65	244	137221	10.08	ng	0.00

Target Compounds						Qvalue
6) 2-Chlorophenol	7.17	128	763	0.28	ng	85
7) Phenol	6.81	94	938	0.09	ng	64
8) bis(2-Chloroethyl)ether	7.02	93	626m	0.28	ng	
9) 2-Methylphenol	8.06	107	577	0.23	ng	# 76
10) 3+4-Methylphenols	8.38	107	564	0.18	ng	# 72
13) Nitrobenzene	8.80	77	684	0.11	ng	# 80
14) 2-Nitrophenol	9.52	139	211	0.20	ng	# 4
15) 2,4-Dimethylphenol	9.58	122	420	0.06	ng	# 77
16) 2,4-Dichlorophenol	10.07	162	430	0.27	ng	# 77
17) Hexachlorobutadiene	10.72	225	395	0.09	ng	# 69
18) 4-Chloro-3-methylphenol	11.69	107	442	0.07	ng	# 89
21) 2,4,6-Trichlorophenol	12.70	196	193	0.17	ng	# 61
22) 2,4,5-Trichlorophenol	12.77	196	274	0.11	ng	88
25) 4-Nitrophenol	14.56	139	72m	0.27	ng	
28) Hexachlorobenzene	16.33	284	644	0.15	ng	# 71
31) Benzidine	19.29	184	688	0.30	ng	90
33) 3,3'-Dichlorobenzidine	21.14	252	626	0.25	ng	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040315\
Data File : BM000837.D
Acq On : 03 Apr 2015 11:18
Operator : TP/IZ
Sample : MDL-06-W
Misc : MDL-WATER-0.1/0.2
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-06-W

Quant Time: Apr 03 18:50:35 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 03 18:16:53 2015
Response via : Initial Calibration

