

Data Path : Z:\HPCHEM1\BNA M\Data\BM040315\
 Data File : BM000841.D
 Acq On : 03 Apr 2015 18:48
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 04 07:18:45 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 18:16:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.756	0.773	-2.2	105	0.00
3	n-Nitrosodimethylamine	0.635	0.711	-12.0	106	-0.02
4 S	2-Fluorophenol	1.377	1.613	-17.1	110	0.00
5 S	Phenol-d6	1.921	2.447	-27.4#	119	0.00
6	2-Chlorophenol	1.365	1.810	-32.6#	126	0.00
7 C	Phenol	2.063	2.192	-6.3	103	0.00
8	bis(2-Chloroethyl)ether	1.435	1.774	-23.6	124	0.00
9	2-Methylphenol	1.472	1.656	-12.5	106	0.00
10	3+4-Methylphenols	1.558	1.982	-27.2#	111	0.00
11 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
12 S	Nitrobenzene-d5	0.233	0.231	0.9	108	0.00
13	Nitrobenzene	0.272	0.257	5.5	101	0.00
14	2-Nitrophenol	0.142	0.138	2.8	109	0.00
15	2,4-Dimethylphenol	0.283	0.297	-4.9	108	0.00
16 C	2,4-Dichlorophenol	0.251	0.277	-10.4	113	0.00
17 C	Hexachlorobutadiene	0.193	0.209	-8.3	111	-0.03
18	4-Chloro-3-methylphenol	0.251	0.270	-7.6	114	0.00
19 I	Acenaphthene-d10	1.000	1.000	0.0	109	-0.02
20 S	2,4,6-Tribromophenol	0.149	0.156	-4.7	104	0.00
21	2,4,6-Trichlorophenol	0.270	0.306	-13.3	103	0.00
22	2,4,5-Trichlorophenol	0.331	0.386	-16.6	107	0.00
23 S	2-Fluorobiphenyl	1.540	1.578	-2.5	110	0.00
24 P	2,4-Dinitrophenol	0.049	0.000#	100.0#	0#	-14.40#
25	4-Nitrophenol	0.098	0.088	10.2	102	0.00
26 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
27	4,6-Dinitro-2-methylphenol	0.044	0.046	-4.5	116	0.00
28	Hexachlorobenzene	0.282	0.244	13.5	100	0.00
29 C	Pentachlorophenol	0.049	0.039	20.4#	88	0.00
30 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
31	Benzidine	0.201	0.327	-62.7#	1370#	-0.02
32 S	Terphenyl-d14	0.602	0.622	-3.3	108	0.00
33	3,3'-Dichlorobenzidine	0.336	0.381	-13.4	144	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 1