

Data Path : Z:\HPCHEM1\BNA M\Data\BM040315\
 Data File : BM000841.D
 Acq On : 03 Apr 2015 18:48
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 04 07:18:45 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 03 18:16:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	95	0.00
2	1,4-Dioxane	1.000	1.169	-16.9	105	0.00
3	n-Nitrosodimethylamine	1.000	1.119	-11.9	106	-0.02
4 S	2-Fluorophenol	1.000	1.171	-17.1	110	0.00
5 S	Phenol-d6	1.000	1.149	-14.9	119	0.00
6	2-Chlorophenol	1.000	1.225	-22.5	126	0.00
7 C	Phenol	1.000	1.063	-6.3	103	0.00
8	bis(2-Chloroethyl)ether	1.000	1.170	-17.0	124	0.00
9	2-Methylphenol	1.000	1.073	-7.3	106	0.00
10	3+4-Methylphenols	1.000	1.101	-10.1	111	0.00
11 I	Naphthalene-d8	5.000	5.000	0.0	105	0.00
12 S	Nitrobenzene-d5	1.000	0.992	0.8	108	0.00
13	Nitrobenzene	1.000	0.942	5.8	101	0.00
14	2-Nitrophenol	1.000	1.039	-3.9	109	0.00
15	2,4-Dimethylphenol	1.000	1.049	-4.9	108	0.00
16 C	2,4-Dichlorophenol	1.000	1.044	-4.4	113	0.00
17 C	Hexachlorobutadiene	1.000	1.083	-8.3	111	-0.03
18	4-Chloro-3-methylphenol	1.000	1.074	-7.4	114	0.00
19 I	Acenaphthene-d10	5.000	5.000	0.0	109	-0.02
20 S	2,4,6-Tribromophenol	1.000	0.904	9.6	104	0.00
21	2,4,6-Trichlorophenol	1.000	0.930	7.0	103	0.00
22	2,4,5-Trichlorophenol	1.000	1.046	-4.6	107	0.00
23 S	2-Fluorobiphenyl	1.000	1.024	-2.4	110	0.00
24 P	2,4-Dinitrophenol	1.000	0.000	100.0#	0	-14.40#
25	4-Nitrophenol	1.000	0.929	7.1	102	0.00
26 I	Phenanthrene-d10	5.000	5.000	0.0	118	0.00
27	4,6-Dinitro-2-methylphenol	1.000	0.933	6.7	116	0.00
28	Hexachlorobenzene	1.000	0.853	14.7	100	0.00
29 C	Pentachlorophenol	1.000	0.843	15.7	88	0.00
30 I	Chrysene-d12	5.000	5.000	0.0	107	0.00
31	Benzidine	1.000	1.795	-79.5#	1370	-0.02
32 S	Terphenyl-d14	1.000	1.033	-3.3	108	0.00
33	3,3'-Dichlorobenzidine	1.000	1.001	-0.1	144	0.00
34 I	Perylene-d12	5.000	5.000	0.0	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0