

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040417\  
 Data File : BE092886.D  
 Acq On : 4 Apr 2017 15:58  
 Operator : SJ/MA  
 Sample : SSTDICV0.4  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 ICVBE040417

Quant Time: Apr 04 16:31:42 2017  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040417.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Apr 04 16:08:03 2017  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.77  | 152  | 8366     | 5.00 | ng    | -0.02    |
| 7) Naphthalene-d8         | 10.54 | 136  | 36405    | 5.00 | ng    | 0.00     |
| 12) Acenaphthene-d10      | 14.38 | 164  | 17942    | 5.00 | ng    | -0.02    |
| 18) Phenanthrene-d10      | 17.13 | 188  | 42879    | 5.00 | ng    | -0.02    |
| 24) Chrysene-d12          | 21.28 | 240  | 48715    | 5.00 | ng    | -0.02    |
| 31) Perylene-d12          | 23.70 | 264  | 41005    | 5.00 | ng    | -0.01    |

|                             |       |     |      |      |    |       |
|-----------------------------|-------|-----|------|------|----|-------|
| System Monitoring Compounds |       |     |      |      |    |       |
| 4) 2-Fluorophenol           | 5.35  | 112 | 674  | 0.40 | ng | 0.00  |
| 5) Phenol-d6                | 6.95  | 99  | 934  | 0.39 | ng | 0.00  |
| 8) Nitrobenzene-d5          | 8.90  | 82  | 867  | 0.42 | ng | -0.02 |
| 13) 2,4,6-Tribromophenol    | 15.87 | 330 | 85   | 0.37 | ng | -0.01 |
| 14) 2-Fluorobiphenyl        | 13.02 | 172 | 2387 | 0.39 | ng | -0.01 |
| 26) Terphenyl-d14           | 19.74 | 244 | 2880 | 0.38 | ng | -0.02 |


|                                |       |     |       |        |    |      |
|--------------------------------|-------|-----|-------|--------|----|------|
| Target Compounds               |       |     |       | Qvalue |    |      |
| 2) 1,4-Dioxane                 | 3.31  | 88  | 458   | 0.41   | ng | 90   |
| 3) n-Nitrosodimethylamine      | 3.61  | 42  | 444   | 0.39   | ng | # 84 |
| 6) bis(2-Chloroethyl)ether     | 7.20  | 93  | 1060  | 0.39   | ng | # 50 |
| 9) Naphthalene                 | 10.58 | 128 | 3221  | 0.40   | ng | 99   |
| 10) Hexachlorobutadiene        | 10.89 | 225 | 431   | 0.40   | ng | 99   |
| 11) 2-Methylnaphthalene        | 12.21 | 142 | 1956  | 0.39   | ng | # 85 |
| 15) Acenaphthylene             | 14.09 | 152 | 9374  | 0.37   | ng | 100  |
| 16) Acenaphthene               | 14.45 | 154 | 1944  | 0.37   | ng | 93   |
| 17) Fluorene                   | 15.44 | 166 | 2178  | 0.38   | ng | 98   |
| 19) Hexachlorobenzene          | 16.47 | 284 | 623   | 0.38   | ng | # 65 |
| 20) Pentachlorophenol          | 16.79 | 266 | 117   | 0.49   | ng | # 90 |
| 21) Phenanthrene               | 17.18 | 178 | 4448  | 0.39   | ng | # 94 |
| 22) Anthracene                 | 17.27 | 178 | 3337  | 0.36   | ng | 99   |
| 23) Fluoranthene               | 19.16 | 202 | 15424 | 0.37   | ng | 99   |
| 25) Pyrene                     | 19.54 | 202 | 16544 | 0.38   | ng | 99   |
| 27) Benzo(a)anthracene         | 21.26 | 228 | 3705  | 0.35   | ng | # 94 |
| 28) Chrysene                   | 21.31 | 228 | 4999  | 0.41   | ng | # 76 |
| 29) Bis(2-ethylhexyl)phthalate | 21.21 | 149 | 869   | 0.44   | ng | # 74 |
| 30) Indeno(1,2,3-cd)pyrene     | 26.24 | 276 | 14460 | 0.36   | ng | # 93 |
| 32) Benzo(b)fluoranthene       | 22.96 | 252 | 3741  | 0.37   | ng | 97   |
| 33) Benzo(k)fluoranthene       | 22.96 | 252 | 3741  | 0.41   | ng | 98   |
| 34) Benzo(a)pyrene             | 23.59 | 252 | 2725  | 0.35   | ng | # 97 |
| 35) Dibenzo(a,h)anthracene     | 26.25 | 278 | 3215  | 0.36   | ng | # 93 |
| 36) Benzo(g,h,i)perylene       | 27.00 | 276 | 13542 | 0.37   | ng | # 93 |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040417\  
Data File : BE092886.D  
Acq On : 4 Apr 2017 15:58  
Operator : SJ/MA  
Sample : SSTDICV0.4  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
ICVBE040417

Quant Time: Apr 04 16:31:42 2017  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040417.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Apr 04 16:08:03 2017  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.77 min Scan# 555

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:152 Resp: 8366

Ion Ratio Lower Upper

152 100

150 177.4 106.0 159.0#




115 83.6 31.2 46.8#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.41 ng

RT: 3.31 min Scan# 62

Delta R.T. -0.01 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion: 88 Resp: 458

Ion Ratio Lower Upper

88 100

58 71.6 63.6 95.4



43 28.4 28.0 42.0




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.39 ng

RT: 3.61 min Scan# 88

Delta R.T. 0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion: 42 Resp: 444

Ion Ratio Lower Upper

42 100

74 124.3 86.0 129.0




44 8.6 5.1 7.7#



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 0.40 ng

RT: 5.35 min Scan# 266

Delta R.T. -0.01 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion: 112 Resp: 674

Ion Ratio Lower Upper

112 100

64 69.1 53.5 80.3



63 34.9 27.9 41.9




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.39 ng

RT: 6.95 min Scan# 519

Delta R.T. 0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion: 99 Resp: 934

Ion Ratio Lower Upper

99 100

42 0.0 16.0 24.0#




71 0.0 30.6 45.8#



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.39 ng

RT: 7.20 min Scan# 530

Delta R.T. 0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion: 93 Resp: 1060

Ion Ratio Lower Upper

93 100

63 75.7 16.0 24.0#



95 31.8 24.0 36.0




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.54 min Scan# 679

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417

Tot Ion:136 Resp: 36405


Ion Ratio Lower Upper

136 100

137 12.5 8.2 12.4#

54 3.5 9.6 14.4#

68 3.8 6.7 10.1#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.42 ng

RT: 8.90 min Scan# 604

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion: 82 Resp: 867

Ion Ratio Lower Upper

82 100

128 20.4 39.0 58.4#



54 75.1 44.8 67.2#




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.40 ng

RT: 10.58 min Scan# 681

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417





Tot Ion:128 Resp: 3221

Ion Ratio Lower Upper

128 100

129 13.0 11.2 16.8

127 14.4 11.6 17.4



#10

Hexachlorobutadiene

Concen: 0.40 ng

RT: 10.89 min Scan# 697

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58




Tgt Ion:225 Resp: 431


Ion Ratio Lower Upper

225 100

223 63.1 50.6 76.0

227 65.2 51.4 77.0





#11

2-Methylnaphthalene

Concen: 0.39 ng

RT: 12.21 min Scan# 770

Delta R.T. 0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:142 Resp: 1956

Ion Ratio Lower Upper

142 100

141 78.9 73.7 110.5




115 31.4 34.0 51.0#



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.38 min Scan# 932

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion:164 Resp: 17942

Ion Ratio Lower Upper

164 100

162 117.2 71.4 107.0#



160 60.6 27.4 41.2#




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13  
 2,4,6-Tribromophenol  
 Concen: 0.37 ng  
 RT: 15.87 min Scan# 1041  
 Delta R.T. -0.01 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58




Instrument :  
 BNA\_M  
 ClientSampleId :  
 ICVBE040417


Tot Ion:330 Resp: 85  
 Ion Ratio Lower Upper  
 330 100  
 332 98.8 75.4 113.0  
 141 58.8 31.0 46.6#



#14  
 2-Fluorobiphenyl  
 Concen: 0.39 ng  
 RT: 13.02 min Scan# 840  
 Delta R.T. -0.01 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58

Tgt Ion:172 Resp: 2387  
 Ion Ratio Lower Upper  
 172 100  
 171 33.8 30.1 45.1  
 170 21.5 21.8 32.6#





#15

Acenaphthylene

Concen: 0.37 ng

RT: 14.09 min Scan# 915

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417





Tot Ion:152 Resp: 9374

Ion Ratio Lower Upper

152 100

151 4.9 3.9 5.9

153 13.1 10.4 15.6



#16

Acenaphthene

Concen: 0.37 ng

RT: 14.45 min Scan# 936

Delta R.T. 0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58




Tgt Ion:154 Resp: 1944


Ion Ratio Lower Upper

154 100

153 454.7 350.8 526.2

152 225.9 171.1 256.7









#17  
**Fluorene**  
 Concen: 0.38 ng  
 RT: 15.44 min Scan# 1004  
 Delta R.T. 0.00 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58

Instrument :  
 BNA\_M  
 ClientSampleId :  
 ICVBE040417

Tot Ion:166 Resp: 2178  




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 166 | 100   |       |       |
| 165 | 99.6  | 78.2  | 117.4 |
| 167 | 14.2  | 11.0  | 16.4  |




#18  
**Phenanthrene-d10**  
 Concen: 5.00 ng  
 RT: 17.13 min Scan# 1108  
 Delta R.T. -0.02 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58

Tgt Ion:188 Resp: 42879  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 188 | 100   |       |       |
| 94  | 14.6  | 3.2   | 4.8#  |
| 80  | 15.6  | 2.6   | 3.8#  |





#19

Hexachlorobenzene

Concen: 0.38 ng

RT: 16.47 min Scan# 1079

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:284 Resp: 623

Ion Ratio Lower Upper

284 100

142 0.0 29.1 43.7#




249 30.3 27.9 41.9



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#20

Pentachlorophenol

Concen: 0.49 ng

RT: 16.79 min Scan# 1093

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion:266 Resp: 117

Ion Ratio Lower Upper

266 100

268 77.0 62.5 93.7



264 88.5 57.2 85.8#




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E





#21

Phenanthrene

Concen: 0.39 ng

RT: 17.18 min Scan# 1110

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417





Tot Ion:178 Resp: 4448

Ion Ratio Lower Upper

178 100

176 18.7 15.8 23.8

179 16.0 16.0 24.0#



#22

Anthracene

Concen: 0.36 ng

RT: 17.27 min Scan# 1114

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58




Tgt Ion:178 Resp: 3337


Ion Ratio Lower Upper

178 100

176 18.0 15.0 22.4

179 15.8 12.2 18.2









#23  
 Fluoranthene  
 Concen: 0.37 ng  
 RT: 19.16 min Scan# 1197  
 Delta R.T. -0.02 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58




Instrument :  
 BNA\_M  
 ClientSampleId :  
 ICVBE040417


Tot Ion:202 Resp: 15424  
 Ion Ratio Lower Upper  
 202 100  
 101 3.2 2.2 3.4  
 203 17.3 13.7 20.5



#24  
 Chrysene-d12  
 Concen: 5.00 ng  
 RT: 21.28 min Scan# 1319  
 Delta R.T. -0.02 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58

Tgt Ion:240 Resp: 48715  
 Ion Ratio Lower Upper  
 240 100  
 120 5.6 2.6 4.0#  
 236 28.2 24.6 37.0









#25  
Pvrene  
Concen: 0.38 ng  
RT: 19.54 min Scan# 1219  
Delta R.T. -0.00 min  
Lab File: BE092886.D  
Acq: 4 Apr 2017 15:58




Instrument :  
BNA\_M  
ClientSampleId :  
ICVBE040417


Tot Ion:202 Resp: 16544  
Ion Ratio Lower Upper  
202 100  
200 5.0 4.2 6.4  
203 17.6 13.9 20.9



#26  
Terphenyl-d14  
Concen: 0.38 ng  
RT: 19.74 min Scan# 1231  
Delta R.T. -0.02 min  
Lab File: BE092886.D  
Acq: 4 Apr 2017 15:58

Tgt Ion:244 Resp: 2880  
Ion Ratio Lower Upper  
244 100  
212 11.0 6.7 10.1#  
122 12.2 3.6 5.4#





#27

Benzo(a)anthracene

Concen: 0.35 ng

RT: 21.26 min Scan# 1318

Delta R.T. -0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:228 Resp: 3705

Ion Ratio Lower Upper

228 100

226 27.1 23.6 35.4




229 20.6 13.3 19.9#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#28

Chrysene

Concen: 0.41 ng

RT: 21.31 min Scan# 1321

Delta R.T. 0.00 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion:228 Resp: 4999

Ion Ratio Lower Upper

228 100

226 28.3 36.3 54.5#



229 20.8 10.6 16.0#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.44 ng

RT: 21.21 min Scan# 1315

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:149 Resp: 869

Ion Ratio Lower Upper

149 100

167 26.4 16.0 24.0#




279 2.3 16.0 24.0#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.36 ng

RT: 26.24 min Scan# 1702

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion:276 Resp: 14460

Ion Ratio Lower Upper

276 100

138 31.8 20.3 30.5#



277 25.0 19.7 29.5




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#31  
 Pervlene-d12  
 Concen: 5.00 ng  
 RT: 23.70 min Scan# 1491  
 Delta R.T. -0.01 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58




Instrument :  
 BNA\_M  
 ClientSampleId :  
 ICVBE040417


Tot Ion:264 Resp: 41005  
 Ion Ratio Lower Upper  
 264 100  
 260 25.2 20.1 30.1  
 265 21.6 17.9 26.9



#32  
 Benzo(b)fluoranthene  
 Concen: 0.37 ng  
 RT: 22.96 min Scan# 1427  
 Delta R.T. -0.01 min  
 Lab File: BE092886.D  
 Acq: 4 Apr 2017 15:58

Tgt Ion:252 Resp: 3741  
 Ion Ratio Lower Upper  
 252 100  
 253 25.0 18.7 28.1  
 125 14.1 10.5 15.7





#33

Benzo(k)fluoranthene

Concen: 0.41 ng

RT: 22.96 min Scan# 1427

Delta R.T. -0.06 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:252 Resp: 3741

Ion Ratio Lower Upper

252 100

253 25.0 20.3 30.5




125 14.1 9.6 14.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#34

Benzo(a)pyrene

Concen: 0.35 ng

RT: 23.59 min Scan# 1481

Delta R.T. -0.01 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion:252 Resp: 2725

Ion Ratio Lower Upper

252 100

253 24.2 19.0 28.6



125 18.0 11.8 17.8#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#35

Dibenzo(a,h)anthracene

Concen: 0.36 ng

RT: 26.25 min Scan# 1703

Delta R.T. -0.03 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58

Instrument :

BNA\_M

ClientSampleId :

ICVBE040417


Tot Ion:278 Resp: 3215

Ion Ratio Lower Upper

278 100

139 25.0 13.8 20.8#




279 26.7 21.4 32.2



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#36

Benzo(g,h,i)perylene

Concen: 0.37 ng

RT: 27.00 min Scan# 1747

Delta R.T. -0.02 min

Lab File: BE092886.D

Acq: 4 Apr 2017 15:58


Tgt Ion:276 Resp: 13542

Ion Ratio Lower Upper

276 100

277 26.6 19.8 29.8

138 29.8 19.8 29.6#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

