

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040422\
 Data File : BM034493.D
 Acq On : 04 Apr 2022 18:46
 Operator : CG/JU
 Sample : SSTD3.220
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2020

Quant Time: Apr 05 11:28:01 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 05 11:24:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.920	152	2900	0.400	ng/ul	0.00
4) Naphthalene-d8	10.724	136	8018	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.555	164	3973	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.299	188	6815	0.400	ng/ul	-0.01
17) Chrysene-d12	21.480	240	5683	0.400	ng/ul	#-0.01
23) Perylene-d12	23.889	264	5085	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.292	96	12537	2.581	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.330	88	12792	1.941	ng/ul#	79
14) Pentachlorophenol	16.957	266	8215	2.843	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040422\
Data File : BM034493.D
Acq On : 04 Apr 2022 18:46
Operator : CG/JU
Sample : SSTD3.220
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2020

Quant Time: Apr 05 11:28:01 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 05 11:24:23 2022
Response via : Initial Calibration

