

Data Path : Z:\HPCHEM1\BNA M\Data\BM040516\
 Data File : BM004883.D
 Acq On : 05 Apr 2016 16:56
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02063

Quant Time: Apr 05 17:28:01 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM040116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 05 06:44:44 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.602	0.562	6.6	104	0.00
3 S	1,4-Dioxane-d8	0.494	0.445	9.9	98	0.00
4	Benzaldehyde	0.829	0.924	-11.5	103	0.00
5 S	Phenol-d5	1.777	1.641	7.7	106	0.00
6	Phenol	1.856	1.764	5.0	108	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.963	0.940	2.4	108	0.00
8	Bis(2-Chloroethyl)ether	1.406	1.348	4.1	108	0.00
9 S	2-Chlorophenol-d4	1.349	1.287	4.6	107	0.00
10	2-Chlorophenol	1.420	1.352	4.8	106	0.00
11	2-Methylphenol	1.465	1.353	7.6	106	0.00
12	2,2'-oxybis(1-Chloropropane	1.715	1.700	0.9	111	0.00
13 S	4-Methylphenol-d8	1.493	1.382	7.4	106	0.00
14	Acetophenone	2.257	2.209	2.1	107	0.00
15 P	N-Nitroso-di-n-propylamine	1.142	1.087	4.8	105	0.00
16	4-Methylphenol	1.632	1.548	5.1	108	0.00
17	Hexachloroethane	0.545	0.499	8.4	103	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
19 S	Nitrobenzene-d5	0.142	0.132	7.0	103	0.00
20	Nitrobenzene	0.340	0.331	2.6	109	0.00
21	Isophorone	0.678	0.635	6.3	104	0.00
22 S	2-Nitrophenol-d4	0.158	0.144	8.9	102	0.00
23 C	2-Nitrophenol	0.172	0.163	5.2	105	0.00
24	2,4-Dimethylphenol	0.364	0.352	3.3	107	0.00
25	Bis(2-Chloroethoxy)methane	0.422	0.400	5.2	107	0.00
26 S	2,4-Dichlorophenol-d3	0.294	0.286	2.7	107	0.00
27 C	2,4-Dichlorophenol	0.304	0.299	1.6	109	0.00
28	Naphthalene	1.011	0.986	2.5	110	0.00
29 S	4-Chloroaniline-d4	0.351	0.395	-12.5	107	0.00
30	4-Chloroaniline	0.370	0.423	-14.3	110	0.00
31 C	Hexachlorobutadiene	0.185	0.176	4.9	108	0.00
32	Caprolactam	0.115	0.097	15.7	99	0.00
33 C	4-Chloro-3-methylphenol	0.348	0.343	1.4	108	0.00
34	2-Methylnaphthalene	0.753	0.743	1.3	110	0.00
35	1-Methylnaphthalene	0.730	0.716	1.9	110	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
37	1,2,4,5-Tetrachlorobenzene	0.619	0.583	5.8	110	0.00
38	Hexachlorocyclopentadiene	0.368	0.303	17.7	100	0.00
39 C	2,4,6-Trichlorophenol	0.409	0.388	5.1	110	0.00
40	2,4,5-Trichlorophenol	0.449	0.429	4.5	111	0.00
41	1,1'-Biphenyl	1.624	1.553	4.4	111	0.00
42	2-Chloronaphthalene	1.223	1.175	3.9	112	0.00
43	2-Nitroaniline	0.352	0.330	6.2	109	0.00
44 S	Dimethylphthalate-d6	1.578	1.504	4.7	111	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM040516\
 Data File : BM004883.D
 Acq On : 05 Apr 2016 16:56
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02063

Quant Time: Apr 05 17:28:01 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM040116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 05 06:44:44 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.607	1.547	3.7	113	0.00
46	2,6-Dinitrotoluene	0.313	0.300	4.2	112	0.00
47 S	Acenaphthylene-d8	1.888	1.820	3.6	112	0.00
48	Acenaphthylene	2.009	1.943	3.3	112	0.00
49	3-Nitroaniline	0.320	0.321	-0.3	112	0.00
50 C	Acenaphthene	1.346	1.288	4.3	111	0.00
51	2,4-Dinitrophenol	0.208	0.153	26.4#	98	0.00
52 S	4-Nitrophenol-d4	0.312	0.287	8.0	110	0.00
53	4-Nitrophenol	0.256	0.241	5.9	112	0.00
54	Dibenzofuran	1.953	1.897	2.9	113	0.00
55	2,4-Dinitrotoluene	0.471	0.457	3.0	112	0.00
56	2,3,4,6-Tetrachlorophenol	0.399	0.384	3.8	110	0.00
57	Diethylphthalate	1.627	1.561	4.1	111	0.00
58 S	Fluorene-d10	1.364	1.329	2.6	114	0.00
59	Fluorene	1.556	1.515	2.6	111	0.00
60	4-Chlorophenyl-phenylether	0.757	0.738	2.5	111	0.00
61	4-Nitroaniline	0.375	0.350	6.7	113	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.120	0.104	13.3	109	0.00
64	4,6-Dinitro-2-methylphenol	0.128	0.115	10.2	110	0.00
65	N-Nitrosodiphenylamine	0.580	0.548	5.5	113	0.00
66	4-Bromophenyl-phenylether	0.199	0.188	5.5	113	0.00
67	Hexachlorobenzene	0.227	0.218	4.0	115	0.00
68	Atrazine	0.213	0.201	5.6	110	0.00
69 C	Pentachlorophenol	0.141	0.124	12.1	110	0.00
70	Phenanthrene	1.103	1.063	3.6	115	0.00
71 S	Anthracene-d10	0.895	0.864	3.5	115	0.00
72	Anthracene	1.106	1.075	2.8	114	0.00
73	1,2,3,4-Tetrachlorobenzene	0.260	0.240	7.7	112	0.00
74	Pentachlorobenzene	0.260	0.244	6.2	114	0.00
75	Carbazole	0.987	0.978	0.9	113	0.00
76	Di-n-butylphthalate	1.172	1.076	8.2	107	0.00
77 C	Fluoranthene	1.225	1.249	-2.0	113	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
79 S	Pyrene-d10	0.868	0.829	4.5	115	0.00
80	Pyrene	1.145	1.094	4.5	114	0.00
81	Butylbenzylphthalate	0.485	0.430	11.3	103	0.00
82	3,3'-Dichlorobenzidine	0.390	0.371	4.9	102	0.00
83	Benzo(a)anthracene	1.147	1.095	4.5	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.694	0.613	11.7	102	0.00
85	Chrysene	1.088	1.042	4.2	114	0.00
86 I	Perylene-d12	1.000	1.000	0.0	122	0.00
87	Di-n-octyl phthalate	1.184	1.048	11.5	102	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM040516\
Data File : BM004883.D
Acq On : 05 Apr 2016 16:56
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02063

Quant Time: Apr 05 17:28:01 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM040116.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 05 06:44:44 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.146	1.085	5.3	116	0.00
89	Benzo(k)fluoranthene	1.099	1.069	2.7	115	0.00
90 S	Benzo(a)pyrene-d12	0.886	0.840	5.2	114	0.00
91 C	Benzo(a)pyrene	1.112	1.067	4.0	115	0.00
92	Indeno(1,2,3-cd)pyrene	1.326	1.246	6.0	112	0.00
93	Dibenzo(a,h)anthracene	1.102	1.055	4.3	114	0.00
94	Benzo(a,h,i)perylene	1.139	1.049	7.9	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0