

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM040519\
 Data File : BM019741.D
 Acq On : 05 Apr 2019 09:41
 Operator : JU/SJ
 Sample : PB118517BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK17

Quant Time: Apr 06 00:21:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 06 00:13:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	133562	20.00	ng/ul	0.00
18) Naphthalene-d8	10.34	136	630620	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.22	164	385470	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.99	188	862396	20.00	ng/ul	0.00
78) Chrysene-d12	21.19	240	803599	20.00	ng/ul	0.00
86) Perylene-d12	23.40	264	730230	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.12	96	14048	4.12	ng/uL	0.00
5) Phenol-d5	6.75	99	274585	23.09	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.92	67	172206	21.63	ng/ul	0.00
9) 2-Chlorophenol-d4	7.10	132	219245	24.32	ng/ul	0.00
13) 4-Methylphenol-d8	8.28	113	227481	25.48	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	105691	23.49	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	121968	24.36	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.97	165	215715	22.83	ng/ul	0.00
29) 4-Chloroaniline-d4	10.51	131	310077	27.33	ng/ul	0.00
44) Dimethylphthalate-d6	13.65	166	751072	24.15	ng/ul	0.00
47) Acenaphthylene-d8	13.92	160	940955	24.19	ng/ul	0.00
52) 4-Nitrophenol-d4	14.49	143	81888	16.80	ng/ul	0.01
58) Fluorene-d10	15.23	176	661721	24.69	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.38	200	85624	15.01	ng/ul	0.00
71) Anthracene-d10	17.09	188	1004856	24.28	ng/ul	0.00
79) Pyrene-d10	19.39	212	1070576	28.84	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.26	264	899302	23.23	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM040519\
Data File : BM019741.D
Acq On : 05 Apr 2019 09:41
Operator : JU/SJ
Sample : PB118517BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK17

Quant Time: Apr 06 00:21:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 06 00:13:11 2019
Response via : Initial Calibration

