


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM040519\
Data File : BM019739.D
Acq On : 05 Apr 2019 08:30
Operator : JU/SJ
Sample : DFTPP62
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP62

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM032219MA.M
Title : SVOA CALIBRATION
Last Update : Mon Apr 08 04:16:20 2019



AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2446

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.3	71829	PASS
68	69	0.00	2	1.7	1987	PASS
69	198	0.00	100	48.8	115764	PASS
70	69	0.00	2	0.7	807	PASS
127	198	10	80	49.7	117890	PASS
197	198	0.00	2	0.7	1755	PASS
198	198	100	100	100.0	237120	PASS
199	198	5	9	6.8	16014	PASS
275	198	10	60	31.3	74285	PASS
365	198	1	100	4.9	11548	PASS
441	198	0.01	100	14.9	35338	PASS
442	198	50	100	96.8	229525	PASS
443	442	15	24	19.4	44584	PASS

m/z	Abundance
37.00	604.0
38.00	1130.0
39.10	7094.0
40.00	555.0
41.10	813.0
43.00	630.0
43.90	302.0
48.80	623.0
50.00	23360.0
51.00	89840.0
52.00	5183.0
55.10	954.0
56.00	3060.0
57.00	7561.0
61.10	1585.0
62.00	1901.0
63.00	4749.0
64.10	666.0
65.10	2392.0
67.00	494.0
68.00	2428.0
69.00	144384.0
70.00	957.0
71.10	558.0
73.10	2289.0
74.00	14084.0
75.00	22040.0
76.00	8164.0
77.10	151616.0
78.10	10657.0
79.00	10508.0
80.00	7335.0
81.00	11632.0
82.00	2713.0
83.00	2663.0
84.10	487.0
85.00	2395.0
86.00	3433.0
87.00	1273.0
88.00	473.0
91.00	2742.0
91.90	2595.0
93.00	17960.0
94.00	1192.0
95.00	375.0
96.00	886.0
97.00	437.0
98.00	13398.0
99.00	10582.0
100.00	792.0
101.00	5859.0
102.00	306.0
103.00	1955.0
104.00	3801.0
105.00	3638.0
106.00	1160.0
107.00	42360.0
108.00	6175.0
109.10	1442.0
110.00	74624.0
111.00	11065.0
112.00	1638.0
113.00	689.0
114.90	483.0
116.00	2670.0
117.00	36656.0
118.00	2645.0
119.00	656.0
119.90	606.0
121.00	472.0
121.90	2512.0
123.00	4037.0
124.00	2084.0
125.00	1889.0
126.10	967.0
127.00	139840.0
128.00	11005.0
129.00	56056.0
130.00	4813.0
131.00	1025.0
131.90	598.0
133.00	554.0
134.00	1945.0
135.00	4860.0
136.00	1637.0
137.00	2298.0
137.90	512.0
138.90	414.0
140.00	938.0
141.00	7006.0
142.00	2264.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

143.00	1872.0
144.00	548.0
145.00	511.0
145.90	1514.0
147.00	4187.0
148.00	9458.0
148.90	1980.0
150.00	466.0
150.90	939.0
151.80	867.0
152.90	2501.0
154.00	2190.0
155.00	4457.0
156.00	6059.0
157.10	1351.0
158.00	1418.0
158.90	1221.0
160.00	2673.0
161.00	3572.0
161.90	1266.0
163.00	365.0
163.90	460.0
165.00	3350.0
166.00	2157.0
167.00	16133.0
168.00	8020.0
169.00	1471.0
169.90	632.0
170.80	940.0
171.90	1501.0
173.00	2085.0
174.00	3336.0
175.00	6084.0
175.90	1702.0
176.90	2893.0
177.90	1198.0
179.00	11821.0
180.00	8028.0
181.00	3812.0
181.90	677.0
182.90	462.0
184.00	1217.0
185.00	5298.0
186.00	40424.0
187.00	12674.0
187.90	1535.0
189.00	3300.0
190.00	404.0
190.90	1646.0
192.00	3356.0
193.00	4707.0
193.90	1125.0
195.00	489.0
196.00	7763.0
197.10	1775.0
197.90	268736.0
198.90	18368.0
200.00	1999.0
201.30	1487.0
203.00	2450.0
204.00	10450.0
205.00	18144.0
206.00	72992.0
207.00	11085.0
208.00	2498.0
209.00	1012.0
209.90	1044.0
211.00	2527.0
211.80	658.0
212.90	333.0
214.90	873.0
215.90	1386.0
217.00	20368.0
217.90	2514.0
218.80	325.0
220.30	669.0
221.00	12716.0
221.80	2823.0
223.00	4330.0
224.00	40208.0
225.00	10792.0
226.10	1258.0
227.00	17720.0
228.00	3013.0
229.00	3508.0
229.90	612.0
231.00	1653.0
233.10	314.0
233.90	911.0
235.00	1296.0
235.90	668.0
237.00	1661.0
237.90	496.0
239.00	727.0
239.90	531.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

240.90	1004.0
242.00	1948.0
243.00	2148.0
244.00	31576.0
245.00	4677.0
246.00	6231.0
246.90	1086.0
249.00	1322.0
250.90	422.0
252.00	378.0
253.00	811.0
254.20	1439.0
255.00	173376.0
256.00	26304.0
257.00	2448.0
258.00	10764.0
259.00	1708.0
260.00	312.0
265.00	4242.0
266.00	758.0
271.00	414.0
272.00	505.0
273.00	5480.0
274.00	13144.0
275.00	76256.0
276.00	10097.0
277.00	6274.0
277.90	1372.0
281.10	619.0
283.00	835.0
284.00	578.0
285.00	1314.0
286.00	365.0
288.90	346.0
292.90	1741.0
294.00	424.0
295.00	550.0
296.00	23624.0
297.00	3026.0
300.90	364.0
302.00	474.0
303.00	2795.0
304.00	697.0
308.00	302.0
310.10	338.0
313.10	335.0
313.90	1219.0
315.00	2850.0
316.00	1293.0
320.90	609.0
322.10	351.0
323.00	7112.0
324.00	1468.0
327.00	1585.0
327.90	698.0
331.90	735.0
333.10	689.0
334.00	4344.0
335.00	1187.0
341.00	1119.0
346.00	1808.0
347.00	351.0
352.10	2451.0
353.00	1728.0
354.00	2390.0
355.00	707.0
365.00	10262.0
365.90	1733.0
371.00	736.0
372.10	3792.0
373.00	1128.0
383.00	979.0
390.10	502.0
390.90	319.0
392.00	335.0
401.00	348.0
402.00	1482.0
403.00	2284.0
404.10	618.0
421.00	1966.0
422.00	1868.0
423.10	12430.0
424.10	3428.0
425.00	453.0
441.10	30536.0
442.10	201152.0
443.00	39648.0
444.00	3787.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
36.90	444.0
38.00	1262.0
39.10	7465.0
40.00	665.0
41.00	879.0
43.00	787.0
49.00	864.0
50.00	21344.0
51.00	85496.0
52.00	4390.0
53.00	347.0
55.00	964.0
56.00	3268.0
57.00	7022.0
58.00	523.0
61.00	1485.0
62.00	1903.0
63.00	5080.0
64.00	703.0
65.00	2532.0
65.90	304.0
67.00	311.0
68.00	2350.0
69.00	137152.0
70.00	774.0
71.10	627.0
73.00	2502.0
74.00	14596.0
75.00	21576.0
76.00	7294.0
77.10	146432.0
78.10	9862.0
79.00	10557.0
80.00	7752.0
81.00	11239.0
82.00	3227.0
83.00	2345.0
84.00	509.0
85.00	1969.0
86.00	3208.0
87.00	1614.0
87.90	570.0
91.00	2944.0
92.00	2487.0
93.00	16592.0
94.00	1538.0
95.10	567.0
96.00	1051.0
97.10	543.0
98.00	13357.0
99.00	10061.0
100.00	1099.0
101.00	5788.0
101.90	454.0
103.00	1682.0
104.00	3675.0
105.00	3577.0
106.00	1042.0
107.00	42912.0
108.00	7189.0
108.90	1183.0
110.00	73816.0
111.00	11104.0
112.00	1226.0
113.10	630.0
115.00	460.0
116.00	2828.0
117.00	36784.0
118.00	3075.0
119.00	653.0
120.00	462.0
121.00	442.0
122.00	2547.0
122.90	4487.0
124.00	1781.0
125.00	1851.0
127.00	139136.0
128.00	10283.0
129.00	53944.0
130.00	4943.0
131.00	879.0
132.10	705.0
133.00	465.0
134.00	1572.0
135.00	4122.0
136.00	1975.0
137.00	2655.0
137.90	559.0
138.90	559.0
140.00	952.0
141.00	8290.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

142.00	2589.0
143.10	1807.0
144.00	518.0
145.10	489.0
146.00	1475.0
147.00	4322.0
148.00	8961.0
149.00	1860.0
149.90	447.0
151.00	1016.0
151.80	668.0
153.00	2898.0
153.90	1944.0
155.00	4149.0
156.00	6281.0
157.00	1246.0
158.00	1288.0
159.00	1259.0
159.90	2484.0
161.00	3391.0
161.90	1096.0
162.90	486.0
163.90	417.0
164.90	3296.0
166.00	2151.0
167.00	16760.0
168.00	8734.0
169.00	1497.0
170.10	536.0
171.00	863.0
171.90	1410.0
173.00	1730.0
174.00	3436.0
175.00	5785.0
175.90	1793.0
176.90	2749.0
178.00	947.0
179.00	12368.0
180.00	8199.0
181.00	3170.0
182.00	592.0
183.00	403.0
184.00	932.0
185.00	5463.0
186.00	44976.0
187.00	12721.0
188.00	1426.0
189.00	2718.0
189.90	550.0
191.00	1842.0
192.00	3749.0
193.00	4882.0
193.90	1028.0
195.10	564.0
196.00	9188.0
197.10	2161.0
197.90	284096.0
198.90	18792.0
199.90	1855.0
201.50	1600.0
203.00	2413.0
204.00	11601.0
205.00	19496.0
206.00	80272.0
207.00	12666.0
208.00	3215.0
209.00	1208.0
210.20	1570.0
210.90	3217.0
215.00	1019.0
216.00	1485.0
216.90	21288.0
218.00	2538.0
221.00	12519.0
221.80	2599.0
223.00	4474.0
224.00	42552.0
225.00	11602.0
226.00	1420.0
227.00	19376.0
228.00	2703.0
229.00	3813.0
230.00	677.0
231.00	1456.0
232.10	354.0
233.00	603.0
233.90	1183.0
234.90	1568.0
235.90	929.0
237.00	1854.0
239.00	796.0
240.00	590.0
240.90	1217.0
242.00	2337.0
243.10	2590.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

244.00	35344.0
245.00	4471.0
246.00	7340.0
246.90	1445.0
247.80	358.0
249.00	1496.0
250.90	531.0
251.90	309.0
252.90	674.0
254.20	1847.0
255.00	196672.0
256.00	27488.0
257.00	2316.0
258.00	11560.0
258.90	1885.0
260.90	325.0
264.00	454.0
265.00	5043.0
266.00	537.0
269.90	347.0
271.00	401.0
271.90	567.0
273.00	5712.0
274.00	15371.0
275.00	90264.0
276.00	12055.0
277.00	7760.0
277.90	1082.0
281.00	600.0
283.00	1083.0
283.90	689.0
285.00	1379.0
288.80	336.0
292.90	1493.0
294.00	557.0
294.90	510.0
296.00	27984.0
297.00	4023.0
301.00	393.0
302.00	448.0
303.00	3198.0
304.00	861.0
307.90	315.0
309.90	394.0
314.00	1333.0
315.00	3282.0
316.10	1505.0
317.00	330.0
321.00	1108.0
322.00	339.0
323.00	8576.0
324.00	1785.0
326.00	321.0
326.90	1908.0
328.00	1011.0
332.00	815.0
332.90	864.0
334.00	5852.0
335.00	1511.0
341.10	1075.0
341.90	368.0
346.00	2110.0
347.00	386.0
352.00	2724.0
353.00	1721.0
354.00	3309.0
355.00	643.0
365.00	13801.0
365.90	1940.0
371.00	1084.0
372.10	4498.0
373.00	1301.0
383.00	1706.0
384.00	539.0
390.00	509.0
390.90	470.0
401.00	319.0
402.00	2188.0
403.00	2800.0
404.00	986.0
421.00	2588.0
422.10	2299.0
423.00	17040.0
424.00	4027.0
425.00	338.0
441.10	42696.0
442.10	273280.0
443.10	53096.0
444.10	4736.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
38.00	653.0
39.10	3349.0
40.00	591.0
41.00	565.0
42.90	521.0
44.00	434.0
48.90	481.0
50.00	10255.0
51.00	41328.0
52.00	2126.0
55.00	910.0
56.00	1605.0
57.00	3899.0
57.90	344.0
61.00	843.0
62.00	541.0
63.00	2572.0
65.00	1270.0
67.00	306.0
68.00	1185.0
69.00	66760.0
70.00	691.0
70.90	630.0
73.10	1839.0
74.00	7589.0
75.00	10004.0
76.00	3540.0
77.00	71448.0
78.00	5054.0
79.00	5190.0
80.00	3235.0
81.00	5670.0
82.00	1343.0
83.10	1448.0
83.90	426.0
85.00	1463.0
86.10	1600.0
86.90	758.0
91.00	1671.0
92.00	1258.0
93.00	9013.0
93.90	594.0
95.00	485.0
96.00	623.0
97.10	462.0
98.00	6799.0
99.00	5140.0
99.90	403.0
101.00	2151.0
103.00	953.0
104.00	1735.0
105.00	2095.0
105.90	688.0
107.00	20920.0
108.00	3381.0
109.00	689.0
110.00	36072.0
111.00	5639.0
112.00	609.0
115.10	437.0
116.00	1144.0
117.00	18808.0
118.00	1655.0
118.90	402.0
119.90	370.0
122.00	1436.0
123.00	2198.0
123.90	926.0
124.90	1087.0
127.00	74696.0
128.00	5744.0
129.00	29104.0
130.00	2534.0
130.90	688.0
132.00	310.0
132.90	481.0
133.90	1163.0
135.00	3011.0
136.00	1118.0
137.00	1249.0
139.80	429.0
141.00	4531.0
142.00	1242.0
142.90	917.0
144.90	378.0
146.00	698.0
147.00	2625.0
147.90	5540.0
149.00	1159.0
149.90	324.0
150.90	511.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

151.60	413.0
153.00	1397.0
153.90	916.0
155.00	2408.0
156.00	3867.0
157.00	667.0
157.90	831.0
158.90	708.0
159.90	1397.0
161.00	1960.0
161.90	743.0
162.90	387.0
164.00	335.0
164.90	2086.0
166.00	1288.0
167.00	9548.0
168.00	4980.0
169.00	888.0
170.00	372.0
170.90	588.0
171.90	965.0
172.90	1191.0
174.00	1963.0
175.00	3208.0
176.00	1241.0
176.90	1716.0
177.90	667.0
179.00	7408.0
180.00	5095.0
180.90	1976.0
181.90	399.0
184.00	552.0
185.00	2795.0
186.00	24440.0
187.00	6293.0
187.90	725.0
189.00	2091.0
190.90	1052.0
192.00	2480.0
193.00	2966.0
194.00	490.0
196.00	4758.0
197.10	1329.0
197.90	158528.0
198.90	10883.0
199.90	1170.0
201.40	914.0
203.00	1279.0
204.00	6279.0
205.00	11056.0
206.00	44280.0
207.00	7365.0
208.00	1912.0
208.90	705.0
210.20	721.0
210.90	1868.0
212.10	308.0
215.00	540.0
216.00	1075.0
217.00	12731.0
218.00	1409.0
221.10	6975.0
221.90	1595.0
223.00	2987.0
224.00	25776.0
225.00	7097.0
226.00	948.0
227.00	11102.0
227.90	1491.0
228.90	2248.0
229.90	369.0
231.00	1011.0
233.90	753.0
235.00	847.0
236.10	402.0
237.00	1066.0
239.00	530.0
239.90	398.0
241.00	710.0
242.00	1504.0
243.00	1611.0
244.00	21280.0
245.00	2785.0
246.00	4436.0
247.00	883.0
248.90	712.0
252.90	433.0
254.10	1395.0
255.00	116960.0
256.00	16624.0
257.00	1481.0
258.00	7793.0
259.00	978.0
263.90	343.0
264.90	3204.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

266.00	534.0
272.00	368.0
273.00	3540.0
274.00	9275.0
275.00	56336.0
276.00	7145.0
276.90	4523.0
278.00	1044.0
281.00	573.0
283.00	922.0
284.00	400.0
285.00	1165.0
293.00	1159.0
294.90	336.0
296.00	17568.0
297.00	2350.0
302.00	305.0
302.90	2177.0
304.00	504.0
314.00	865.0
315.00	2175.0
316.00	1284.0
321.10	703.0
322.00	324.0
323.00	6405.0
324.00	1055.0
326.90	1251.0
327.90	519.0
333.00	631.0
334.00	4085.0
335.10	994.0
341.00	1001.0
346.00	1629.0
352.10	2117.0
353.00	1041.0
354.00	2014.0
355.10	542.0
365.00	10581.0
366.00	1504.0
371.00	541.0
372.00	3667.0
373.00	869.0
383.00	1092.0
383.90	405.0
390.00	580.0
391.00	335.0
402.00	1350.0
403.00	2441.0
404.00	675.0
421.00	1893.0
421.90	1749.0
423.00	12904.0
424.00	2982.0
441.10	32784.0
442.10	214144.0
443.10	41008.0
444.10	3560.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
39.90	415.0
40.90	382.0
43.00	349.0
51.00	392.0
55.10	388.0
67.10	311.0
69.10	334.0
73.00	1321.0
76.90	419.0
81.00	320.0
83.90	391.0
90.90	438.0
95.00	301.0
119.00	385.0
133.10	369.0
147.00	485.0
193.00	407.0
207.00	1449.0
208.10	381.0
281.00	535.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62