

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM040521\
 Data File : BM029343.D
 Acq On : 05 Apr 2021 15:12
 Operator : CG/JU
 Sample : SSTD3.281
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.281

Quant Time: Apr 05 15:49:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM040521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 05 14:08:43 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	2105	0.40	ng/ul	0.00
2) Naphthalene-d8	10.48	136	7048	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.33	164	3680	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	7428	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	8393	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	8495	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.72	266	5679	2.926	ng/ul	Ovalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM040521\
Data File : BM029343.D
Acq On : 05 Apr 2021 15:12
Operator : CG/JU
Sample : SST3.281
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.281

Quant Time: Apr 05 15:49:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM040521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 05 14:08:43 2021
Response via : Initial Calibration

