




Data Path : Z:\HPCHEM1\BNA M\DATA\BM040618\
Data File : BM014682.D
Acq On : 06 Apr 2018 14:11
Operator : SJ/JU
Sample : J2246-06DL 5X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Apr 07 00:54:27 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM032818MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 07 00:52:04 2018
Response via : Initial Calibration



TIC: BM014682.D

(36) 1,2,4,5-Tetrachlorobenzene

13.392min (-0.006) 9.72ng/ul

response 287295

Ion	Exp%	Act%
216.00	100	100
214.00	78.10	80.91
179.00	22.70	22.45
108.00	17.70	17.47