

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040622\
 Data File : BM034520.D
 Acq On : 06 Apr 2022 17:25
 Operator : CG/JU
 Sample : PB143886BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK886

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/07/2022
 Supervised By :Yogesh Patel 04/09/2022

Quant Time: Apr 07 01:02:43 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 05 11:50:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	2488	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.719	136	6776	0.400	ng/u1	#-0.02
9) Acenaphthene-d10	14.555	164	3648	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.295	188	6431	0.400	ng/u1	-0.02
17) Chrysene-d12	21.518	240	4301	0.400	ng/u1	# 0.02
23) Perylene-d12	23.866	264	2077m	0.400	ng/u1	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.292	96	21303	5.112	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.346	152	2947	0.331	ng/u1	0.01
18) Fluoranthene-d10	19.324	212	5553	0.440	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040622\
Data File : BM034520.D
Acq On : 06 Apr 2022 17:25
Operator : CG/JU
Sample : PB143886BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK886

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/07/2022
Supervised By :Yogesh Patel 04/09/2022

Quant Time: Apr 07 01:02:43 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 05 11:50:53 2022
Response via : Initial Calibration

