

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040623\
Data File : BM039323.D
Acq On : 06 Apr 2023 19:13
Operator : CG/JU
Sample : PB151749BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK749

Quant Time: Apr 07 02:17:14 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 23:58:10 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	6072	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.697	136	23156	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.536	164	14161	0.400	ng/u1	# 0.00
13) Phenanthrene-d10	17.282	188	26470	0.400	ng/u1	0.00
17) Chrysene-d12	21.480	240	16984	0.400	ng/u1	# 0.00
23) Perylene-d12	23.860	264	26036	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	44745	3.980	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.302	152	8005	0.299	ng/u1	0.01
18) Fluoranthene-d10	19.315	212	18643	0.352	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040623\
Data File : BM039323.D
Acq On : 06 Apr 2023 19:13
Operator : CG/JU
Sample : PB151749BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK749

Quant Time: Apr 07 02:17:14 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 23:58:10 2023
Response via : Initial Calibration

