

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040623\
 Data File : BM039324.D
 Acq On : 06 Apr 2023 19:52
 Operator : CG/JU
 Sample : PB151753BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK753

Quant Time: Apr 07 02:17:21 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 06 23:58:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	5987	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.691	136	23461	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.532	164	14488	0.400	ng/u1	# 0.00
13) Phenanthrene-d10	17.282	188	26789	0.400	ng/u1	0.00
17) Chrysene-d12	21.480	240	16891	0.400	ng/u1	# 0.00
23) Perylene-d12	23.860	264	25090	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.259	96	46917	4.233	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.302	152	8966	0.331	ng/u1	0.01
18) Fluoranthene-d10	19.315	212	20865	0.396	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040623\
Data File : BM039324.D
Acq On : 06 Apr 2023 19:52
Operator : CG/JU
Sample : PB151753BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK753

Quant Time: Apr 07 02:17:21 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 06 23:58:10 2023
Response via : Initial Calibration

