

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040715\  
 Data File : BM000897.D  
 Acq On : 08 Apr 2015 02:33  
 Operator : TP/IZ  
 Sample : PB82337BL  
 Misc : LOD/LOO-WATER- BLK  
 ALS Vial : 57 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 PB82337BL

Quant Time: Apr 08 05:49:17 2015  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040715.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Apr 07 15:59:39 2015  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.59  | 152  | 21684    | 5.00  | ng    | -0.04    |
| 11) Naphthalene-d8          | 10.39 | 136  | 111602   | 5.00  | ng    | 0.00     |
| 19) Acenaphthene-d10        | 14.26 | 164  | 60750    | 5.00  | ng    | 0.00     |
| 26) Phenanthrene-d10        | 17.01 | 188  | 116555   | 5.00  | ng    | 0.00     |
| 30) Chrysene-d12            | 21.20 | 240  | 103797   | 5.00  | ng    | 0.00     |
| 34) Perylene-d12            | 23.38 | 264  | 95833    | 5.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 4) 2-Fluorophenol           | 5.19  | 112  | 99180    | 13.63 | ng    | -0.02    |
| 5) Phenol-d6                | 6.77  | 99   | 170599   | 11.42 | ng    | 0.00     |
| 12) Nitrobenzene-d5         | 8.73  | 82   | 48488    | 8.71  | ng    | -0.04    |
| 20) 2,4,6-Tribromophenol    | 15.75 | 330  | 28626    | 11.20 | ng    | 0.00     |
| 23) 2-Fluorobiphenyl        | 12.87 | 172  | 198265   | 10.59 | ng    | 0.00     |
| 32) Terphenyl-d14           | 19.65 | 244  | 145893   | 11.39 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040715\  
Data File : BM000897.D  
Acq On : 08 Apr 2015 02:33  
Operator : TP/IZ  
Sample : PB82337BL  
Misc : LOD/LOO-WATER- BLK  
ALS Vial : 57 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
PB82337BL

Quant Time: Apr 08 05:49:17 2015  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040715.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Apr 07 15:59:39 2015  
Response via : Initial Calibration

