

Data Path : Z:\HPCHEM1\BNA M\Data\BM040715\
 Data File : BM000889.D
 Acq On : 07 Apr 2015 21:47
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 08 06:59:12 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 07 15:59:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	1.024	-2.4	99	0.00
3	n-Nitrosodimethylamine	1.000	1.015	-1.5	115	-0.02
4 S	2-Fluorophenol	1.000	1.035	-3.5	109	0.00
5 S	Phenol-d6	1.000	0.992	0.8	114	0.00
6	2-Chlorophenol	1.000	0.997	0.3	116	0.00
7 C	Phenol	1.000	0.932	6.8	101	0.00
8	bis(2-Chloroethyl)ether	1.000	1.009	-0.9	121	0.00
9	2-Methylphenol	1.000	0.918	8.2	96	0.00
10	3+4-Methylphenols	1.000	0.975	2.5	112	0.00
11 I	Naphthalene-d8	5.000	5.000	0.0	105	0.00
12 S	Nitrobenzene-d5	1.000	0.954	4.6	108	-0.04
13	Nitrobenzene	1.000	0.899	10.1	98	0.00
14	2-Nitrophenol	1.000	0.835	16.5	87	0.00
15	2,4-Dimethylphenol	1.000	0.974	2.6	104	0.00
16 C	2,4-Dichlorophenol	1.000	1.019	-1.9	121	0.00
17 C	Hexachlorobutadiene	1.000	1.044	-4.4	109	0.00
18	4-Chloro-3-methylphenol	1.000	0.993	0.7	115	0.00
19 I	Acenaphthene-d10	5.000	5.000	0.0	105	0.00
20 S	2,4,6-Tribromophenol	1.000	0.998	0.2	115	0.02
21	2,4,6-Trichlorophenol	1.000	1.108	-10.8	158	0.00
22	2,4,5-Trichlorophenol	1.000	1.044	-4.4	137	0.00
23 S	2-Fluorobiphenyl	1.000	1.002	-0.2	105	0.00
24 P	2,4-Dinitrophenol	1.000	0.000	100.0#	0	-14.40#
25	4-Nitrophenol	1.000	1.191	-19.1	143	-0.01
26 I	Phenanthrene-d10	5.000	5.000	0.0	119	0.00
27	4,6-Dinitro-2-methylphenol	1.000	0.859	14.1	104	0.00
28	Hexachlorobenzene	1.000	0.865	13.5	82	0.02
29 C	Pentachlorophenol	1.000	1.347	-34.7#	305	0.00
30 I	Chrysene-d12	5.000	5.000	0.0	102	0.00
31	Benzidine	1.000	1.148	-14.8	127	0.00
32 S	Terphenyl-d14	1.000	1.055	-5.5	108	0.00
33	3,3'-Dichlorobenzidine	1.000	1.170	-17.0	130	0.00
34 I	Perylene-d12	5.000	5.000	0.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1