

Data Path : Z:\HPCHEM1\BNA M\Data\BM040715\
 Data File : BM000900.D
 Acq On : 08 Apr 2015 04:20
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleID :
 SSTDCCC001EC

Quant Time: Apr 08 06:04:23 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 07 15:59:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	109	-0.04
2	1,4-Dioxane	1.000	1.013	-1.3	104	0.00
3	n-Nitrosodimethylamine	1.000	0.939	6.1	111	0.00
4 S	2-Fluorophenol	1.000	0.986	1.4	111	0.00
5 S	Phenol-d6	1.000	0.937	6.3	113	0.00
6	2-Chlorophenol	1.000	0.986	1.4	122	0.00
7 C	Phenol	1.000	0.858	14.2	99	0.00
8	bis(2-Chloroethyl)ether	1.000	0.984	1.6	125	0.00
9	2-Methylphenol	1.000	0.851	14.9	95	0.00
10	3+4-Methylphenols	1.000	0.905	9.5	108	0.00
11 I	Naphthalene-d8	5.000	5.000	0.0	108	-0.03
12 S	Nitrobenzene-d5	1.000	0.966	3.4	112	-0.04
13	Nitrobenzene	1.000	0.886	11.4	99	0.00
14	2-Nitrophenol	1.000	0.791	20.9	84	0.00
15	2,4-Dimethylphenol	1.000	0.924	7.6	101	0.00
16 C	2,4-Dichlorophenol	1.000	0.967	3.3	116	0.00
17 C	Hexachlorobutadiene	1.000	1.095	-9.5	117	0.00
18	4-Chloro-3-methylphenol	1.000	0.953	4.7	112	0.00
19 I	Acenaphthene-d10	5.000	5.000	0.0	109	0.00
20 S	2,4,6-Tribromophenol	1.000	0.900	10.0	103	0.02
21	2,4,6-Trichlorophenol	1.000	0.946	5.4	134	0.00
22	2,4,5-Trichlorophenol	1.000	0.953	4.7	127	0.00
23 S	2-Fluorobiphenyl	1.000	0.996	0.4	109	0.00
24 P	2,4-Dinitrophenol	1.000	0.000	100.0#	0	-14.40#
25	4-Nitrophenol	1.000	0.882	11.8	95	-0.01
26 I	Phenanthrene-d10	5.000	5.000	0.0	124	0.00
27	4,6-Dinitro-2-methylphenol	1.000	0.773	22.7	94	0.00
28	Hexachlorobenzene	1.000	0.855	14.5	85	0.00
29 C	Pentachlorophenol	1.000	0.964	3.6	166	0.00
30 I	Chrysene-d12	5.000	5.000	0.0	106	0.00
31	Benzidine	1.000	0.942	5.8	102	0.00
32 S	Terphenyl-d14	1.000	1.002	-0.2	106	0.00
33	3,3'-Dichlorobenzidine	1.000	0.929	7.1	100	0.00
34 I	Perylene-d12	5.000	5.000	0.0	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0