

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040715\
 Data File : BM000899.D
 Acq On : 08 Apr 2015 03:44
 Operator : TP/IZ
 Sample : G1614-06
 Misc : LOO-WATER-0.2PPM
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 LOQ-WATER2015-02

Manual Integrations
 APPROVED

apatel
 4/9/2015 8:11:50 AM

Quant Time: Apr 08 18:31:18 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 07 15:59:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	19867	5.00	ng	-0.04
11) Naphthalene-d8	10.39	136	102284	5.00	ng	0.00
19) Acenaphthene-d10	14.26	164	56049	5.00	ng	0.00
26) Phenanthrene-d10	17.00	188	143570	5.00	ng	0.00
30) Chrysene-d12	21.19	240	106751	5.00	ng	0.00
34) Perylene-d12	23.39	264	97145	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.19	112	93494	14.02	ng	-0.02
5) Phenol-d6	6.77	99	160124	11.70	ng	0.00
12) Nitrobenzene-d5	8.73	82	45742	8.97	ng	-0.04
20) 2,4,6-Tribromophenol	15.77	330	22440	9.56	ng	0.02
23) 2-Fluorobiphenyl	12.87	172	187805	10.87	ng	0.00
32) Terphenyl-d14	19.65	244	148987	11.31	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.12	88	846	0.17	ng	# 52
3) n-Nitrosodimethylamine	3.46	42	408	0.31	ng	# 1
6) 2-Chlorophenol	7.17	128	1413	0.30	ng	93
7) Phenol	6.81	94	1043	0.13	ng	66
8) bis(2-Chloroethyl)ether	7.02	93	1105	0.29	ng	84
9) 2-Methylphenol	8.06	107	819	0.13	ng	# 89
10) 3+4-Methylphenols	8.38	107	990	0.31	ng	88
13) Nitrobenzene	8.80	77	887	0.17	ng	# 68
14) 2-Nitrophenol	9.52	139	538	0.29	ng	# 49
15) 2,4-Dimethylphenol	9.58	122	652	0.11	ng	# 85
16) 2,4-Dichlorophenol	10.04	162	712	0.32	ng	# 72
17) Hexachlorobutadiene	10.72	225	645	0.14	ng	93
18) 4-Chloro-3-methylphenol	11.69	107	682	0.31	ng	# 84
21) 2,4,6-Trichlorophenol	12.69	196	375	0.32	ng	# 95
22) 2,4,5-Trichlorophenol	12.77	196	435m	0.27	ng	
25) 4-Nitrophenol	14.54	139	56m	0.17	ng	
27) 4,6-Dinitro-2-methylphenol	15.41	198	52	0.19	ng	# 1
28) Hexachlorobenzene	16.31	284	718	0.08	ng	# 9
29) Pentachlorophenol	16.67	266	71	0.55	ng	# 84
31) Benzidine	19.29	184	460	0.11	ng	# 30
33) 3,3'-Dichlorobenzidine	21.13	252	860	0.13	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM040715\
Data File : BM000899.D
Acq On : 08 Apr 2015 03:44
Operator : TP/IZ
Sample : G1614-06
Misc : LOO-WATER-0.2PPM
ALS Vial : 59 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM

Quant Time: Apr 08 18:31:18 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 07 15:59:39 2015
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.59 min Scan# 261
Delta R.T. -0.04 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02




Tot Ion	Ratio	Lower	Upper
152	100		
150	207.2	96.5	144.7#
115	172.0	11.9	17.9#

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.17 ng
RT: 3.12 min Scan# 29
Delta R.T. 0.00 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Lower	Upper
88	100		
58	34.9	56.7	85.1#
43	0.0	25.0	37.6#



Abundance Ion 88.00 (87.70 to 88.70): BM0
Ion 58.00 (57.70 to 58.70): BM0
Ion 43.00 (42.70 to 43.70): BM0





#3





n-Nitrosodimethylamine
 Concen: 0.31 ng
 RT: 3.46 min Scan# 51
 Delta R.T. 0.02 min
 Lab File: BM000899.D
 Acq: 08 Apr 2015 03:44

Instrument :
 BNA_M
 ClientSampleId :
 LOQ-WATER2015-02

Tot Ion	Ratio	Lower	Upper
42	100		
74	67.2	81.0	121.4#
44	243.7	53.3	79.9#

Manual Integrations
 APPROVED




apatel
 4/9/2015 8:11:50 AM






#4

2-Fluorophenol
 Concen: 14.02 ng
 RT: 5.19 min Scan# 163
 Delta R.T. -0.02 min
 Lab File: BM000899.D
 Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Lower	Upper
112	100		
64	78.7	35.7	53.5#
63	42.0	17.0	25.4#





#5

Phenol-d6

Concen: 11.70 ng

RT: 6.77 min Scan# 238

Delta R.T. 0.00 min

Lab File: BM000899.D





Acq: 08 Apr 2015 03:44

Tot Ion:	99	Resp:	160124
Ion Ratio	Lower	Upper	
99	100		
42	2.8	7.5	11.3#
71	16.6	21.3	31.9#

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



#6

2-Chlorophenol

Concen: 0.30 ng


RT: 7.17 min Scan# 249


Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Tgt Ion:	128	Resp:	1413
Ion Ratio	Lower	Upper	
128	100		
130	36.2	9.4	49.4
64	22.1	2.1	42.1





#7

Phenol

Concen: 0.13 ng

RT: 6.81 min Scan# 239

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M





ClientSampleId :

LOQ-WATER2015-02

Tot Ion:	94	Resp:	1043
Ion	Ratio	Lower	Upper
94	100		
65	21.0	0.0	30.9
66	32.3	0.0	35.8

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



#8

bis(2-Chloroethyl)ether

Concen: 0.29 ng




RT: 7.02 min Scan# 245




Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Tgt Ion:	93	Resp:	1105
Ion	Ratio	Lower	Upper
93	100		
63	78.0	73.3	113.3
95	0.0	0.0	20.0









#9
2-Methylphenol
Concen: 0.13 ng
RT: 8.06 min Scan# 274
Delta R.T. 0.00 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tot Ion:	107	Resp:	819
Ion	Ratio	Lower	Upper
107	100		
108	116.9	96.7	145.1
77	39.2	18.3	27.5#
79	36.5	19.9	29.9#

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02


Manual Integrations
APPROVED


apatel
4/9/2015 8:11:50 AM



#10
3+4-Methylphenols
Concen: 0.31 ng
RT: 8.38 min Scan# 283
Delta R.T. 0.00 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion:	107	Resp:	990
Ion	Ratio	Lower	Upper
107	100		
108	85.3	67.6	107.6
77	46.2	10.0	50.0
79	40.1	8.4	48.4









#11
Naphthalene-d8
Concen: 5.00 ng
RT: 10.39 min Scan# 350
Delta R.T. 0.00 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	12.4	9.6	14.4	
54	1.7	1.5	2.3	
68	1.7	1.6	2.4	




Manual Integrations
APPROVED


apatel
4/9/2015 8:11:50 AM



#12
Nitrobenzene-d5
Concen: 8.97 ng
RT: 8.73 min Scan# 293
Delta R.T. -0.04 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	3.5	96.2	144.2#	
54	153.1	15.0	22.6#	









#13
Nitrobenzene
Concen: 0.17 ng
RT: 8.80 min Scan# 295
Delta R.T. 0.00 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Tot Ion	Ratio	Lower	Upper
77	100		
123	106.2	64.0	96.0#
65	28.2	8.7	13.1#




Manual Integrations
APPROVED


apatel
4/9/2015 8:11:50 AM



#14
2-Nitrophenol
Concen: 0.29 ng
RT: 9.52 min Scan# 318
Delta R.T. 0.00 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Lower	Upper
139	100		
109	41.4	15.8	23.8#
65	45.4	16.2	24.2#





#15

2,4-Dimethylphenol

Concen: 0.11 ng

RT: 9.58 min Scan# 320

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M


ClientSampleId :

LOQ-WATER2015-02




Tot Ion:	122	Resp:	652
Ion Ratio	Lower	Upper	
122	100		
107	97.9	67.4	101.2
121	65.5	43.5	65.3#

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E



#16

2,4-Dichlorophenol

Concen: 0.32 ng


RT: 10.04 min Scan# 337

Delta R.T. 0.00 min



Lab File: BM000899.D


Acq: 08 Apr 2015 03:44

Tgt Ion:	162	Resp:	712
Ion Ratio	Lower	Upper	
162	100		
164	68.9	39.9	79.9
98	63.1	11.3	51.3#



Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BM





#17

Hexachlorobutadiene

Concen: 0.14 ng

RT: 10.72 min Scan# 362

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M


ClientSampleId :

LOQ-WATER2015-02




Tot Ion:	225	Resp:	645
Ion Ratio	Lower	Upper	
225	100		
223	78.2	58.3	87.5
227	60.9	44.5	66.7

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E



#18

4-Chloro-3-methylphenol

Concen: 0.31 ng


RT: 11.69 min Scan# 408

Delta R.T. 0.02 min



Lab File: BM000899.D


Acq: 08 Apr 2015 03:44

Tgt Ion:	107	Resp:	682
Ion Ratio	Lower	Upper	
107	100		
144	39.1	18.4	27.6#
142	84.3	59.9	89.9



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E





#19

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.26 min Scan# 562

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M





ClientSampleId :

LOQ-WATER2015-02

Tot Ion:	164	Resp:	56049
Ion	Ratio	Lower	Upper
164	100		
162	105.1	91.8	137.6
160	47.8	44.6	67.0

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



#20

2,4,6-Tribromophenol

Concen: 9.56 ng




RT: 15.77 min Scan# 638


Delta R.T. 0.02 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Tgt Ion:	330	Resp:	22440
Ion	Ratio	Lower	Upper
330	100		
332	99.0	72.6	109.0
141	0.0	29.8	44.6#





#21

2,4,6-Trichlorophenol

Concen: 0.32 ng

RT: 12.69 min Scan# 466

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M


ClientSampleId :

LOQ-WATER2015-02




Tot Ion:	196	Resp:	375
Ion Ratio	Lower	Upper	
196	100		
198	84.9	67.5	101.3
200	37.8	22.1	33.1#

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



Abundance Ion 196.00 (195.70 to 196.70): E
Ion 198.00 (197.70 to 198.70): E
Ion 200.00 (199.70 to 200.70): E



#22

2,4,5-Trichlorophenol

Concen: 0.27 ng m

RT: 12.77 min Scan# 471




Delta R.T. 0.02 min


Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Tgt Ion:	196	Resp:	435
Ion Ratio	Lower	Upper	
196	100		
198	94.9	62.7	94.1#
97	0.0	0.0	0.0
132	0.0	0.0	0.0

Abundance Ion 196.00 (195.70 to 196.70): E
Ion 198.00 (197.70 to 198.70): E
Ion 97.00 (96.70 to 97.70): BM
Ion 132.00 (131.70 to 132.70): E





#23

2-Fluorobiphenyl

Concen: 10.87 ng

RT: 12.87 min Scan# 477

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M





ClientSampleId :

LOQ-WATER2015-02

Tot Ion:	172	Resp:	187805
Ion Ratio	Lower	Upper	
172	100		
171	32.6	27.8	41.8
170	20.5	19.0	28.6

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM



#25

4-Nitrophenol

Concen: 0.17 ng m




RT: 14.54 min Scan# 580


Delta R.T. 0.03 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Tgt Ion:	139	Resp:	56
Ion Ratio	Lower	Upper	
139	100		
109	98.3	58.4	98.4
65	120.0	87.7	127.7









#26
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.00 min Scan# 686
Delta R.T. -0.01 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Tot Ion	Ratio	Resp Lower	Upper
188	100		
94	9.2	4.9	7.3#
80	9.2	4.4	6.6#




Manual Integrations
APPROVED


apatel
4/9/2015 8:11:50 AM



#27
4,6-Dinitro-2-methylphenol
Concen: 0.19 ng
RT: 15.41 min Scan# 624
Delta R.T. 0.02 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Resp Lower	Upper
198	100		
51	207.6	44.6	84.6#
105	150.0	36.0	76.0#









#28
Hexachlorobenzene
Concen: 0.08 ng
RT: 16.31 min Scan# 659
Delta R.T. -0.01 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Tot Ion	Ratio	Lower	Upper
284	100		
142	102.5	29.8	44.8#
249	84.0	31.0	46.4#




Manual Integrations
APPROVED


apatel
4/9/2015 8:11:50 AM



#29
Pentachlorophenol
Concen: 0.55 ng
RT: 16.67 min Scan# 673
Delta R.T. -0.01 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Lower	Upper
266	100		
268	94.7	61.8	92.8#
264	76.3	53.8	80.6









#30
Chrysene-d12
Concen: 5.00 ng
RT: 21.19 min Scan# 969
Delta R.T. -0.01 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Instrument :
BNA_M
ClientSampleId :
LOQ-WATER2015-02

Tot Ion	Ratio	Lower	Upper
240	100		
120	21.1	11.1	16.7#
236	30.4	20.9	31.3




Manual Integrations
APPROVED


apatel
4/9/2015 8:11:50 AM



#31
Benzidine
Concen: 0.11 ng
RT: 19.29 min Scan# 824
Delta R.T. 0.02 min
Lab File: BM000899.D
Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Lower	Upper
184	100		
185	48.4	14.4	21.6#
183	46.0	12.6	19.0#





#32

Terphenyl-d14

Concen: 11.31 ng

RT: 19.65 min Scan# 854

Delta R.T. 0.00 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M


ClientSampleId :

LOQ-WATER2015-02

Tot Ion	Ratio	Lower	Upper
244	100		
212	7.8	8.1	12.1#
122	5.8	13.6	20.4#

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM






Abundance

Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#33

3,3'-Dichlorobenzidine

Concen: 0.13 ng


RT: 21.13 min Scan# 965

Delta R.T. 0.01 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Tgt Ion	Ratio	Lower	Upper
252	100		
254	68.4	43.2	64.8#
126	24.7	15.5	23.3#








Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#34

Pervlene-d12

Concen: 5.00 ng

RT: 23.39 min Scan# 1111

Delta R.T. 0.01 min

Lab File: BM000899.D

Acq: 08 Apr 2015 03:44

Instrument :

BNA_M

ClientSampleId :

LOQ-WATER2015-02

Tot Ion: 264 Resp: 97145

Ion Ratio Lower Upper

264 100

260 23.2 21.6 32.4

265 22.9 16.1 24.1

Manual Integrations
APPROVED

apatel
4/9/2015 8:11:50 AM

