

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040716\
 Data File : BM004922.D
 Acq On : 07 Apr 2016 18:46
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02036

Quant Time: Apr 08 01:00:21 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM040116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 00:57:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.602	0.582	3.3	86	0.00
3 S	1,4-Dioxane-d8	0.494	0.494	0.0	86	0.00
4	Benzaldehyde	0.829	0.928	-11.9	82	0.00
5 S	Phenol-d5	1.777	1.566	11.9	80	0.00
6	Phenol	1.856	1.681	9.4	82	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.963	0.894	7.2	82	0.00
8	Bis(2-Chloroethyl)ether	1.406	1.293	8.0	82	0.00
9 S	2-Chlorophenol-d4	1.349	1.235	8.5	82	0.00
10	2-Chlorophenol	1.420	1.303	8.2	82	0.00
11	2-Methylphenol	1.465	1.276	12.9	80	0.00
12	2,2'-oxybis(1-Chloropropane	1.715	1.599	6.8	83	0.00
13 S	4-Methylphenol-d8	1.493	1.293	13.4	79	0.00
14	Acetophenone	2.257	2.082	7.8	81	0.00
15 P	N-Nitroso-di-n-propylamine	1.142	1.019	10.8	78	0.00
16	4-Methylphenol	1.632	1.433	12.2	80	0.00
17	Hexachloroethane	0.545	0.511	6.2	83	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
19 S	Nitrobenzene-d5	0.142	0.132	7.0	79	0.00
20	Nitrobenzene	0.340	0.324	4.7	81	0.00
21	Isophorone	0.678	0.616	9.1	76	0.00
22 S	2-Nitrophenol-d4	0.158	0.144	8.9	78	0.00
23 C	2-Nitrophenol	0.172	0.166	3.5	82	0.00
24	2,4-Dimethylphenol	0.364	0.348	4.4	80	0.00
25	Bis(2-Chloroethoxy)methane	0.422	0.394	6.6	80	0.00
26 S	2,4-Dichlorophenol-d3	0.294	0.285	3.1	81	0.00
27 C	2,4-Dichlorophenol	0.304	0.294	3.3	82	0.00
28	Naphthalene	1.011	0.970	4.1	82	0.00
29 S	4-Chloroaniline-d4	0.351	0.391	-11.4	81	0.00
30	4-Chloroaniline	0.370	0.408	-10.3	80	0.00
31 C	Hexachlorobutadiene	0.185	0.182	1.6	84	0.00
32	Caprolactam	0.115	0.059	48.7#	46	0.00
33 C	4-Chloro-3-methylphenol	0.348	0.342	1.7	82	0.00
34	2-Methylnaphthalene	0.753	0.731	2.9	82	0.00
35	1-Methylnaphthalene	0.730	0.710	2.7	83	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
37	1,2,4,5-Tetrachlorobenzene	0.619	0.579	6.5	84	0.00
38	Hexachlorocyclopentadiene	0.368	0.297	19.3	76	0.00
39 C	2,4,6-Trichlorophenol	0.409	0.370	9.5	81	0.00
40	2,4,5-Trichlorophenol	0.449	0.406	9.6	81	0.00
41	1,1'-Biphenyl	1.624	1.512	6.9	83	0.00
42	2-Chloronaphthalene	1.223	1.144	6.5	84	0.00
43	2-Nitroaniline	0.352	0.326	7.4	83	0.00
44 S	Dimethylphthalate-d6	1.578	1.510	4.3	86	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040716\
 Data File : BM004922.D
 Acq On : 07 Apr 2016 18:46
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02036

Quant Time: Apr 08 01:00:21 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM040116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 00:57:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.607	1.539	4.2	86	0.00
46	2,6-Dinitrotoluene	0.313	0.297	5.1	85	0.00
47 S	Acenaphthylene-d8	1.888	1.790	5.2	84	0.00
48	Acenaphthylene	2.009	1.927	4.1	85	0.00
49	3-Nitroaniline	0.320	0.334	-4.4	89	0.00
50 C	Acenaphthene	1.346	1.273	5.4	85	0.00
51	2,4-Dinitrophenol	0.208	0.148	28.8#	72	0.00
52 S	4-Nitrophenol-d4	0.312	0.290	7.1	85	0.00
53	4-Nitrophenol	0.256	0.248	3.1	88	0.00
54	Dibenzofuran	1.953	1.899	2.8	87	0.00
55	2,4-Dinitrotoluene	0.471	0.473	-0.4	89	0.00
56	2,3,4,6-Tetrachlorophenol	0.399	0.386	3.3	85	0.00
57	Diethylphthalate	1.627	1.567	3.7	86	0.00
58 S	Fluorene-d10	1.364	1.340	1.8	88	0.00
59	Fluorene	1.556	1.535	1.3	86	0.00
60	4-Chlorophenyl-phenylether	0.757	0.747	1.3	87	0.00
61	4-Nitroaniline	0.375	0.376	-0.3	93	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.120	0.101	15.8	86	0.00
64	4,6-Dinitro-2-methylphenol	0.128	0.112	12.5	87	0.00
65	N-Nitrosodiphenylamine	0.580	0.527	9.1	88	0.00
66	4-Bromophenyl-phenylether	0.199	0.182	8.5	88	0.00
67	Hexachlorobenzene	0.227	0.213	6.2	91	0.00
68	Atrazine	0.213	0.202	5.2	89	0.00
69 C	Pentachlorophenol	0.141	0.125	11.3	90	0.00
70	Phenanthrene	1.103	1.057	4.2	93	0.00
71 S	Anthracene-d10	0.895	0.869	2.9	94	0.00
72	Anthracene	1.106	1.075	2.8	92	0.00
73	1,2,3,4-Tetrachlorobenzene	0.260	0.225	13.5	85	0.00
74	Pentachlorobenzene	0.260	0.232	10.8	87	0.00
75	Carbazole	0.987	0.997	-1.0	93	0.00
76	Di-n-butylphthalate	1.172	1.078	8.0	87	0.00
77 C	Fluoranthene	1.225	1.315	-7.3	97	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
79 S	Pyrene-d10	0.868	0.801	7.7	98	0.00
80	Pyrene	1.145	1.056	7.8	97	0.00
81	Butylbenzylphthalate	0.485	0.416	14.2	89	0.00
82	3,3'-Dichlorobenzidine	0.390	0.372	4.6	91	0.00
83	Benzo(a)anthracene	1.147	1.079	5.9	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.694	0.605	12.8	90	0.00
85	Chrysene	1.088	1.033	5.1	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Di-n-octyl phthalate	1.184	1.120	5.4	91	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM040716\
Data File : BM004922.D
Acq On : 07 Apr 2016 18:46
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02036

Quant Time: Apr 08 01:00:21 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM040116.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 08 00:57:37 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.146	1.153	-0.6	103	0.00
89	Benzo(k)fluoranthene	1.099	1.063	3.3	95	0.00
90 S	Benzo(a)pyrene-d12	0.886	0.851	4.0	96	0.00
91 C	Benzo(a)pyrene	1.112	1.070	3.8	96	0.00
92	Indeno(1,2,3-cd)pyrene	1.326	1.114	16.0	84	-0.01
93	Dibenzo(a,h)anthracene	1.102	0.946	14.2	85	0.00
94	Benzo(a,h,i)perylene	1.139	0.901	20.9#	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0