

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040815\
 Data File : BM000902.D
 Acq On : 08 Apr 2015 10:33
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 09 03:33:30 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	90	-0.01
2	1,4-Dioxane	1.000	0.000	100.0#	0	-2.47#
3	n-Nitrosodimethylamine	1.000	0.960	4.0	88	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	90	-0.01
5 S	Nitrobenzene-d5	1.000	0.934	6.6	86	-0.01
6	Naphthalene	1.000	0.987	1.3	87	-0.01
7	2-Methylnaphthalene	1.000	1.016	-1.6	90	-0.01
8 I	Acenaphthene-d10	5.000	5.000	0.0	95	-0.02
9 S	2-Fluorobiphenyl	1.000	0.977	2.3	90	-0.01
10	Acenaphthylene	1.000	0.956	4.4	91	-0.02
11	Acenaphthene	1.000	0.950	5.0	91	0.00
12	Fluorene	1.000	0.994	0.6	93	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	97	0.00
14	Hexachlorobenzene	1.000	0.925	7.5	88	0.00
15	Pentachlorophenol	1.000	1.028	-2.8	110	-0.02
16	Phenanthrene	1.000	0.931	6.9	89	0.00
17	Anthracene	1.000	1.055	-5.5	102	-0.02
18	Fluoranthene	1.000	1.006	-0.6	98	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	99	0.00
20	Pyrene	1.000	1.012	-1.2	99	0.00
21 S	Terphenyl-d14	1.000	1.038	-3.8	100	-0.01
22	Benzo(a)anthracene	1.000	0.964	3.6	99	-0.01
23	Chrysene	1.000	1.002	-0.2	97	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.941	5.9	94	-0.02
25 I	Perylene-d12	5.000	5.000	0.0	101	-0.01
26	Benzo(b)fluoranthene	1.000	0.984	1.6	98	-0.01
27	Benzo(k)fluoranthene	1.000	0.988	1.2	97	-0.01
28 C	Benzo(a)pyrene	1.000	0.973	2.7	98	-0.02
29	Dibenzo(a,h)anthracene	1.000	0.918	8.2	91	-0.02
30	Benzo(g,h,i)perylene	1.000	0.973	2.7	96	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0