

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040815\
 Data File : BM000902.D
 Acq On : 08 Apr 2015 10:33
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 09 03:33:30 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.01
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.466	-0.6	88	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	90	-0.01
5 S	Nitrobenzene-d5	0.247	0.233	5.7	86	-0.01
6	Naphthalene	1.162	1.146	1.4	87	-0.01
7	2-Methylnaphthalene	0.734	0.746	-1.6	90	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.02
9 S	2-Fluorobiphenyl	1.551	1.515	2.3	90	-0.01
10	Acenaphthylene	2.213	2.115	4.4	91	-0.02
11	Acenaphthene	1.533	1.456	5.0	91	0.00
12	Fluorene	1.716	1.706	0.6	93	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
14	Hexachlorobenzene	0.293	0.271	7.5	88	0.00
15	Pentachlorophenol	0.069	0.052	24.6	110	-0.02
16	Phenanthrene	1.370	1.275	6.9	89	0.00
17	Anthracene	1.148	1.211	-5.5	102	-0.02
18	Fluoranthene	1.353	1.362	-0.7	98	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
20	Pyrene	1.631	1.650	-1.2	99	0.00
21 S	Terphenyl-d14	0.602	0.624	-3.7	100	-0.01
22	Benzo(a)anthracene	1.428	1.376	3.6	99	-0.01
23	Chrysene	1.497	1.500	-0.2	97	0.00
24	Indeno(1,2,3-cd)pyrene	1.549	1.459	5.8	94	-0.02
25 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
26	Benzo(b)fluoranthene	1.575	1.551	1.5	98	-0.01
27	Benzo(k)fluoranthene	1.511	1.492	1.3	97	-0.01
28 C	Benzo(a)pyrene	1.352	1.316	2.7	98	-0.02
29	Dibenzo(a,h)anthracene	1.335	1.225	8.2	91	-0.02
30	Benzo(g,h,i)perylene	1.441	1.402	2.7	96	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0