

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040815\
 Data File : BM000918.D
 Acq On : 08 Apr 2015 22:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 09 09:57:15 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.01
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.466	-0.6	92	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	94	-0.01
5 S	Nitrobenzene-d5	0.247	0.243	1.6	94	-0.01
6	Naphthalene	1.162	1.141	1.8	91	-0.01
7	2-Methylnaphthalene	0.734	0.733	0.1	93	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	98	-0.02
9 S	2-Fluorobiphenyl	1.551	1.516	2.3	93	-0.01
10	Acenaphthylene	2.213	2.195	0.8	97	-0.02
11	Acenaphthene	1.533	1.474	3.8	95	-0.02
12	Fluorene	1.716	1.717	-0.1	97	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
14	Hexachlorobenzene	0.293	0.262	10.6	91	0.00
15	Pentachlorophenol	0.069	0.074	-7.2	167#	-0.02
16	Phenanthrene	1.370	1.263	7.8	95	0.00
17	Anthracene	1.148	1.204	-4.9	109	-0.02
18	Fluoranthene	1.353	1.354	-0.1	105	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
20	Pyrene	1.631	1.638	-0.4	105	0.00
21 S	Terphenyl-d14	0.602	0.630	-4.7	108	-0.01
22	Benzo(a)anthracene	1.428	1.436	-0.6	110	-0.02
23	Chrysene	1.497	1.485	0.8	103	-0.01
24	Indeno(1,2,3-cd)pyrene	1.549	1.521	1.8	105	-0.03
25 I	Perylene-d12	1.000	1.000	0.0	110	-0.02
26	Benzo(b)fluoranthene	1.575	1.501	4.7	103	-0.02
27	Benzo(k)fluoranthene	1.511	1.548	-2.4	109	-0.02
28 C	Benzo(a)pyrene	1.352	1.339	1.0	109	-0.02
29	Dibenzo(a,h)anthracene	1.335	1.272	4.7	102	-0.03
30	Benzo(g,h,i)perylene	1.441	1.402	2.7	105	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0