

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040815\
 Data File : BM000920.D
 Acq On : 09 Apr 2015 09:19
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 09 11:18:47 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	-0.01
2	1,4-Dioxane	1.000	0.000	100.0#	0	-2.47#
3	n-Nitrosodimethylamine	1.000	0.972	2.8	104	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	105	-0.01
5 S	Nitrobenzene-d5	1.000	0.958	4.2	103	0.00
6	Naphthalene	1.000	0.989	1.1	102	-0.01
7	2-Methylnaphthalene	1.000	1.012	-1.2	104	-0.01
8 I	Acenaphthene-d10	5.000	5.000	0.0	111	-0.02
9 S	2-Fluorobiphenyl	1.000	0.980	2.0	105	-0.01
10	Acenaphthylene	1.000	0.977	2.3	109	-0.02
11	Acenaphthene	1.000	0.958	4.2	107	-0.02
12	Fluorene	1.000	1.008	-0.8	111	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	116	0.00
14	Hexachlorobenzene	1.000	0.924	7.6	104	-0.02
15	Pentachlorophenol	1.000	1.335	-33.5#	196	-0.02
16	Phenanthrene	1.000	0.962	3.8	109	0.00
17	Anthracene	1.000	1.071	-7.1	123	-0.02
18	Fluoranthene	1.000	1.043	-4.3	121	-0.01
19 I	Chrysene-d12	5.000	5.000	0.0	122	-0.01
20	Pyrene	1.000	1.006	-0.6	121	-0.01
21 S	Terphenyl-d14	1.000	1.022	-2.2	121	-0.01
22	Benzo(a)anthracene	1.000	0.993	0.7	125	-0.01
23	Chrysene	1.000	0.985	1.5	117	-0.01
24	Indeno(1,2,3-cd)pyrene	1.000	0.945	5.5	115	-0.03
25 I	Perylene-d12	5.000	5.000	0.0	125	-0.02
26	Benzo(b)fluoranthene	1.000	0.952	4.8	117	-0.02
27	Benzo(k)fluoranthene	1.000	1.023	-2.3	124	-0.02
28 C	Benzo(a)pyrene	1.000	0.986	1.4	123	-0.02
29	Dibenzo(a,h)anthracene	1.000	0.909	9.1	111	-0.03
30	Benzo(g,h,i)perylene	1.000	0.971	2.9	118	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0