

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040815\
 Data File : BM000920.D
 Acq On : 09 Apr 2015 09:19
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 09 11:18:47 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	-0.01
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.472	-1.9	104	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	105	-0.01
5 S	Nitrobenzene-d5	0.247	0.240	2.8	103	0.00
6	Naphthalene	1.162	1.149	1.1	102	-0.01
7	2-Methylnaphthalene	0.734	0.743	-1.2	104	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	111	-0.02
9 S	2-Fluorobiphenyl	1.551	1.520	2.0	105	-0.01
10	Acenaphthylene	2.213	2.163	2.3	109	-0.02
11	Acenaphthene	1.533	1.468	4.2	107	-0.02
12	Fluorene	1.716	1.731	-0.9	111	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
14	Hexachlorobenzene	0.293	0.271	7.5	104	-0.02
15	Pentachlorophenol	0.069	0.078	-13.0	196#	-0.02
16	Phenanthrene	1.370	1.317	3.9	109	0.00
17	Anthracene	1.148	1.230	-7.1	123	-0.02
18	Fluoranthene	1.353	1.412	-4.4	121	-0.01
19 I	Chrysene-d12	1.000	1.000	0.0	122	-0.01
20	Pyrene	1.631	1.640	-0.6	121	-0.01
21 S	Terphenyl-d14	0.602	0.615	-2.2	121	-0.01
22	Benzo(a)anthracene	1.428	1.418	0.7	125	-0.01
23	Chrysene	1.497	1.474	1.5	117	-0.01
24	Indeno(1,2,3-cd)pyrene	1.549	1.464	5.5	115	-0.03
25 I	Perylene-d12	1.000	1.000	0.0	125	-0.02
26	Benzo(b)fluoranthene	1.575	1.499	4.8	117	-0.02
27	Benzo(k)fluoranthene	1.511	1.546	-2.3	124	-0.02
28 C	Benzo(a)pyrene	1.352	1.333	1.4	123	-0.02
29	Dibenzo(a,h)anthracene	1.335	1.214	9.1	111	-0.03
30	Benzo(g,h,i)perylene	1.441	1.398	3.0	118	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0