

Data Path : Z:\HPCHEM1\BNA_M\Data\BM040815\
 Data File : BM000925.D
 Acq On : 09 Apr 2015 13:43
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 09 14:37:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM033115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 01 04:17:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.01
2	1,4-Dioxane	0.000	0.000	0.0	0#	-2.47#
3	n-Nitrosodimethylamine	0.463	0.465	-0.4	106	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	109	-0.01
5 S	Nitrobenzene-d5	0.247	0.253	-2.4	114	0.00
6	Naphthalene	1.162	1.143	1.6	105	-0.01
7	2-Methylnaphthalene	0.734	0.739	-0.7	108	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	115	-0.02
9 S	2-Fluorobiphenyl	1.551	1.498	3.4	108	-0.01
10	Acenaphthylene	2.213	2.205	0.4	115	-0.02
11	Acenaphthene	1.533	1.469	4.2	112	-0.02
12	Fluorene	1.716	1.718	-0.1	114	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
14	Hexachlorobenzene	0.293	0.264	9.9	104	0.00
15	Pentachlorophenol	0.069	0.092	-33.3#	237#	-0.02
16	Phenanthrene	1.370	1.293	5.6	110	0.00
17	Anthracene	1.148	1.267	-10.4	130	-0.02
18	Fluoranthene	1.353	1.416	-4.7	124	-0.01
19 I	Chrysene-d12	1.000	1.000	0.0	127	-0.01
20	Pyrene	1.631	1.639	-0.5	126	-0.01
21 S	Terphenyl-d14	0.602	0.618	-2.7	126	-0.01
22	Benzo(a)anthracene	1.428	1.432	-0.3	131	-0.02
23	Chrysene	1.497	1.469	1.9	121	-0.01
24	Indeno(1,2,3-cd)pyrene	1.549	1.515	2.2	124	-0.03
25 I	Perylene-d12	1.000	1.000	0.0	131	-0.02
26	Benzo(b)fluoranthene	1.575	1.488	5.5	122	-0.02
27	Benzo(k)fluoranthene	1.511	1.546	-2.3	130	-0.01
28 C	Benzo(a)pyrene	1.352	1.341	0.8	130	-0.02
29	Dibenzo(a,h)anthracene	1.335	1.263	5.4	121	-0.02
30	Benzo(g,h,i)perylene	1.441	1.401	2.8	125	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0