


Data Path : Z:\HPCHEM1\BNA M\DATA\BM040816\
Data File : BM004923.D
Acq On : 08 Apr 2016 10:10
Operator : UM/SJ
Sample : DFTPP48
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP48

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM032516.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 12 01:10:05 2016



AutoFind: Scans 2541, 2542, 2543; Background Corrected with Scan 2534

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.3	41261	PASS
68	69	0.00	2	1.7	766	PASS
69	198	0.00	100	31.1	43778	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.0	60557	PASS
197	198	0.00	2	0.7	980	PASS
198	198	100	100	100.0	140880	PASS
199	198	5	9	6.8	9638	PASS
275	198	10	60	26.8	37797	PASS
365	198	1	100	3.6	5067	PASS
441	443	0.01	100	79.2	21116	PASS
442	198	50	100	97.9	137949	PASS
443	442	15	24	19.3	26645	PASS

m/z	Abundance
207.00	448.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	483.0
39.10	2902.0
40.00	376.0
49.10	352.0
50.10	11112.0
51.10	42776.0
52.10	2083.0
56.00	1294.0
57.00	3115.0
61.00	529.0
62.00	675.0
63.10	1825.0
64.00	311.0
65.00	957.0
68.10	734.0
69.00	45368.0
73.00	414.0
74.10	4718.0
75.00	7448.0
76.10	2598.0
77.10	55256.0
78.10	3958.0
79.00	3517.0
80.00	2795.0
81.00	3992.0
82.00	921.0
83.00	901.0
85.00	670.0
86.00	1106.0
87.00	489.0
91.00	833.0
92.00	939.0
93.00	6483.0
94.00	404.0
96.10	384.0
98.00	4846.0
99.00	3927.0
99.90	309.0
101.00	2178.0
103.00	628.0
104.00	1319.0
105.00	1307.0
106.00	558.0
107.00	17720.0
108.00	2740.0
109.00	420.0
110.00	32232.0
111.00	4964.0
112.00	606.0
116.00	936.0
117.00	15029.0
118.00	1235.0
122.00	1060.0
123.00	1583.0
124.00	773.0
125.00	685.0
126.20	347.0
127.00	59520.0
128.10	4637.0
129.00	22992.0
130.00	2167.0
131.00	413.0
134.00	682.0
135.00	1922.0
136.00	800.0
137.00	1168.0
141.00	3231.0
142.00	1021.0
143.00	710.0
146.00	501.0
147.10	1706.0
148.00	3691.0
149.00	861.0
151.10	337.0
153.00	1024.0
154.00	815.0
155.00	1763.0
156.10	2538.0
157.10	489.0
158.00	556.0
159.00	450.0
160.00	1047.0
161.00	1428.0
162.00	524.0
165.00	1130.0
166.00	892.0
167.00	6690.0
168.00	3835.0
168.90	608.0
172.00	578.0
173.00	793.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

174.00	1292.0
175.10	2757.0
176.10	688.0
177.00	1233.0
178.00	459.0
179.00	4769.0
180.00	3712.0
181.00	1594.0
183.90	401.0
185.10	2458.0
186.10	18776.0
187.10	5292.0
188.00	652.0
189.00	1186.0
191.10	573.0
192.00	1640.0
193.10	1813.0
194.00	342.0
196.10	3965.0
197.10	863.0
198.00	133248.0
199.00	9332.0
200.10	636.0
201.40	514.0
203.00	758.0
204.10	4726.0
205.10	8210.0
206.10	35184.0
207.10	5357.0
208.00	1214.0
209.00	380.0
210.10	578.0
211.10	1380.0
215.00	455.0
216.00	762.0
217.00	8942.0
218.00	1226.0
221.10	5106.0
221.80	1223.0
223.00	2059.0
224.10	19168.0
225.10	4771.0
226.10	588.0
227.00	8029.0
228.00	1096.0
229.00	1559.0
231.00	689.0
233.90	412.0
235.00	597.0
236.00	504.0
237.00	807.0
239.10	325.0
241.00	501.0
242.00	920.0
243.00	1037.0
244.10	15158.0
245.10	2121.0
246.00	3180.0
247.00	519.0
249.00	543.0
253.00	359.0
254.20	735.0
255.10	75016.0
256.10	11409.0
257.00	981.0
258.00	4528.0
259.00	752.0
265.10	1707.0
273.00	2183.0
274.00	6070.0
275.10	32528.0
276.10	4353.0
277.00	2899.0
278.00	547.0
285.10	442.0
293.00	688.0
296.00	9403.0
297.00	1361.0
303.10	1045.0
314.10	447.0
315.00	989.0
316.00	580.0
323.10	2929.0
324.00	547.0
326.90	514.0
328.10	309.0
334.00	1810.0
335.10	497.0
341.00	383.0
346.00	622.0
352.10	832.0
353.00	646.0
354.10	1059.0
365.00	3827.0
366.00	620.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

372.10	1474.0
373.10	383.0
383.00	395.0
402.00	558.0
403.00	908.0
421.10	821.0
422.10	840.0
423.10	5328.0
424.00	1835.0
441.10	15533.0
442.10	102872.0
443.10	20560.0
444.10	1963.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	507.0
39.10	3055.0
49.10	341.0
50.10	11938.0
51.10	47136.0
52.10	2217.0
56.00	1291.0
57.00	3170.0
61.00	569.0
62.00	789.0
63.00	1912.0
64.00	344.0
65.00	1124.0
68.00	966.0
69.00	49976.0
73.00	520.0
74.10	5402.0
75.00	8531.0
76.10	2980.0
77.10	62624.0
78.10	4147.0
79.00	4087.0
80.00	2956.0
81.00	4193.0
82.00	1081.0
83.00	1041.0
85.00	836.0
86.00	1152.0
87.00	657.0
91.00	910.0
92.00	1103.0
93.00	7250.0
93.90	380.0
96.10	369.0
98.00	5596.0
99.00	4603.0
100.00	378.0
101.00	2598.0
103.00	991.0
104.00	1577.0
105.00	1558.0
106.10	463.0
107.00	20824.0
108.00	3055.0
109.00	516.0
110.00	36816.0
111.00	5591.0
112.00	640.0
116.10	1141.0
117.00	18080.0
118.00	1327.0
122.00	1247.0
123.00	1970.0
124.00	929.0
125.00	756.0
126.20	515.0
127.10	69520.0
128.00	5377.0
129.00	26536.0
130.00	2187.0
131.00	434.0
134.00	720.0
135.10	2363.0
136.00	866.0
137.10	1231.0
139.90	358.0
141.00	3626.0
142.00	1157.0
143.00	843.0
146.00	682.0
147.00	1870.0
148.00	4564.0
149.00	928.0
151.10	496.0
151.80	304.0
153.00	1262.0
154.00	860.0
155.00	2151.0
156.10	3317.0
157.10	575.0
158.00	723.0
159.00	610.0
160.00	1241.0
161.00	1858.0
161.90	455.0
164.00	309.0
165.00	1512.0
166.00	1243.0
167.00	8032.0
168.00	4381.0
169.00	835.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

171.00	351.0
172.00	687.0
173.10	892.0
174.00	1620.0
175.10	3259.0
176.00	869.0
177.00	1263.0
178.10	537.0
179.00	5904.0
180.00	4064.0
181.00	1912.0
184.00	481.0
185.10	2965.0
186.10	23320.0
187.10	6805.0
188.00	812.0
189.00	1429.0
191.10	731.0
192.00	1983.0
193.00	2348.0
194.10	489.0
196.00	4636.0
197.10	1117.0
198.00	160192.0
199.00	11172.0
200.00	872.0
201.50	648.0
203.00	1112.0
204.10	5861.0
205.10	10461.0
206.10	42560.0
207.10	5762.0
208.00	1637.0
209.00	470.0
210.10	680.0
211.00	1636.0
214.90	467.0
216.00	1135.0
217.00	11082.0
218.00	1352.0
221.10	6011.0
221.90	1534.0
223.00	2489.0
224.10	25200.0
225.10	5933.0
226.10	790.0
227.00	10454.0
228.00	1405.0
229.00	2277.0
229.90	327.0
231.10	876.0
234.00	542.0
235.00	724.0
236.00	545.0
237.10	1016.0
239.00	432.0
240.00	319.0
241.00	560.0
242.00	1287.0
243.10	1461.0
244.10	18112.0
245.10	2679.0
246.00	3647.0
247.00	856.0
249.00	617.0
253.00	439.0
254.70	969.0
255.10	97288.0
256.10	14028.0
257.00	1070.0
258.00	5622.0
259.00	961.0
265.00	2454.0
273.00	2616.0
274.10	8448.0
275.10	44304.0
276.10	5462.0
277.00	3888.0
278.00	601.0
283.00	461.0
285.10	687.0
293.00	764.0
296.10	12870.0
297.10	1907.0
303.10	1505.0
304.00	360.0
314.00	657.0
315.00	1409.0
316.00	778.0
321.00	389.0
323.10	4179.0
324.00	727.0
327.00	740.0
328.00	412.0
332.00	345.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

333.10	409.0
334.10	2560.0
335.10	644.0
341.10	517.0
346.00	978.0
352.10	1263.0
353.10	884.0
354.10	1172.0
365.00	6040.0
366.00	782.0
371.10	304.0
372.10	2240.0
373.00	494.0
383.00	611.0
402.00	898.0
403.00	1281.0
404.00	526.0
421.00	1231.0
422.10	1282.0
423.10	8255.0
424.10	2128.0
441.10	24424.0
442.10	157824.0
443.10	30400.0
444.10	2795.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	432.0
39.10	2365.0
44.00	330.0
49.00	309.0
50.00	8146.0
51.10	33872.0
52.10	1744.0
56.00	1006.0
57.00	2285.0
61.00	439.0
62.00	519.0
63.00	1602.0
65.10	765.0
68.10	599.0
69.00	35992.0
73.00	431.0
74.00	3700.0
75.00	6151.0
76.10	2143.0
77.10	44200.0
78.10	2927.0
79.00	2964.0
80.00	2146.0
81.00	3137.0
82.00	763.0
83.10	716.0
85.00	634.0
86.00	877.0
87.10	452.0
91.00	658.0
92.00	831.0
93.00	5119.0
94.00	381.0
98.00	4113.0
99.00	3230.0
100.10	333.0
101.00	1901.0
103.00	576.0
104.00	1097.0
105.00	1220.0
106.00	374.0
107.00	15164.0
108.00	2523.0
109.10	421.0
110.00	26376.0
111.00	4176.0
112.00	514.0
116.00	741.0
117.00	12512.0
118.00	901.0
122.00	935.0
123.00	1603.0
124.00	714.0
125.00	646.0
126.20	305.0
127.00	52632.0
128.00	3982.0
129.00	19896.0
130.00	1549.0
131.00	329.0
134.00	600.0
135.00	1615.0
135.90	804.0
137.00	1020.0
141.00	2765.0
142.00	866.0
143.00	595.0
146.00	387.0
147.00	1528.0
148.00	3380.0
149.00	646.0
151.10	334.0
153.00	882.0
154.00	579.0
155.00	1623.0
156.10	2599.0
157.00	425.0
157.90	491.0
159.00	360.0
160.00	955.0
161.00	1382.0
162.00	398.0
165.00	1086.0
166.00	844.0
167.00	5676.0
168.00	3343.0
169.00	532.0
172.00	574.0
173.00	800.0
174.00	1335.0
175.10	2201.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

176.00	646.0
177.00	1273.0
178.00	367.0
179.00	4326.0
180.00	3378.0
181.00	1520.0
184.00	397.0
185.10	2213.0
186.10	17560.0
187.00	4824.0
188.10	561.0
189.00	1114.0
191.00	515.0
192.00	1551.0
193.00	1829.0
194.00	336.0
196.10	3656.0
197.10	960.0
198.00	129200.0
199.00	8411.0
200.00	718.0
201.40	484.0
203.10	883.0
204.10	4889.0
205.10	8015.0
206.10	34864.0
207.00	4610.0
208.00	1215.0
209.10	364.0
210.10	538.0
211.10	1215.0
215.00	450.0
216.10	836.0
217.00	9249.0
218.00	1177.0
221.10	4620.0
221.90	1292.0
223.00	2092.0
224.10	19800.0
225.10	5196.0
226.10	557.0
227.10	8121.0
228.00	1194.0
229.00	1792.0
231.00	698.0
234.00	630.0
235.00	617.0
236.00	370.0
237.00	817.0
239.00	317.0
241.00	491.0
242.10	948.0
243.10	1075.0
244.10	15267.0
245.10	1979.0
246.10	2893.0
247.00	534.0
249.00	508.0
254.20	901.0
255.10	80936.0
256.10	11587.0
257.00	869.0
258.00	4563.0
259.00	800.0
265.10	1866.0
273.10	2139.0
274.10	6851.0
275.10	36560.0
276.10	4830.0
277.00	3336.0
278.00	490.0
283.00	346.0
285.10	515.0
293.00	769.0
296.10	10886.0
297.00	1536.0
303.10	1433.0
304.10	314.0
314.00	540.0
315.00	1260.0
316.00	657.0
321.10	415.0
323.10	3339.0
324.10	612.0
327.00	726.0
328.10	301.0
333.00	380.0
334.10	2296.0
335.00	706.0
341.00	436.0
346.00	823.0
352.10	1162.0
353.00	786.0
354.00	1185.0
365.00	5334.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

366.00	896.0
372.10	1937.0
373.00	509.0
383.00	547.0
390.10	331.0
402.00	816.0
403.00	1167.0
404.00	440.0
421.00	998.0
422.10	1125.0
423.10	8033.0
424.00	1800.0
441.10	23392.0
442.10	153152.0
443.10	28976.0
444.10	2662.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48