



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM040824\
Data File : BM045043.D
Acq On : 08 Apr 2024 16:12
Operator : MA/JU
Sample : DFTPP012
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP012

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 09 02:13:29 2024



AutoFind: Scans 2446, 2447, 2448; Background Corrected with Scan 2439

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.7	36427	PASS
68	69	0.00	2	1.7	670	PASS
69	198	0.00	100	47.3	38483	PASS
70	69	0.00	2	0.4	166	PASS
127	198	10	80	55.8	45475	PASS
197	198	0.00	2	0.5	440	PASS
198	198	100	100	100.0	81445	PASS
199	198	5	9	6.6	5402	PASS
275	198	10	60	27.3	22229	PASS
365	198	1	100	5.1	4141	PASS
441	198	0.01	100	14.4	11729	PASS
442	198	50	100	86.4	70392	PASS
443	442	15	24	19.4	13676	PASS

m/z	Abundance
39.90	619.0
43.90	873.0
73.00	571.0
76.90	242.0
83.90	239.0
85.80	199.0
95.90	249.0
132.90	210.0
134.90	169.0
146.90	167.0
190.80	192.0
193.00	237.0
206.90	1965.0
207.90	610.0
209.00	282.0
221.00	161.0
280.90	342.0

Instrument :

BNA_M

ClientSampleId :

DFTPP012

m/z	Abundance
36.10	153.0
37.00	407.0
38.00	493.0
39.00	3525.0
39.90	869.0
43.90	1115.0
49.00	588.0
50.00	12725.0
51.00	44120.0
52.00	2491.0
55.00	293.0
56.00	1305.0
57.00	2735.0
57.90	162.0
61.00	380.0
61.90	603.0
63.00	1771.0
64.00	318.0
64.90	867.0
67.90	645.0
69.00	46152.0
70.00	174.0
73.00	1228.0
74.00	5787.0
75.00	7654.0
76.00	3069.0
77.00	61288.0
78.00	4573.0
79.00	3644.0
79.90	2579.0
81.00	3440.0
82.00	982.0
82.90	805.0
83.90	397.0
84.90	684.0
86.00	1334.0
86.90	565.0
88.00	248.0
91.00	756.0
91.90	1084.0
93.00	7134.0
93.90	432.0
96.10	598.0
97.10	167.0
98.00	5540.0
99.00	4733.0
99.90	298.0
101.00	2113.0
103.00	840.0
104.00	1361.0
104.90	1343.0
105.90	375.0
107.00	16162.0
108.00	2935.0
108.90	540.0
110.00	23224.0
110.90	4478.0
112.00	771.0
112.90	175.0
115.90	745.0
116.90	18576.0
117.90	1491.0
119.90	169.0
122.00	1335.0
123.00	1554.0
123.90	945.0
124.80	660.0
126.00	302.0
127.00	51176.0
128.00	3499.0
129.00	19464.0
129.90	1489.0
130.90	236.0
131.80	224.0
132.90	272.0
133.80	443.0
134.90	1910.0
136.00	915.0
137.00	1230.0
137.80	159.0
139.90	362.0
140.90	2583.0
141.90	936.0
143.00	746.0
145.00	166.0
145.90	588.0
146.90	1687.0
148.00	4344.0
148.90	1127.0
150.00	407.0
150.90	269.0
151.70	212.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

152.90	755.0
154.00	530.0
155.00	1584.0
155.90	2094.0
157.00	427.0
157.80	472.0
158.90	435.0
160.00	879.0
160.90	1193.0
161.80	361.0
162.90	197.0
164.00	163.0
165.00	830.0
165.90	756.0
167.00	7491.0
168.00	4684.0
169.00	757.0
169.90	165.0
170.70	168.0
171.90	645.0
173.00	583.0
173.90	1175.0
174.90	1948.0
176.00	561.0
176.90	929.0
177.90	237.0
178.90	4435.0
179.90	2588.0
180.90	1258.0
181.80	182.0
184.00	411.0
185.00	2003.0
186.00	13566.0
187.00	3599.0
188.00	614.0
188.90	1117.0
191.10	520.0
191.90	1410.0
192.90	1732.0
193.90	289.0
196.00	2865.0
196.90	526.0
197.90	86720.0
198.90	5625.0
199.90	481.0
200.70	162.0
201.50	717.0
203.00	759.0
204.00	3693.0
205.00	6950.0
206.00	24992.0
207.00	5695.0
208.00	1260.0
208.90	389.0
209.80	371.0
210.90	1148.0
211.80	152.0
215.00	297.0
215.90	432.0
216.90	6611.0
217.90	943.0
220.90	3525.0
221.70	1154.0
222.90	1855.0
224.00	14944.0
225.00	3766.0
225.90	413.0
226.90	5956.0
227.90	722.0
228.90	1150.0
229.90	259.0
230.90	456.0
233.90	321.0
234.90	414.0
235.90	415.0
236.90	749.0
237.90	168.0
239.00	237.0
239.90	251.0
240.80	309.0
241.90	974.0
242.90	749.0
244.00	11677.0
245.00	1874.0
245.90	2201.0
246.90	506.0
248.90	538.0
250.90	158.0
253.10	275.0
253.90	371.0
255.00	56576.0
255.90	8826.0
257.00	905.0
257.90	3692.0
258.90	637.0
264.90	1123.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

265.80	157.0
273.00	1806.0
273.90	5120.0
275.00	21840.0
276.00	2873.0
276.90	2441.0
277.90	347.0
280.90	333.0
282.90	387.0
284.00	242.0
285.00	515.0
292.90	660.0
294.00	173.0
294.80	243.0
295.90	8347.0
296.90	1281.0
302.90	989.0
304.00	300.0
313.80	350.0
315.00	656.0
316.00	460.0
320.90	179.0
323.00	2148.0
323.80	252.0
324.10	226.0
326.90	460.0
327.90	342.0
331.80	189.0
332.90	271.0
333.90	1429.0
335.00	429.0
340.90	367.0
345.90	544.0
346.80	154.0
351.90	484.0
352.90	311.0
354.00	678.0
354.90	194.0
364.90	3575.0
366.00	537.0
370.80	222.0
372.00	1283.0
372.90	344.0
382.90	344.0
389.90	222.0
402.00	364.0
402.90	552.0
403.80	464.0
421.00	577.0
422.00	639.0
423.00	3081.0
423.90	1232.0
424.90	193.0
441.00	9325.0
442.00	57824.0
443.00	11955.0
443.90	1227.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

m/z	Abundance
36.90	243.0
38.00	589.0
39.00	2948.0
40.00	781.0
44.00	1149.0
44.90	234.0
49.00	449.0
50.00	11890.0
51.00	42616.0
52.00	2209.0
53.10	204.0
54.90	253.0
56.00	1286.0
57.00	2557.0
58.00	210.0
60.90	424.0
61.90	765.0
63.00	2143.0
64.10	254.0
65.00	927.0
67.90	924.0
69.00	44928.0
69.90	171.0
73.00	1026.0
74.00	5530.0
75.00	7712.0
76.00	3229.0
77.00	58528.0
78.00	4356.0
79.00	3485.0
79.90	2314.0
81.00	3197.0
81.90	853.0
82.90	553.0
83.90	298.0
84.90	571.0
86.00	1059.0
86.90	430.0
87.90	213.0
91.00	1021.0
91.90	841.0
93.00	6607.0
93.90	354.0
95.90	563.0
96.90	159.0
98.00	4963.0
99.00	4221.0
99.90	425.0
100.90	1978.0
101.80	209.0
102.90	691.0
104.00	1239.0
104.90	1018.0
106.00	273.0
107.00	15702.0
108.00	2569.0
108.90	378.0
110.00	23920.0
111.00	3759.0
111.90	547.0
113.00	336.0
115.90	1253.0
117.00	17848.0
117.90	1435.0
118.90	318.0
121.90	635.0
122.90	1645.0
123.90	723.0
125.00	927.0
125.90	229.0
127.00	54280.0
128.00	4680.0
129.00	21928.0
129.90	1918.0
131.00	473.0
132.90	334.0
133.90	511.0
134.90	1916.0
135.90	524.0
137.00	1306.0
137.70	254.0
139.90	371.0
140.90	3208.0
142.00	819.0
142.90	792.0
143.80	188.0
144.90	198.0
145.90	762.0
146.90	1437.0
147.90	5056.0
148.90	996.0
149.80	287.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

150.90	502.0
152.90	835.0
153.80	489.0
154.90	1780.0
155.90	2223.0
156.80	491.0
158.00	620.0
158.90	268.0
159.90	1216.0
160.90	1599.0
162.10	311.0
165.10	1079.0
166.00	769.0
166.90	7738.0
167.90	5394.0
168.90	613.0
170.00	246.0
170.80	264.0
171.90	446.0
173.00	555.0
174.00	1359.0
175.00	2554.0
175.90	591.0
176.90	910.0
177.90	393.0
178.90	4047.0
180.00	2919.0
180.90	1261.0
182.00	303.0
183.90	310.0
184.90	2241.0
186.00	14832.0
187.00	4688.0
188.00	458.0
188.90	917.0
190.00	203.0
191.00	617.0
191.90	1715.0
193.00	2189.0
193.90	518.0
194.90	239.0
195.90	2869.0
196.80	481.0
197.90	96256.0
198.90	6794.0
199.90	623.0
201.20	433.0
202.90	732.0
203.90	5062.0
205.00	7853.0
206.00	28272.0
207.00	6042.0
207.90	1408.0
208.90	571.0
210.00	522.0
211.00	1282.0
211.90	204.0
214.90	290.0
216.00	690.0
216.90	8036.0
217.90	901.0
219.90	172.0
221.00	4169.0
221.80	1241.0
222.90	1733.0
224.00	16864.0
225.00	3859.0
226.00	634.0
227.00	7223.0
227.90	870.0
228.90	1386.0
229.90	283.0
230.90	512.0
233.90	566.0
234.90	661.0
235.90	497.0
236.90	854.0
239.00	288.0
239.90	246.0
240.80	482.0
241.90	1075.0
243.00	1252.0
244.00	13451.0
245.00	1810.0
245.90	2702.0
246.90	761.0
248.90	436.0
250.90	170.0
252.00	173.0
252.90	300.0
254.00	280.0
255.00	66352.0
255.90	9798.0
256.90	746.0
258.00	3773.0
259.00	811.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

264.90	1812.0
265.90	248.0
267.00	176.0
272.90	2364.0
274.00	5089.0
275.00	26776.0
276.00	3806.0
276.90	2800.0
278.00	505.0
278.90	187.0
281.00	253.0
283.00	426.0
283.80	297.0
284.90	470.0
291.80	183.0
292.90	804.0
294.90	193.0
295.90	9881.0
296.90	1086.0
302.90	1128.0
303.90	162.0
307.90	176.0
309.00	173.0
309.80	176.0
313.90	335.0
314.90	1116.0
315.90	678.0
320.90	303.0
322.90	2749.0
324.00	489.0
326.80	637.0
328.00	276.0
331.90	253.0
332.80	216.0
334.00	2418.0
335.00	487.0
340.90	365.0
345.80	742.0
346.80	165.0
352.00	907.0
353.10	592.0
354.00	906.0
364.90	5058.0
365.90	805.0
372.00	1404.0
372.80	357.0
382.70	197.0
383.10	198.0
384.00	185.0
390.00	246.0
401.90	605.0
402.90	964.0
403.90	433.0
404.90	155.0
421.00	701.0
421.90	682.0
423.00	4554.0
424.00	1473.0
424.90	221.0
441.00	14036.0
442.00	86688.0
443.00	16202.0
444.00	1288.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

m/z	Abundance
37.00	240.0
38.00	350.0
39.00	1867.0
40.00	696.0
41.00	165.0
44.00	1216.0
44.90	217.0
49.00	346.0
50.00	6908.0
51.00	22544.0
52.00	1321.0
56.00	413.0
57.00	1809.0
61.00	238.0
61.90	432.0
62.90	1225.0
65.00	634.0
68.00	442.0
69.00	24368.0
69.90	153.0
72.90	989.0
74.00	3405.0
75.00	5132.0
76.00	2002.0
77.00	33456.0
78.00	2563.0
78.90	1774.0
80.00	1435.0
81.00	1792.0
82.00	434.0
83.00	479.0
84.00	239.0
84.90	348.0
85.90	469.0
87.00	381.0
90.90	655.0
91.90	498.0
93.00	4051.0
93.90	244.0
95.90	290.0
98.00	3188.0
99.00	2127.0
99.80	207.0
101.00	1326.0
103.00	488.0
103.90	807.0
104.90	695.0
105.90	183.0
107.00	9031.0
108.00	1256.0
109.10	250.0
110.00	13407.0
111.00	2377.0
111.90	505.0
115.90	412.0
117.00	10267.0
118.00	604.0
119.90	162.0
122.00	663.0
122.90	850.0
123.90	530.0
125.00	485.0
126.10	179.0
127.00	30968.0
128.00	2168.0
128.90	12899.0
130.00	1252.0
131.00	253.0
132.90	230.0
133.90	342.0
134.90	957.0
135.90	461.0
137.00	693.0
139.80	186.0
141.00	1727.0
141.90	540.0
143.00	284.0
145.10	164.0
146.00	259.0
146.90	999.0
147.90	2853.0
148.90	665.0
150.90	280.0
151.80	252.0
152.80	570.0
154.00	226.0
155.00	920.0
155.90	1354.0
157.00	167.0
157.80	387.0
159.10	182.0
159.80	451.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

160.90	889.0
162.00	205.0
162.80	192.0
164.80	520.0
165.90	507.0
167.00	4931.0
167.90	3175.0
168.90	426.0
170.80	155.0
171.90	519.0
172.90	425.0
173.90	844.0
174.90	1284.0
175.90	378.0
176.90	526.0
177.80	238.0
178.90	2891.0
179.90	2126.0
180.90	1090.0
183.70	223.0
184.90	1335.0
186.00	9434.0
186.90	2520.0
187.70	164.0
188.90	666.0
191.00	482.0
191.90	914.0
193.00	988.0
193.80	235.0
194.90	166.0
195.90	1724.0
196.90	312.0
197.90	61360.0
198.90	3788.0
199.90	335.0
201.40	224.0
202.80	487.0
204.00	3027.0
205.00	5263.0
206.00	18352.0
207.00	4181.0
207.90	868.0
209.00	538.0
210.00	286.0
210.30	297.0
210.90	964.0
215.00	283.0
215.90	359.0
216.90	4716.0
217.90	892.0
220.90	2748.0
221.80	701.0
222.90	1502.0
224.00	11140.0
225.00	2537.0
226.00	260.0
226.90	4372.0
227.90	574.0
229.00	1026.0
229.90	202.0
230.90	385.0
233.90	307.0
234.90	370.0
235.80	355.0
236.90	545.0
239.90	202.0
240.80	319.0
241.90	590.0
243.00	687.0
244.00	8974.0
244.90	1082.0
245.90	1409.0
246.80	243.0
247.90	150.0
248.90	350.0
252.90	169.0
254.10	154.0
255.00	44168.0
256.00	6516.0
256.90	547.0
257.90	2782.0
259.00	392.0
264.90	1317.0
266.10	153.0
272.90	1323.0
274.00	3424.0
275.00	18072.0
276.00	2680.0
276.90	1625.0
277.90	462.0
281.00	397.0
282.00	163.0
283.00	235.0
283.90	206.0
285.00	318.0
292.90	414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012

295.90	6095.0
296.90	1157.0
302.00	153.0
302.90	826.0
303.90	166.0
307.90	162.0
313.90	301.0
314.90	917.0
315.90	428.0
321.00	156.0
323.00	2187.0
324.00	256.0
327.00	553.0
327.90	305.0
332.10	152.0
333.00	238.0
333.90	1003.0
334.90	193.0
341.00	182.0
345.90	584.0
346.90	168.0
352.00	905.0
353.00	515.0
353.90	652.0
355.00	158.0
365.00	3789.0
365.90	439.0
372.00	1409.0
373.00	265.0
382.90	424.0
383.90	186.0
389.90	310.0
391.00	208.0
401.90	567.0
403.00	524.0
404.00	239.0
421.00	722.0
421.90	536.0
423.00	3887.0
424.00	1097.0
441.00	11826.0
442.00	66664.0
443.00	12871.0
443.90	1150.0

Instrument :
BNA_M
ClientSampleId :
DFTPP012