

Data Path : Z:\HPCHEM1\BNA M\Data\BM040915\
 Data File : BM000935.D
 Acq On : 09 Apr 2015 23:09
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM040915

Quant Time: Apr 10 16:34:28 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 10 16:15:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.497	0.499	-0.4	106	0.00
3	n-Nitrosodimethylamine	0.486	0.498	-2.5	108	0.00
4 S	2-Fluorophenol	1.058	1.125	-6.3	109	0.00
5 S	Phenol-d6	1.267	1.343	-6.0	112	0.00
6 C	Phenol	1.302	1.367	-5.0	111	0.00
7	bis(2-Chloroethyl)ether	1.142	1.215	-6.4	110	0.00
8 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
9 S	Nitrobenzene-d5	0.241	0.259	-7.5	112	0.00
10	Nitrobenzene	0.283	0.309	-9.2	112	0.00
11	2,4-Dimethylphenol	0.275	0.295	-7.3	110	0.00
12 C	2,4-Dichlorophenol	0.243	0.257	-5.8	112	0.00
13	Naphthalene	1.009	1.058	-4.9	107	0.00
14 C	Hexachlorobutadiene	0.184	0.197	-7.1	109	0.00
15 C	2-Methylnaphthalene	0.689	0.738	-7.1	109	0.00
16	1-Methylnaphthalene	0.694	0.751	-8.2	110	-0.01
17 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.01
18 S	2,4,6-Tribromophenol	0.179	0.164	8.4	93	0.01
19 S	2-Fluorobiphenyl	1.401	1.437	-2.6	110	0.00
20	Acenaphthylene	2.032	1.913	5.9	88	0.01
21 C	Acenaphthene	1.380	1.460	-5.8	123	-0.02
22 P	2,4-Dinitrophenol	0.058	0.046#	20.7	140	-0.01
23	Fluorene	1.606	1.810	-12.7	134	-0.02
24 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.01
25	Hexachlorobenzene	0.270	0.277	-2.6	105	0.01
26 C	Pentachlorophenol	0.086	0.080	7.0	111	0.00
27	Phenanthrene	1.292	1.325	-2.6	104	0.01
28	Anthracene	1.184	1.248	-5.4	117	-0.01
29 C	Fluoranthene	1.350	1.409	-4.4	108	-0.01
30 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
31	Benzidine	0.300	0.295	1.7	121	0.00
32	Pyrene	1.625	1.737	-6.9	108	0.00
33 S	Terphenyl-d14	0.609	0.669	-9.9	114	0.00
34	Benzo(a)anthracene	1.443	1.545	-7.1	113	-0.01
35	3,3'-Dichlorobenzidine	0.389	0.412	-5.9	119	0.00
36	Chrysene	1.475	1.555	-5.4	106	0.00
37	Indeno(1,2,3-cd)pyrene	1.506	1.642	-9.0	112	0.00
38 I	Perylene-d12	1.000	1.000	0.0	102	0.00
39	Benzo(b)fluoranthene	1.565	1.754	-12.1	120	0.00
40	Benzo(k)fluoranthene	1.506	1.534	-1.9	101	0.00
41 C	Benzo(a)pyrene	1.378	1.468	-6.5	111	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM040915\
Data File : BM000935.D
Acq On : 09 Apr 2015 23:09
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM040915

Quant Time: Apr 10 16:34:28 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 10 16:15:17 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
42	Dibenzo(a,h)anthracene	1.278	1.412	-10.5	114	0.00
43	Benzo(q,h,i)perylene	1.410	1.514	-7.4	110	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0