

Data Path : Z:\HPCHEM1\BNA M\Data\BM040915\
 Data File : BM000935.D
 Acq On : 09 Apr 2015 23:09
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM040915

Quant Time: Apr 10 16:34:28 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 10 16:15:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	0.00
2	1,4-Dioxane	1.000	1.072	-7.2	106	0.00
3	n-Nitrosodimethylamine	1.000	1.025	-2.5	108	0.00
4 S	2-Fluorophenol	1.000	1.063	-6.3	109	0.00
5 S	Phenol-d6	1.000	1.060	-6.0	112	0.00
6 C	Phenol	1.000	1.050	-5.0	111	0.00
7	bis(2-Chloroethyl)ether	1.000	1.064	-6.4	110	0.00
8 I	Naphthalene-d8	5.000	5.000	0.0	99	0.00
9 S	Nitrobenzene-d5	1.000	1.075	-7.5	112	0.00
10	Nitrobenzene	1.000	1.027	-2.7	112	0.00
11	2,4-Dimethylphenol	1.000	1.069	-6.9	110	0.00
12 C	2,4-Dichlorophenol	1.000	1.059	-5.9	112	0.00
13	Naphthalene	1.000	1.048	-4.8	107	0.00
14 C	Hexachlorobutadiene	1.000	1.067	-6.7	109	0.00
15 C	2-Methylnaphthalene	1.000	1.072	-7.2	109	0.00
16	1-Methylnaphthalene	1.000	1.082	-8.2	110	-0.01
17 I	Acenaphthene-d10	5.000	5.000	0.0	125	0.01
18 S	2,4,6-Tribromophenol	1.000	0.908	9.2	93	0.01
19 S	2-Fluorobiphenyl	1.000	1.026	-2.6	110	0.00
20	Acenaphthylene	1.000	0.898	10.2	88	0.01
21 C	Acenaphthene	1.000	1.058	-5.8	123	-0.02
22 P	2,4-Dinitrophenol	1.000	1.015	-1.5	140	-0.01
23	Fluorene	1.000	1.127	-12.7	134	-0.02
24 I	Phenanthrene-d10	5.000	5.000	0.0	107	0.01
25	Hexachlorobenzene	1.000	1.026	-2.6	105	0.01
26 C	Pentachlorophenol	1.000	0.900	10.0	111	0.00
27	Phenanthrene	1.000	1.026	-2.6	104	0.01
28	Anthracene	1.000	1.055	-5.5	117	-0.01
29 C	Fluoranthene	1.000	1.043	-4.3	108	-0.01
30 I	Chrysene-d12	5.000	5.000	0.0	102	0.00
31	Benzidine	1.000	1.095	-9.5	121	0.00
32	Pyrene	1.000	1.069	-6.9	108	0.00
33 S	Terphenyl-d14	1.000	1.098	-9.8	114	0.00
34	Benzo(a)anthracene	1.000	1.071	-7.1	113	-0.01
35	3,3'-Dichlorobenzidine	1.000	1.004	-0.4	119	0.00
36	Chrysene	1.000	1.054	-5.4	106	0.00
37	Indeno(1,2,3-cd)pyrene	1.000	1.090	-9.0	112	0.00
38 I	Perylene-d12	5.000	5.000	0.0	102	0.00
39	Benzo(b)fluoranthene	1.000	1.120	-12.0	120	0.00
40	Benzo(k)fluoranthene	1.000	1.018	-1.8	101	0.00
41 C	Benzo(a)pyrene	1.000	1.065	-6.5	111	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM040915\
Data File : BM000935.D
Acq On : 09 Apr 2015 23:09
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM040915

Quant Time: Apr 10 16:34:28 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-BM040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 10 16:15:17 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area	% Dev(min)
42	Dibenzo(a,h)anthracene	1.000	1.105	-10.5	114	0.00
43	Benzo(q,h,i)perylene	1.000	1.073	-7.3	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0