

Data Path : S:\HPCHEM1\BNA_M\Data\BM041015\
 Data File : BM000937.D
 Acq On : 10 Apr 2015 10:48
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 10 23:13:47 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 10 16:15:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.497	0.464	6.6	103	0.00
3	n-Nitrosodimethylamine	0.486	0.463	4.7	104	0.00
4 S	2-Fluorophenol	1.058	1.051	0.7	106	-0.02
5 S	Phenol-d6	1.267	1.236	2.4	107	0.00
6 C	Phenol	1.302	1.273	2.2	108	-0.02
7	bis(2-Chloroethyl)ether	1.142	1.126	1.4	106	0.00
8 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
9 S	Nitrobenzene-d5	0.241	0.232	3.7	106	0.00
10	Nitrobenzene	0.283	0.281	0.7	108	0.00
11	2,4-Dimethylphenol	0.275	0.268	2.5	106	0.00
12 C	2,4-Dichlorophenol	0.243	0.234	3.7	107	0.00
13	Naphthalene	1.009	1.017	-0.8	108	0.00
14 C	Hexachlorobutadiene	0.184	0.180	2.2	105	0.00
15 C	2-Methylnaphthalene	0.689	0.691	-0.3	107	0.00
16	1-Methylnaphthalene	0.694	0.694	0.0	107	0.00
17 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.01
18 S	2,4,6-Tribromophenol	0.179	0.151	15.6	90	0.01
19 S	2-Fluorobiphenyl	1.401	1.336	4.6	108	0.00
20	Acenaphthylene	2.032	1.820	10.4	89	0.01
21 C	Acenaphthene	1.380	1.312	4.9	117	-0.02
22 P	2,4-Dinitrophenol	0.058	0.038#	34.5#	124	-0.01
23	Fluorene	1.606	1.602	0.2	126	-0.02
24 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.01
25	Hexachlorobenzene	0.270	0.267	1.1	107	0.01
26 C	Pentachlorophenol	0.086	0.079	8.1	117	0.01
27	Phenanthrene	1.292	1.255	2.9	104	0.01
28	Anthracene	1.184	1.167	1.4	116	-0.01
29 C	Fluoranthene	1.350	1.310	3.0	107	-0.01
30 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
31	Benzidine	0.300	0.299	0.3	133	0.00
32	Pyrene	1.625	1.604	1.3	107	0.00
33 S	Terphenyl-d14	0.609	0.623	-2.3	115	0.00
34	Benzo(a)anthracene	1.443	1.367	5.3	107	0.00
35	3,3'-Dichlorobenzidine	0.389	0.352	9.5	109	0.00
36	Chrysene	1.475	1.491	-1.1	110	0.00
37	Indeno(1,2,3-cd)pyrene	1.506	1.460	3.1	107	0.00
38 I	Perylene-d12	1.000	1.000	0.0	109	0.01
39	Benzo(b)fluoranthene	1.565	1.629	-4.1	120	0.01
40	Benzo(k)fluoranthene	1.506	1.415	6.0	100	0.01
41 C	Benzo(a)pyrene	1.378	1.361	1.2	110	0.00

Data Path : S:\HPCHEM1\BNA_M\Data\BM041015\
Data File : BM000937.D
Acq On : 10 Apr 2015 10:48
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Apr 10 23:13:47 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 10 16:15:17 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
42	Dibenzo(a,h)anthracene	1.278	1.238	3.1	107	0.01
43	Benzo(g,h,i)perylene	1.410	1.378	2.3	107	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0