


Data Path : S:\HPCHEM1\BNA_M\Data\BM041015\
 Data File : BM000936.D
 Acq On : 10 Apr 2015 10:07
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 10 16:15:17 2015



AutoFind: Scans 2511, 2512, 2513; Background Corrected with Scan 2504

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.8	187072	PASS
68	69	0.00	2	1.6	2745	PASS
69	198	0.00	100	39.0	174397	PASS
70	69	0.00	2	0.5	787	PASS
127	198	10	80	48.4	216597	PASS
197	198	0.00	2	0.3	1414	PASS
198	198	100	100	100.0	447381	PASS
199	198	5	9	6.7	29901	PASS
275	198	10	60	23.8	106437	PASS
365	198	1	100	2.8	12451	PASS
441	442	0.01	24	14.8	59805	PASS
442	198	50	100	90.1	403093	PASS
443	442	15	24	19.0	76640	PASS