

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041023\
 Data File : BM039386.D
 Acq On : 10 Apr 2023 20:30
 Operator : CG/JU
 Sample : 02222-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MID-H4(ELEV-24)

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 04/11/2023
 Supervised By : Jagrut Upadhyay 04/11/2023

Quant Time: Apr 11 02:14:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 11 02:05:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	284577	20.000	ng	0.00
21) Naphthalene-d8	10.681	136	1100407	20.000	ng	0.00
39) Acenaphthene-d10	14.522	164	545715	20.000	ng	0.00
64) Phenanthrene-d10	17.275	188	945195	20.000	ng	0.00
76) Chrysene-d12	21.468	240	850861	20.000	ng	0.00
86) Perylene-d12	23.862	264	1049090	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	1960859	103.811	ng	0.00
7) Phenol-d6	7.046	99	2511635	102.805	ng	0.00
23) Nitrobenzene-d5	9.040	82	1495529	66.054	ng	0.00
42) 2,4,6-Tribromophenol	16.016	330	582315	84.061	ng	0.00
45) 2-Fluorobiphenyl	13.145	172	2841136	70.540	ng	0.00
79) Terphenyl-d14	19.904	244	3192106	69.313	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.734	128	6175832	104.156	ng	99
37) 2-Methylnaphthalene	12.346	142	1645463	39.864	ng	99
38) 1-Methylnaphthalene	12.563	142	1214355	31.419	ng	100
46) 1,1'-Biphenyl	13.357	154	462882	10.741	ng	100
49) Acenaphthylene	14.245	152	276620	5.158	ng	# 91
52) Acenaphthene	14.587	154	2210418	69.803	ng	100
55) Dibenzofuran	14.922	168	2202072	43.194	ng	99
58) Fluorene	15.575	166	2867090	71.021	ng	99
71) Phenanthrene	17.328	178	18551014m	358.357	ng	
72) Anthracene	17.416	178	5619253	106.455	ng	99
73) Carbazole	17.686	167	1734121	35.158	ng	100
75) Fluoranthene	19.345	202	16913240	286.616	ng	88
78) Pyrene	19.710	202	17030255	281.156	ng	# 86
81) Benzo(a)anthracene	21.457	228	8075658	134.576	ng	96
83) Chrysene	21.510	228	7854157m	135.581	ng	
87) Indeno(1,2,3-cd)pyrene	26.351	276	4318966	55.875	ng	96
88) Benzo(b)fluoranthene	23.145	252	8712858	135.732	ng	97
89) Benzo(k)fluoranthene	23.180	252	3188105m	48.759	ng	
90) Benzo(a)pyrene	23.762	252	7349185	116.543	ng	98
91) Dibenzo(a,h)anthracene	26.351	278	1088117m	16.765	ng	
92) Benzo(g,h,i)perylene	27.121	276	4387727	68.741	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041023\
Data File : BM039386.D
Acq On : 10 Apr 2023 20:30
Operator : CG/JU
Sample : 02222-02
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)

Quant Time: Apr 11 02:14:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 11 02:05:37 2023
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : Christian Giraldo 04/11/2023
Supervised By : Jagrut Upadhyay 04/11/2023





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.869 min Scan# 796
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)




Tgt Ion:152 Resp: 284572
Ion Ratio Lower Upper
152 100
150 152.5 125.5 188.3
115 60.1 48.1 72.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023




#5
2-Fluorophenol
Concen: 103.811 ng
RT: 5.428 min Scan# 381
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30



Tgt Ion:112 Resp: 1960859
Ion Ratio Lower Upper
112 100
64 58.9 44.3 66.5
63 30.2 23.0 34.6





#7

Phenol-d6

Concen: 102.805 ng

RT: 7.046 min Scan# 61

Delta R.T. 0.000 min

Lab File: BM039386.D


Acq: 10 Apr 2023 20:30

Instrument :

BNA_M

ClientSampleId :

MID-H4(ELEV-24)



Tgt Ion: 99 Resp: 2511635

Ion Ratio Lower Upper

99 100

42 16.9 13.0 19.6




71 32.4 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/11/2023

Supervised By :Jagrut Upadhyay 04/11/2023



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.681 min Scan# 1274

Delta R.T. 0.000 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30

Tgt Ion:136 Resp: 1100407




Ion Ratio Lower Upper


136 100

137 11.1 8.9 13.3

54 8.1 6.2 9.4


68 6.6 4.9 7.3





#23
Nitrobenzene-d5
Concen: 66.054 ng
RT: 9.040 min Scan# 995
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)




Tgt Ion: 82 Resp: 1495529
Ion Ratio Lower Upper
82 100
128 42.8 35.1 52.7
54 48.3 38.5 57.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023




#31
Naphthalene
Concen: 104.156 ng
RT: 10.734 min Scan# 1283
Delta R.T. 0.006 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Tgt Ion:128 Resp: 6175832
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.1 10.3 15.5





#37
2-Methylnaphthalene
Concen: 39.864 ng
RT: 12.346 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)




Tgt Ion:142 Resp: 1645463
Ion Ratio Lower Upper
142 100
141 88.6 69.8 104.8
115 35.6 27.8 41.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023




#38
1-Methylnaphthalene
Concen: 31.419 ng
RT: 12.563 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Tgt Ion:142 Resp: 1214355
Ion Ratio Lower Upper
142 100
141 91.8 73.2 109.8
116 3.8 3.0 4.4





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.522 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)




Tgt Ion:164 Resp: 545715
Ion Ratio Lower Upper
164 100
162 99.1 79.1 118.7
160 43.7 35.8 53.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023




#42
2,4,6-Tribromophenol
Concen: 84.061 ng
RT: 16.016 min Scan# 2181
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30



Tgt Ion:330 Resp: 582315
Ion Ratio Lower Upper
330 100
332 96.8 77.4 116.0
141 39.0 29.8 44.6





#45

2-Fluorobiphenyl

Concen: 70.540 ng

RT: 13.145 min Scan# 1685

Delta R.T. 0.000 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30

Instrument :

BNA_M

ClientSampleId :

MID-H4(ELEV-24)

Tgt Ion:172 Resp: 2841136

Ion Ratio Lower Upper

172 100

171 35.1 28.0 42.0





170 23.2 18.7 28.1

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/11/2023

Supervised By :Jagrut Upadhyay 04/11/2023



#46

1,1'-Biphenyl

Concen: 10.741 ng

RT: 13.357 min Scan# 1729

Delta R.T. 0.000 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30




Tgt Ion:154 Resp: 462882


Ion Ratio Lower Upper

154 100

153 39.8 19.9 59.9

76 13.8 0.0 33.5





#49

Acenaphthylene

Concen: 5.158 ng

RT: 14.245 min Scan# 1872

Delta R.T. 0.000 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30

Instrument :

BNA_M

ClientSampleId :





MID-H4(ELEV-24)

Tgt Ion:152	Resp: 276620
Ion Ratio	Lower Upper
152 100	
151 20.9	15.7 23.5
153 20.6	10.3 15.5

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#52

Acenaphthene

Concen: 69.803 ng




RT: 14.587 min Scan# 1938


Delta R.T. 0.000 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30


Tgt Ion:154	Resp: 2210418
Ion Ratio	Lower Upper
154 100	
153 111.2	89.1 133.7
152 53.6	42.6 64.0





#55
Dibenzenofuran
Concen: 43.194 ng
RT: 14.922 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)



Tgt Ion:168 Resp: 2202072
Ion Ratio Lower Upper
168 100
139 37.3 29.4 44.2
169 14.3 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#58
Fluorene
Concen: 71.021 ng
RT: 15.575 min Scan# 2106
Delta R.T. 0.006 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Tgt Ion:166 Resp: 2867090
Ion Ratio Lower Upper
166 100
165 97.4 76.9 115.3
167 13.9 10.8 16.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.275 min Scan# 2386
Delta R.T. 0.006 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)

Tgt Ion:188 Resp: 945195
Ion Ratio Lower Upper
188 100
94 10.5 8.2 12.4
80 11.9 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#71
Phenanthrene
Concen: 358.357 ng m
RT: 17.328 min Scan# 2404
Delta R.T. 0.012 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Tgt Ion:178 Resp:18551014
Ion Ratio Lower Upper
178 100
176 25.6 15.7 23.5#
179 20.4 12.4 18.6#





#72

Anthracene

Concen: 106.455 ng

RT: 17.416 min Scan# 24

Delta R.T. 0.012 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30

Instrument :

BNA_M

ClientSampleId :

MID-H4(ELEV-24)

Tgt Ion:178 Resp: 561925

Ion Ratio Lower Upper

178 100

176 19.1 14.9 22.3





179 15.6 12.2 18.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/11/2023

Supervised By :Jagrut Upadhyay 04/11/2023



#73

Carbazole

Concen: 35.158 ng

RT: 17.686 min Scan# 2465

Delta R.T. 0.006 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30




Tgt Ion:167 Resp: 1734121


Ion Ratio Lower Upper

167 100

166 21.4 17.2 25.8


139 13.3 10.6 16.0





#75
Fluoranthene
Concen: 286.616 ng
RT: 19.345 min Scan# 21
Delta R.T. 0.012 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)



Tgt Ion:202 Resp:16913240
Ion Ratio Lower Upper
202 100
101 16.0 0.0 32.5
203 23.4 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.468 min Scan# 3108
Delta R.T. 0.006 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Tgt Ion:240 Resp: 850861
Ion Ratio Lower Upper
240 100
120 13.1 10.2 15.2
236 26.5 20.9 31.3









#78
Pyrene
Concen: 281.156 ng
RT: 19.710 min Scan# 2809
Delta R.T. 0.012 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)

Tgt Ion:202 Resp:1703025
Ion Ratio Lower Upper
202 100
200 27.9 17.1 25.7
203 23.8 13.9 20.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#79
Terphenyl-d14
Concen: 69.313 ng
RT: 19.904 min Scan# 2842
Delta R.T. 0.006 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30


Tgt Ion:244 Resp: 3192106
Ion Ratio Lower Upper
244 100
212 7.0 5.6 8.4
122 11.2 9.1 13.7





#81
Benzo(a)anthracene
Concen: 134.576 ng
RT: 21.457 min Scan# 3106
Delta R.T. 0.012 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)



Tgt Ion:228 Resp: 8075658
Ion Ratio Lower Upper
228 100
226 29.0 21.9 32.9
229 21.7 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#83
Chrysene
Concen: 135.581 ng m
RT: 21.510 min Scan# 3115
Delta R.T. 0.012 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Tgt Ion:228 Resp: 7854157
Ion Ratio Lower Upper
228 100
226 31.6 24.1 36.1
229 23.1 15.5 23.3





#86

Perylene-d12

Concen: 20.000 ng

RT: 23.862 min Scan# 3515

Delta R.T. 0.018 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30

Instrument :

BNA_M

ClientSampleId :

MID-H4(ELEV-24)

Tgt Ion:264 Resp: 1049090

Ion Ratio Lower Upper

264 100

260 24.4 19.0 28.6





265 25.6 18.1 27.1

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/11/2023

Supervised By :Jagrut Upadhyay 04/11/2023



#87

Indeno(1,2,3-cd)pyrene

Concen: 55.875 ng

RT: 26.351 min Scan# 3938

Delta R.T. 0.024 min

Lab File: BM039386.D

Acq: 10 Apr 2023 20:30




Tgt Ion:276 Resp: 4318966


Ion Ratio Lower Upper

276 100

138 23.9 22.2 33.2

277 25.1 20.2 30.2









#88
Benzo(b)fluoranthene
Concen: 135.732 ng
RT: 23.145 min Scan# 3383
Delta R.T. 0.024 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)

Tgt Ion:252 Resp: 8712858
Ion Ratio Lower Upper
252 100
253 23.4 17.6 26.4
125 12.3 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#89
Benzo(k)fluoranthene
Concen: 48.759 ng m
RT: 23.180 min Scan# 3399
Delta R.T. 0.012 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30


Tgt Ion:252 Resp: 3188105
Ion Ratio Lower Upper
252 100
253 23.7 17.8 26.8
125 12.0 8.5 12.7





#90
Benzo(a)pyrene
Concen: 116.543 ng
RT: 23.762 min Scan# 3490
Delta R.T. 0.018 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30




Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)



Tgt Ion:252 Resp: 7349185
Ion Ratio Lower Upper
252 100
253 23.5 17.6 26.4
125 12.4 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023



#91
Dibenzo(a,h)anthracene
Concen: 16.765 ng m
RT: 26.351 min Scan# 3938
Delta R.T. 0.006 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30


Tgt Ion:278 Resp: 1088117
Ion Ratio Lower Upper
278 100
139 0.0 13.9 20.9#
279 26.6 18.8 28.2





#92
Benzo(g,h,i)perylene
Concen: 68.741 ng
RT: 27.121 min Scan# 4069
Delta R.T. 0.036 min
Lab File: BM039386.D
Acq: 10 Apr 2023 20:30

Instrument :
BNA_M
ClientSampleId :
MID-H4(ELEV-24)



Tgt Ion:276 Resp: 438772
Ion Ratio Lower Upper
276 100
277 24.5 19.1 28.7
138 23.8 19.5 29.3

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 04/11/2023
Supervised By :Jagrut Upadhyay 04/11/2023

