

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041116\
 Data File : BM004937.D
 Acq On : 11 Apr 2016 14:19
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.440

Quant Time: Apr 12 01:33:57 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 12 01:10:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2 SURR	1,4-Dioxane-d8	0.180	0.187	-3.9	123	0.00
3	1,4-Dioxane	0.220	0.227	-3.2	121	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
5	Naphthalene	0.966	0.959	0.7	109	0.00
6 SURR	2-Methylnaphthalene-d10	0.558	0.506	9.3	100	0.00
7	2-Methylnaphthalene	0.692	0.668	3.5	104	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
9	Acenaphthylene	1.840	1.662	9.7	90	0.00
10 C	Acenaphthene	1.302	1.350	-3.7	103	0.00
11	Fluorene	1.554	1.574	-1.3	101	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
13	Pentachlorophenol	0.100	0.065	35.0#	66	0.00
14	Phenanthrene	1.124	1.159	-3.1	104	0.00
15	Anthracene	1.062	0.995	6.3	95	0.00
16 SURR	Fluoranthene-d10	0.946	0.939	0.7	100	0.00
17 C	Fluoranthene	1.279	1.315	-2.8	104	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
19	Pyrene	1.370	1.353	1.2	103	0.00
20	Benzo(a)anthracene	1.230	1.183	3.8	99	0.00
21	Chrysene	1.188	1.213	-2.1	104	0.00
22 I	Perylene-d12	1.000	1.000	0.0	94	0.00
23	Benzo(b)fluoranthene	1.348	1.448	-7.4	104	0.00
24	Benzo(k)fluoranthene	1.249	1.297	-3.8	96	0.00
25 C	Benzo(a)pyrene	1.232	1.253	-1.7	95	0.00
26	Indeno(1,2,3-cd)pyrene	1.140	1.219	-6.9	101	0.00
27	Dibenzo(a,h)anthracene	0.898	0.974	-8.5	103	0.00
28	Benzo(q,h,i)perylene	0.944	1.035	-9.6	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0