


Data Path : Z:\HPCHEM1\BNA M\Data\BM041116\
 Data File : BM004931.D
 Acq On : 11 Apr 2016 09:24
 Operator : UM/SJ
 Sample : DFTPP49
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP49

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM032516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Apr 12 01:10:05 2016



AutoFind: Scans 2540, 2541, 2542; Background Corrected with Scan 2534

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.8	44378	PASS
68	69	0.00	2	1.8	847	PASS
69	198	0.00	100	31.4	46760	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.5	64765	PASS
197	198	0.00	2	0.7	1111	PASS
198	198	100	100	100.0	149024	PASS
199	198	5	9	6.8	10172	PASS
275	198	10	60	26.3	39240	PASS
365	198	1	100	3.5	5174	PASS
441	443	0.01	100	77.5	21677	PASS
442	198	50	100	95.7	142589	PASS
443	442	15	24	19.6	27973	PASS

Scan 2534 (17.592 min): BM004931.D
DFTPP49

m/z	Abundance
44.00	321.0
73.00	311.0
207.00	451.0

Instrument :

BNA_M

ClientSampleId :

DFTPP49

m/z	Abundance
38.00	391.0
39.10	2749.0
43.90	308.0
49.10	324.0
50.10	10274.0
51.10	41320.0
52.00	2167.0
55.00	303.0
56.00	1186.0
57.00	2894.0
61.00	555.0
62.00	583.0
63.10	1734.0
65.10	818.0
68.10	703.0
69.00	42712.0
73.10	651.0
74.10	4745.0
75.00	7089.0
76.10	2549.0
77.10	52584.0
78.10	3964.0
79.00	3409.0
80.00	2805.0
81.00	3659.0
82.00	907.0
83.10	920.0
85.00	750.0
86.00	945.0
87.00	503.0
91.00	881.0
92.00	1010.0
93.00	5974.0
94.00	438.0
96.00	321.0
98.00	4812.0
99.00	3999.0
100.00	371.0
101.00	1975.0
103.00	639.0
104.00	1297.0
105.00	1213.0
106.00	480.0
107.00	16464.0
108.00	2485.0
109.10	476.0
110.00	29568.0
111.00	4333.0
112.00	536.0
116.00	899.0
117.00	14141.0
118.00	992.0
122.00	1041.0
123.00	1691.0
124.00	789.0
125.00	651.0
126.20	314.0
127.00	56288.0
128.00	4430.0
129.00	21176.0
130.00	2002.0
131.00	428.0
134.00	603.0
135.00	1635.0
136.00	729.0
137.00	996.0
141.00	2803.0
142.00	962.0
143.00	670.0
146.00	643.0
147.00	1527.0
148.00	3479.0
149.00	784.0
151.00	315.0
153.00	905.0
153.90	660.0
155.00	1761.0
156.10	2430.0
157.10	528.0
158.00	501.0
159.10	431.0
160.00	824.0
161.00	1466.0
162.00	442.0
165.00	1154.0
166.10	829.0
167.00	6281.0
168.00	3628.0
169.10	540.0
172.00	498.0
173.00	741.0
174.00	1271.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

175.10	2088.0
176.10	591.0
177.00	1067.0
178.00	504.0
179.00	4294.0
180.00	3535.0
181.00	1504.0
184.00	475.0
185.00	2306.0
186.10	17136.0
187.10	4822.0
188.00	518.0
189.00	978.0
191.00	543.0
192.00	1361.0
193.00	1634.0
194.00	331.0
196.10	3378.0
197.10	918.0
198.00	120800.0
199.00	8511.0
200.00	719.0
201.60	519.0
203.00	765.0
204.10	4669.0
205.10	7365.0
206.10	32136.0
207.10	4512.0
208.00	1201.0
209.00	417.0
210.00	469.0
211.00	1185.0
215.00	393.0
216.10	639.0
217.00	8273.0
218.00	1135.0
221.10	4482.0
221.80	1190.0
223.10	1845.0
224.10	17544.0
225.10	4541.0
226.00	579.0
227.00	7663.0
228.00	1130.0
229.00	1608.0
231.00	623.0
234.00	470.0
235.00	585.0
236.00	347.0
237.00	822.0
241.00	393.0
242.10	880.0
243.10	975.0
244.10	13130.0
245.10	1794.0
246.00	2535.0
247.00	583.0
249.00	474.0
254.20	694.0
255.10	66064.0
256.10	10412.0
257.10	888.0
258.10	3752.0
259.00	582.0
265.00	1646.0
273.10	2055.0
274.10	5204.0
275.10	28832.0
276.10	4064.0
277.00	2468.0
278.00	514.0
285.00	457.0
293.00	773.0
296.00	7959.0
297.00	1263.0
303.10	968.0
314.10	438.0
315.00	829.0
316.00	502.0
323.10	2654.0
324.10	535.0
327.00	547.0
334.10	1782.0
335.00	430.0
341.00	340.0
346.00	546.0
352.10	677.0
353.00	488.0
354.00	940.0
365.00	3470.0
366.00	471.0
372.10	1278.0
373.10	311.0
383.00	358.0
402.00	548.0
403.00	833.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

421.00	713.0
422.00	578.0
423.10	4563.0
424.10	1464.0
441.10	13039.0
442.10	86136.0
443.10	17728.0
444.10	1579.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

m/z	Abundance
38.00	512.0
39.10	3250.0
49.00	367.0
50.10	12521.0
51.10	50152.0
52.10	2419.0
55.00	307.0
56.00	1425.0
57.00	3446.0
61.00	719.0
62.00	780.0
63.00	2064.0
64.00	305.0
65.00	1098.0
68.10	1041.0
69.00	52784.0
73.00	619.0
74.00	5549.0
75.00	8801.0
76.10	3162.0
77.10	64656.0
78.10	4679.0
79.00	4246.0
80.00	3345.0
81.00	4739.0
82.00	1234.0
83.00	1110.0
85.00	871.0
86.00	1380.0
87.00	575.0
91.00	992.0
92.00	1127.0
93.00	7510.0
94.00	514.0
98.00	5946.0
99.00	4901.0
100.00	590.0
101.00	2828.0
103.00	890.0
104.00	1550.0
105.00	1588.0
106.00	547.0
107.00	21976.0
108.00	3477.0
109.00	647.0
110.00	39008.0
111.00	5936.0
112.00	777.0
113.10	339.0
116.00	1194.0
117.00	18360.0
118.00	1460.0
120.00	323.0
122.00	1407.0
123.00	2199.0
124.00	1027.0
125.10	905.0
126.20	590.0
127.10	73792.0
128.10	5723.0
129.00	28432.0
130.00	2449.0
131.00	406.0
134.00	753.0
135.00	2420.0
136.00	984.0
137.00	1431.0
140.00	303.0
141.00	3957.0
142.10	1248.0
143.10	778.0
146.00	706.0
147.10	2044.0
148.00	4692.0
149.00	1106.0
151.00	523.0
153.00	1315.0
154.00	865.0
155.10	2403.0
156.10	3458.0
157.00	723.0
157.90	733.0
159.00	597.0
160.00	1251.0
161.10	1794.0
162.00	536.0
165.00	1511.0
166.00	1193.0
167.00	8443.0
168.00	4656.0
169.00	745.0
170.90	371.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

172.00	714.0
173.00	1118.0
174.10	1788.0
175.10	3299.0
176.00	924.0
177.10	1571.0
178.00	510.0
179.00	6101.0
180.10	4490.0
181.00	2049.0
184.00	514.0
185.00	2925.0
186.10	24168.0
187.10	6876.0
188.00	850.0
189.00	1388.0
191.00	707.0
192.00	2203.0
193.10	2372.0
194.10	473.0
195.10	350.0
196.00	4780.0
197.20	1234.0
198.00	169216.0
199.00	11721.0
200.00	1014.0
201.50	618.0
203.00	1267.0
204.10	6596.0
205.10	11245.0
206.10	46128.0
207.10	6744.0
208.10	1697.0
209.00	518.0
210.10	668.0
211.10	1759.0
215.00	465.0
216.00	1008.0
217.00	12687.0
218.10	1543.0
221.10	6712.0
221.80	1664.0
223.10	2619.0
224.10	25296.0
225.10	6336.0
226.10	727.0
227.00	10694.0
228.00	1440.0
229.00	2257.0
231.00	821.0
234.00	762.0
235.00	803.0
236.00	523.0
237.10	1046.0
239.00	441.0
241.00	671.0
242.00	1347.0
243.10	1592.0
244.10	19088.0
245.10	2651.0
246.00	4031.0
247.00	898.0
249.00	637.0
253.00	408.0
254.20	1206.0
255.10	103352.0
256.10	14170.0
257.00	1071.0
258.10	5952.0
259.00	1003.0
265.00	2565.0
266.00	343.0
273.10	3011.0
274.10	8355.0
275.10	45544.0
276.10	5942.0
277.00	3952.0
278.00	791.0
283.00	435.0
284.00	316.0
285.00	638.0
293.00	808.0
295.10	309.0
296.10	13084.0
297.10	1817.0
303.10	1604.0
304.00	494.0
314.10	650.0
315.00	1465.0
316.10	806.0
321.00	454.0
323.10	4182.0
324.10	835.0
327.00	836.0
328.10	404.0
332.00	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

333.10	489.0
334.10	2800.0
335.00	738.0
341.00	488.0
346.00	1026.0
352.10	1352.0
353.00	750.0
354.10	1521.0
355.00	312.0
365.10	6063.0
366.00	871.0
371.10	358.0
372.10	2491.0
373.10	532.0
383.00	615.0
402.00	939.0
403.00	1417.0
404.00	527.0
421.10	1263.0
422.00	1391.0
423.10	8440.0
424.10	2438.0
440.20	334.0
441.10	24720.0
442.10	163136.0
443.10	31736.0
444.10	2703.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

m/z	Abundance
38.00	421.0
39.10	2875.0
49.10	319.0
50.10	10677.0
51.10	41664.0
52.00	2116.0
56.00	1097.0
57.00	3029.0
61.00	559.0
62.00	639.0
63.10	1794.0
65.00	984.0
68.00	797.0
69.00	44784.0
73.00	621.0
74.00	4947.0
75.10	7054.0
76.10	2698.0
77.10	56208.0
78.10	3762.0
79.00	3753.0
80.00	2700.0
81.00	3893.0
82.00	1041.0
83.00	1013.0
85.00	754.0
86.00	990.0
87.00	552.0
91.00	985.0
92.00	1005.0
93.00	6533.0
94.00	475.0
98.00	5232.0
99.00	3846.0
100.00	378.0
101.00	2533.0
103.00	772.0
104.00	1514.0
105.00	1384.0
106.00	539.0
107.00	19472.0
108.00	2743.0
109.00	567.0
110.00	33376.0
111.00	5183.0
112.00	553.0
116.10	1093.0
117.00	16552.0
118.00	1229.0
122.00	1265.0
123.00	1854.0
124.00	894.0
125.00	899.0
126.20	550.0
127.10	64216.0
128.00	4618.0
129.00	24504.0
130.00	2164.0
131.10	391.0
134.00	735.0
135.00	2088.0
136.00	918.0
137.10	1133.0
140.10	313.0
141.00	3280.0
142.00	1082.0
143.00	866.0
146.00	701.0
147.00	1825.0
148.00	4350.0
149.10	807.0
151.10	447.0
153.00	1160.0
154.00	781.0
155.10	2176.0
156.10	2635.0
157.10	539.0
158.10	707.0
159.10	482.0
160.00	1042.0
161.00	1732.0
162.00	456.0
165.00	1163.0
166.00	1051.0
167.00	7435.0
168.00	3876.0
169.00	675.0
170.90	303.0
172.00	773.0
173.00	892.0
174.00	1632.0
175.10	2834.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

176.00	791.0
177.00	1438.0
178.00	544.0
179.00	5446.0
180.10	4030.0
181.00	1877.0
184.10	451.0
185.10	2687.0
186.10	21912.0
187.10	6320.0
188.10	724.0
189.00	1467.0
191.10	554.0
192.00	1910.0
193.10	1996.0
194.00	482.0
196.10	4300.0
197.10	1182.0
198.00	157056.0
199.00	10284.0
200.00	808.0
201.50	600.0
203.10	1177.0
204.10	5792.0
205.10	9982.0
206.10	41888.0
207.10	5913.0
208.10	1501.0
209.00	411.0
210.00	551.0
211.10	1680.0
215.00	448.0
216.00	998.0
217.00	10831.0
218.00	1575.0
221.10	6085.0
221.80	1426.0
223.00	2336.0
224.10	24032.0
225.10	6104.0
226.10	710.0
227.00	10092.0
228.00	1507.0
229.10	2019.0
230.00	312.0
231.10	826.0
234.00	609.0
235.10	774.0
236.00	488.0
237.00	907.0
239.00	361.0
241.00	567.0
242.10	1272.0
243.10	1454.0
244.10	18232.0
245.10	2371.0
246.10	3926.0
247.00	774.0
249.00	606.0
253.00	398.0
254.10	1012.0
255.10	94640.0
256.10	14248.0
257.10	1032.0
258.10	5595.0
259.00	885.0
265.00	2294.0
266.00	361.0
273.10	2742.0
274.10	7927.0
275.10	43344.0
276.10	5662.0
277.00	4033.0
278.10	683.0
283.00	439.0
284.00	302.0
285.10	726.0
293.00	863.0
295.10	313.0
296.10	12903.0
297.10	1525.0
303.00	1398.0
304.00	413.0
314.00	537.0
315.00	1500.0
316.00	860.0
321.00	353.0
323.10	4118.0
324.00	695.0
327.00	850.0
328.00	459.0
333.00	440.0
334.10	2741.0
335.10	631.0
341.00	439.0
346.00	894.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49

352.10	1328.0
353.10	1110.0
354.10	1378.0
355.20	310.0
365.00	5990.0
366.00	940.0
372.10	2630.0
373.10	543.0
383.10	668.0
402.10	1006.0
403.10	1367.0
404.00	554.0
421.00	1291.0
422.10	1279.0
423.10	8524.0
424.10	2147.0
441.10	27272.0
442.10	178496.0
443.10	34456.0
444.10	2987.0

Instrument :
BNA_M
ClientSampleId :
DFTPP49