

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041122\
 Data File : BM034613.D
 Acq On : 11 Apr 2022 17:04
 Operator : CG/JU
 Sample : N2338-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0J90

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
 Supervised By :Yogesh Patel 04/14/2022

Quant Time: Apr 12 02:56:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 08 14:30:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.878	152	90570	20.000	ng/u1	0.00
20) Naphthalene-d8	10.689	136	364152	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.524	164	209499	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.265	188	417330	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.442	240	413140	20.000	ng/u1	# 0.00
88) Perylene-d12	23.824	264	492213	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.255	96	11295	4.447	ng/uL	0.00
4) Pyridine-d5	3.684	84	53864	8.788	ng/u1	0.00
7) Phenol-d5	7.043	99	47430m	6.292	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.201	67	159199	31.693	ng/u1	0.00
11) 2-Chlorophenol-d4	7.407	132	150428	25.814	ng/u1	0.00
15) 4-Methylphenol-d8	8.589	113	89568	15.211	ng/u1	0.00
21) Nitrobenzene-d5	9.042	128	87202m	32.583	ng/u1	0.00
24) 2-Nitrophenol-d4	9.766	143	83491	29.964	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.313	165	137119	26.134	ng/u1	0.00
31) 4-Chloroaniline-d4	10.825	131	167053m	23.078	ng/u1	0.00
46) Dimethylphthalate-d6	13.936	166	534831	35.331	ng/u1	0.00
49) Acenaphthylene-d8	14.219	160	639836m	32.885	ng/u1	0.00
54) 4-Nitrophenol-d4	14.771	143	5142m	2.228	ng/u1	0.05
60) Fluorene-d10	15.518	176	475316	36.334	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.636	200	66558	26.125	ng/u1	0.00
73) Anthracene-d10	17.365	188	808784	40.649	ng/u1	0.00
81) Pyrene-d10	19.653	212	1019904	45.231	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	919087	37.121	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041122\
Data File : BM034613.D
Acq On : 11 Apr 2022 17:04
Operator : CG/JU
Sample : N2338-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0J90

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/14/2022

Quant Time: Apr 12 02:56:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 08 14:30:08 2022
Response via : Initial Calibration

