

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041223\
 Data File : BM039406.D
 Acq On : 12 Apr 2023 16:45
 Operator : CG/JU
 Sample : 02287-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BS-3-4

Quant Time: Apr 13 01:48:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 11 16:05:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	179392	20.000	ng	0.00
21) Naphthalene-d8	10.669	136	679908	20.000	ng	0.00
39) Acenaphthene-d10	14.515	164	390655	20.000	ng	0.00
64) Phenanthrene-d10	17.262	188	817407	20.000	ng	0.00
76) Chrysene-d12	21.456	240	844120	20.000	ng	0.00
86) Perylene-d12	23.838	264	981814	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.446	112	521963	43.836	ng	0.02
7) Phenol-d6	7.069	99	821191	53.321	ng	0.04
23) Nitrobenzene-d5	9.034	82	823616	58.875	ng	0.00
42) 2,4,6-Tribromophenol	16.010	330	179496	36.196	ng	0.00
45) 2-Fluorobiphenyl	13.139	172	1792796	62.180	ng	0.00
79) Terphenyl-d14	19.892	244	2901175	63.499	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041223\
Data File : BM039406.D
Acq On : 12 Apr 2023 16:45
Operator : CG/JU
Sample : 02287-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BS-3-4

Quant Time: Apr 13 01:48:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 11 16:05:42 2023
Response via : Initial Calibration

