

Data Path : Z:\HPCHEM1\BNA_M\Data\BM041315\
 Data File : BM000952.D
 Acq On : 13 Apr 2015 16:23
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 14 06:19:09 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 10 16:15:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
2	1,4-Dioxane	0.497	0.517	-4.0	80	0.00
3	n-Nitrosodimethylamine	0.486	0.482	0.8	75	0.00
4 S	2-Fluorophenol	1.058	1.069	-1.0	75	-0.02
5 S	Phenol-d6	1.267	1.223	3.5	74	0.00
6 C	Phenol	1.302	1.250	4.0	74	-0.02
7	bis(2-Chloroethyl)ether	1.142	1.131	1.0	74	0.00
8 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
9 S	Nitrobenzene-d5	0.241	0.239	0.8	75	0.00
10	Nitrobenzene	0.283	0.283	0.0	75	0.00
11	2,4-Dimethylphenol	0.275	0.265	3.6	72	0.00
12 C	2,4-Dichlorophenol	0.243	0.224	7.8	71	0.00
13	Naphthalene	1.009	1.039	-3.0	76	0.00
14 C	Hexachlorobutadiene	0.184	0.191	-3.8	77	0.00
15 C	2-Methylnaphthalene	0.689	0.711	-3.2	77	0.00
16	1-Methylnaphthalene	0.694	0.718	-3.5	77	-0.01
17 I	Acenaphthene-d10	1.000	1.000	0.0	74	0.00
18 S	2,4,6-Tribromophenol	0.179	0.183	-2.2	61	0.00
19 S	2-Fluorobiphenyl	1.401	1.689	-20.6	76	0.00
20	Acenaphthylene	2.032	2.543	-25.1#	70	0.00
21 C	Acenaphthene	1.380	1.425	-3.3	71	0.00
22 P	2,4-Dinitrophenol	0.058	0.031#	46.6#	55	0.00
23	Fluorene	1.606	1.585	1.3	70	0.00
24 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
25	Hexachlorobenzene	0.270	0.302	-11.9	72	0.00
26 C	Pentachlorophenol	0.086	0.071	17.4	62	0.00
27	Phenanthrene	1.292	1.434	-11.0	70	0.00
28	Anthracene	1.184	1.155	2.4	68	0.00
29 C	Fluoranthene	1.350	1.341	0.7	65	0.00
30 I	Chrysene-d12	1.000	1.000	0.0	62	0.00
31	Benzidine	0.300	0.202	32.7#	50	0.02
32	Pyrene	1.625	1.692	-4.1	63	0.00
33 S	Terphenyl-d14	0.609	0.625	-2.6	65	0.02
34	Benzo(a)anthracene	1.443	1.521	-5.4	67	0.00
35	3,3'-Dichlorobenzidine	0.389	0.315	19.0	55	0.02
36	Chrysene	1.475	1.483	-0.5	61	0.00
37	Indeno(1,2,3-cd)pyrene	1.506	1.500	0.4	62	0.00
38 I	Perylene-d12	1.000	1.000	0.0	60	0.00
39	Benzo(b)fluoranthene	1.565	1.734	-10.8	70	0.01
40	Benzo(k)fluoranthene	1.506	1.508	-0.1	58	0.01
41 C	Benzo(a)pyrene	1.378	1.393	-1.1	62	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM041315\
Data File : BM000952.D
Acq On : 13 Apr 2015 16:23
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Apr 14 06:19:09 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 10 16:15:17 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
42	Dibenzo(a,h)anthracene	1.278	1.294	-1.3	61	0.00
43	Benzo(g,h,i)perylene	1.410	1.493	-5.9	64	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0