

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041323\
 Data File : BM039411.D
 Acq On : 13 Apr 2023 14:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 13 16:08:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 11 16:05:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.01
2	1,4-Dioxane	40.000	43.080	-7.7	103	0.01
3	Pyridine	40.000	42.015	-5.0	97	0.01
4	n-Nitrosodimethylamine	40.000	42.100	-5.3	100	0.00
5 S	2-Fluorophenol	80.000	80.964	-1.2	96	0.02
6	Aniline	40.000	41.523	-3.8	97	0.01
7 S	Phenol-d6	80.000	83.273	-4.1	98	0.02
8	2-Chlorophenol	40.000	41.107	-2.8	97	0.01
9	Benzaldehyde	40.000	42.831	-7.1	101	0.00
10 C	Phenol	40.000	40.719	-1.8	96	0.02
11	bis(2-Chloroethyl)ether	40.000	40.728	-1.8	97	0.01
12	1,3-Dichlorobenzene	40.000	41.363	-3.4	98	0.01
13 C	1,4-Dichlorobenzene	40.000	41.123	-2.8	97	0.01
14	1,2-Dichlorobenzene	40.000	41.443	-3.6	99	0.01
15	Benzyl Alcohol	40.000	39.204	2.0	91	0.01
16	2,2'-oxybis(1-Chloropropane	40.000	42.688	-6.7	102	0.01
17	2-Methylphenol	40.000	40.755	-1.9	96	0.02
18	Hexachloroethane	40.000	40.233	-0.6	96	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	40.969	-2.4	97	0.01
20	3+4-Methylphenols	40.000	41.266	-3.2	96	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.01
22	Acetophenone	40.000	42.551	-6.4	95	0.01
23 S	Nitrobenzene-d5	80.000	86.713	-8.4	97	0.00
24	Nitrobenzene	40.000	42.596	-6.5	96	0.00
25	Isophorone	40.000	42.943	-7.4	96	0.01
26 C	2-Nitrophenol	40.000	41.617	-4.0	92	0.01
27	2,4-Dimethylphenol	40.000	42.943	-7.4	96	0.01
28	bis(2-Chloroethoxy)methane	40.000	41.512	-3.8	94	0.01
29 C	2,4-Dichlorophenol	40.000	42.897	-7.2	95	0.01
30	1,2,4-Trichlorobenzene	40.000	42.438	-6.1	96	0.00
31	Naphthalene	40.000	42.650	-6.6	96	0.01
32	Benzoic acid	40.000	37.774	5.6	81	0.01
33	4-Chloroaniline	40.000	40.143	-0.4	89	0.01
34 C	Hexachlorobutadiene	40.000	42.535	-6.3	96	0.01
35	Caprolactam	40.000	41.089	-2.7	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.782	-7.0	94	0.01
37	2-Methylnaphthalene	40.000	42.112	-5.3	95	0.01
38	1-Methylnaphthalene	40.000	42.456	-6.1	96	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.337	-8.3	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.680	-1.7	86	0.01
42 S	2,4,6-Tribromophenol	80.000	80.916	-1.1	89	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.568	-6.4	92	0.01
44	2,4,5-Trichlorophenol	40.000	41.904	-4.8	90	0.01
45 S	2-Fluorobiphenyl	80.000	86.120	-7.7	95	0.00
46	1,1'-Biphenyl	40.000	42.681	-6.7	95	0.01
47	2-Chloronaphthalene	40.000	42.033	-5.1	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041323\
 Data File : BM039411.D
 Acq On : 13 Apr 2023 14:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 13 16:08:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 11 16:05:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.499	-3.7	90	0.01
49	Acenaphthylene	40.000	42.520	-6.3	94	0.00
50	Dimethylphthalate	40.000	41.342	-3.4	92	0.00
51	2,6-Dinitrotoluene	40.000	41.842	-4.6	91	0.01
52 C	Acenaphthene	40.000	42.676	-6.7	95	0.00
53	3-Nitroaniline	40.000	37.785	5.5	81	0.01
54 P	2,4-Dinitrophenol	40.000	37.408	6.5	80	0.00
55	Dibenzofuran	40.000	41.694	-4.2	93	0.00
56 P	4-Nitrophenol	40.000	38.966	2.6	80	0.01
57	2,4-Dinitrotoluene	40.000	41.698	-4.2	89	0.00
58	Fluorene	40.000	42.035	-5.1	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.834	-4.6	92	0.01
60	Diethylphthalate	40.000	40.404	-1.0	90	0.01
61	4-Chlorophenyl-phenylether	40.000	41.190	-3.0	92	0.01
62	4-Nitroaniline	40.000	35.749	10.6	75	0.01
63	Azobenzene	40.000	42.037	-5.1	93	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	85	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.029	-5.1	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.963	-7.4	91	0.00
67	4-Bromophenyl-phenylether	40.000	42.407	-6.0	90	0.00
68	Hexachlorobenzene	40.000	42.471	-6.2	91	0.00
69	Atrazine	40.000	41.481	-3.7	88	0.01
70 C	Pentachlorophenol	40.000	42.853	-7.1	87	0.00
71	Phenanthrene	40.000	42.474	-6.2	91	0.00
72	Anthracene	40.000	42.956	-7.4	91	0.00
73	Carbazole	40.000	41.622	-4.1	88	0.00
74	Di-n-butylphthalate	40.000	41.270	-3.2	86	0.01
75 C	Fluoranthene	40.000	42.418	-6.0	90	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	80	0.00
77	Benzidine	40.000	45.195	-13.0	84	0.01
78	Pyrene	40.000	45.169	-12.9	90	0.00
79 S	Terphenyl-d14	80.000	86.456	-8.1	86	0.00
80	Butylbenzylphthalate	40.000	42.745	-6.9	84	0.00
81	Benzo(a)anthracene	40.000	41.655	-4.1	83	0.00
82	3,3'-Dichlorobenzidine	40.000	42.017	-5.0	83	0.00
83	Chrysene	40.000	42.327	-5.8	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.882	-2.2	81	0.00
85 c	Di-n-octyl phthalate	40.000	38.509	3.7	76	0.00
86 I	Perylene-d12	20.000	20.000	0.0	78	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.557	-3.9	79	0.01
88	Benzo(b)fluoranthene	40.000	40.766	-1.9	79	0.00
89	Benzo(k)fluoranthene	40.000	42.933	-7.3	83	0.00
90 C	Benzo(a)pyrene	40.000	42.224	-5.6	81	0.00
91	Dibenzo(a,h)anthracene	40.000	41.378	-3.4	79	0.01
92	Benzo(g,h,i)perylene	40.000	41.819	-4.5	80	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041323\
Data File : BM039411.D
Acq On : 13 Apr 2023 14:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Apr 13 16:08:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 11 16:05:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0