

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041324\
 Data File : BM045210.D
 Acq On : 13 Apr 2024 18:17
 Operator : MA/JU
 Sample : P2014-18
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E0ME3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/16/2024
 Supervised By :mohammad ahmed 04/17/2024

Quant Time: Apr 15 00:36:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 15 00:32:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.606	152	1454	0.400	ng/u1	0.00
4) Naphthalene-d8	10.375	136	4484	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.251	164	2433	0.400	ng/u1	-0.01
13) Phenanthrene-d10	17.014	188	5357m	0.400	ng/u1	-0.02
17) Chrysene-d12	21.237	240	4975m	0.400	ng/u1	0.00
23) Perylene-d12	23.438	264	5037m	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.201	96	6626	3.469	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.991	152	2201	0.362	ng/u1	0.00
18) Fluoranthene-d10	19.057	212	5434	0.432	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041324\
Data File : BM045210.D
Acq On : 13 Apr 2024 18:17
Operator : MA/JU
Sample : P2014-18
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E0ME3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/16/2024
Supervised By :mohammad ahmed 04/17/2024

Quant Time: Apr 15 00:36:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM040424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 15 00:32:25 2024
Response via : Initial Calibration

