

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
 Data File : BM000961.D
 Acq On : 13 Apr 2015 22:33
 Operator : TP/IZ
 Sample : PB82711BL
 Misc : SBLK44
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK44

Quant Time: Apr 14 18:52:41 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 14 08:28:33 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	134322	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	549766	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.26	164	308904	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.00	188	676123	20.00	ng/ul	0.00
75) Chrysene-d12	21.20	240	727227	20.00	ng/ul	0.00
83) Perylene-d12	23.38	264	693627	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
5) Phenol-d5	6.78	99	406549	40.26	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	248747	39.45	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	339925	40.78	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	316568	40.24	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	142505	39.16	ng/ul	0.00
22) 2-Nitrophenol-d4	9.47	143	147567	37.40	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.01	165	287261	36.52	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	394328	41.86	ng/ul	0.00
43) Dimethylphthalate-d6	13.67	166	833844	41.23	ng/ul	0.00
46) Acenaphthylene-d8	13.95	160	1104993	37.75	ng/ul	0.00
51) 4-Nitrophenol-d4	14.47	143	91872	25.91	ng/ul	0.00
57) Fluorene-d10	15.26	176	793500	38.64	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.39	200	74784	20.70	ng/ul	0.00
70) Anthracene-d10	17.10	188	1217238	39.41	ng/ul	0.00
76) Pyrene-d10	19.40	212	1345371	40.98	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.24	264	1227510	40.45	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
Data File : BM000961.D
Acq On : 13 Apr 2015 22:33
Operator : TP/IZ
Sample : PB82711BL
Misc : SBLK44
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK44

Quant Time: Apr 14 18:52:41 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 14 08:28:33 2015
Response via : Initial Calibration

