

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
 Data File : BM000976.D
 Acq On : 14 Apr 2015 08:06
 Operator : TP/IZ
 Sample : G1830-03MSD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P001-SS003-0006-01MSD

Quant Time: Apr 15 15:38:34 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 14 08:59:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	203838	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	828514	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.26	164	454798	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.00	188	969169	20.00	ng/ul	0.00
75) Chrysene-d12	21.20	240	995092	20.00	ng/ul	0.00
83) Perylene-d12	23.38	264	954344	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
5) Phenol-d5	6.78	99	544733	35.55	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	343838	35.94	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	445568	35.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	430936	36.10	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	221111	40.32	ng/ul	0.00
22) 2-Nitrophenol-d4	9.47	143	237644	39.96	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.01	165	413657	34.90	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	567705	39.99	ng/ul	0.00
43) Dimethylphthalate-d6	13.67	166	1304219	43.80	ng/ul	0.00
46) Acenaphthylene-d8	13.95	160	1615999	37.50	ng/ul	0.00
51) 4-Nitrophenol-d4	14.47	143	197137	37.76	ng/ul	0.00
57) Fluorene-d10	15.26	176	1093138	36.16	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.38	200	129437	24.99	ng/ul	0.00
70) Anthracene-d10	17.10	188	1594289	36.01	ng/ul	0.00
76) Pyrene-d10	19.40	212	1625323	36.18	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.24	264	1312254	31.43	ng/ul	0.00

Target Compounds

						Qvalue
6) Phenol	6.81	94	635959	38.13	ng/ul	98
10) 2-Chlorophenol	7.17	128	465069	34.04	ng/ul	99
15) N-Nitroso-di-n-propylamine	8.40	70	372479	38.57	ng/ul	98
33) 4-Chloro-3-methylphenol	11.68	107	495031	40.23	ng/ul	98
44) Dimethylphthalate	13.72	163	113751	3.34	ng/ul	98
49) Acenaphthene	14.32	153	1157099	36.54	ng/ul	99
52) 4-Nitrophenol	14.49	109	161675	39.46	ng/ul	94
54) 2,4-Dinitrotoluene	14.63	165	377294	39.93	ng/ul	99
68) Pentachlorophenol	16.66	266	192910	35.96	ng/ul	96
77) Pyrene	19.43	202	2372234	37.37	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
Data File : BM000976.D
Acq On : 14 Apr 2015 08:06
Operator : TP/IZ
Sample : G1830-03MSD
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P001-SS003-0006-01MSD

Quant Time: Apr 15 15:38:34 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 14 08:59:44 2015
Response via : Initial Calibration

