

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
 Data File : BM000987.D
 Acq On : 14 Apr 2015 17:14
 Operator : TP/IZ
 Sample : G1858-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P001-SD008-0006-02

Quant Time: Apr 15 17:10:24 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 15 06:02:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	363527	20.00	ng/ul	0.00
18) Naphthalene-d8	10.38	136	1492562	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.26	164	811981	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.00	188	1733353	20.00	ng/ul	0.00
75) Chrysene-d12	21.20	240	1839914	20.00	ng/ul	0.00
83) Perylene-d12	23.38	264	1754517	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
5) Phenol-d5	6.78	99	680072	24.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	416244	24.39	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	525721	23.30	ng/ul	0.00
13) 4-Methylphenol-d8	8.31	113	567491	26.66	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	235905	23.88	ng/ul	0.00
22) 2-Nitrophenol-d4	9.47	143	247932	23.14	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.01	165	525215	24.60	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	675048	26.39	ng/ul	0.00
43) Dimethylphthalate-d6	13.67	166	1918618	36.09	ng/ul	0.00
46) Acenaphthylene-d8	13.95	160	2030648	26.39	ng/ul	0.00
51) 4-Nitrophenol-d4	14.47	143	226570	24.31	ng/ul	0.00
57) Fluorene-d10	15.26	176	1443181	26.74	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.38	200	151715	16.38	ng/ul	0.00
70) Anthracene-d10	17.10	188	2316079	29.25	ng/ul	0.00
76) Pyrene-d10	19.40	212	2216270	26.68	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.24	264	1789559	23.32	ng/ul	0.00

Target Compounds					Qvalue
6) Phenol	6.80	94	67410	2.27 ng/ul	97
44) Dimethylphthalate	13.72	163	571289	9.38 ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
Data File : BM000987.D
Acq On : 14 Apr 2015 17:14
Operator : TP/IZ
Sample : G1858-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P001-SD008-0006-02

Quant Time: Apr 15 17:10:24 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 15 06:02:42 2015
Response via : Initial Calibration

