

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
 Data File : BM000989.D
 Acq On : 14 Apr 2015 18:25
 Operator : TP/IZ
 Sample : G1858-13
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P001-SS017-0006-01

Quant Time: Apr 15 17:12:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 15 06:02:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	370508	20.00	ng/ul	0.00
18) Naphthalene-d8	10.38	136	1487616	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.26	164	808085	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.00	188	1733385	20.00	ng/ul	0.00
75) Chrysene-d12	21.20	240	1827405	20.00	ng/ul	-0.01
83) Perylene-d12	23.38	264	1760107	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
5) Phenol-d5	6.78	99	697225	25.03	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	454351	26.12	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	574995	25.01	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	538194	24.80	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	271834	27.61	ng/ul	0.00
22) 2-Nitrophenol-d4	9.47	143	288962	27.06	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.01	165	556412	26.14	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	713966	28.01	ng/ul	0.00
43) Dimethylphthalate-d6	13.67	166	1838889	34.75	ng/ul	0.00
46) Acenaphthylene-d8	13.95	160	2203887	28.78	ng/ul	0.00
51) 4-Nitrophenol-d4	14.47	143	221593	23.89	ng/ul	0.00
57) Fluorene-d10	15.25	176	1544765	28.76	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.38	200	181371	19.58	ng/ul	0.00
70) Anthracene-d10	17.10	188	2298631	29.03	ng/ul	0.00
76) Pyrene-d10	19.40	212	2349444	28.48	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.24	264	2003061	26.01	ng/ul	0.00

Target Compounds

						Qvalue
6) Phenol	6.80	94	69278	2.29	ng/ul	99
44) Dimethylphthalate	13.72	163	301394	4.97	ng/ul	98
69) Phenanthrene	17.04	178	131534	1.32	ng/ul	99
74) Fluoranthene	19.07	202	266334	2.41	ng/ul	99
77) Pyrene	19.43	202	208467	1.79	ng/ul	98
80) Benzo(a)anthracene	21.19	228	128352	1.18	ng/ul	99
82) Chrysene	21.24	228	108321	1.04	ng/ul	99
85) Benzo(b)fluoranthene	22.73	252	114188	1.05	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041415\
Data File : BM000989.D
Acq On : 14 Apr 2015 18:25
Operator : TP/IZ
Sample : G1858-13
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P001-SS017-0006-01

Quant Time: Apr 15 17:12:49 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 15 06:02:42 2015
Response via : Initial Calibration

